

METHODOLOGY

Open Access

Performance of coumarin-derived dendrimer-based fluorescence-linked immunosorbent assay (FLISA) to detect malaria antigen

Seon-Ju Yeo¹, Dinh Thi Huong¹, Jin-Hee Han², Jung-Yeon Kim³, Won-Ja Lee³, Ho-Joon Shin⁴, Eun-Taek Han^{2*} and Hyun Park^{1*}

Abstract

Background: Due to limitation of conventional malaria diagnostics, including microscopy, polymerase chain reaction (PCR), and enzyme-linked immunosorbent assay (ELISA), alternative accurate diagnostics have been demanded for improvement of sensitivity and specificity.

Methods: Serially diluted *Plasmodium* LDH antigens, *Plasmodium falciparum*-infected human red blood cells (RBC) derived from *in vitro* culture or patient's samples were used for evaluation of the performance of fluorescence-linked immunosorbent assay (FLISA). Microscopic examination was used to determine parasite density and the performance of FLISA was compared to ELISA. Finally, sensitivity and specificity of FLISA was determined by human specimens infected with *P. falciparum*, *Plasmodium vivax*, *Toxoplasma gondii*, and amoebae.

Results: As a result of FLISA, the fluorescent intensity was highly correlated with antigen amount and FLISA was more sensitive than ELISA. FLISA detected at least 0.01 ng/ml of pLDH antigen, which showed 1,000-fold higher sensitivity than ELISA. *In vitro*-cultured *P. falciparum* was detected up to 20 parasite number/ μ L in FLISA but 5120 parasite number/ μ L in sandwich ELISA. *In vitro* *P. falciparum*-infected RBC number was highly correlated with fluorescent intensity ($R^2 = 0.979$), showing that FLISA was reliable for detection of *P. falciparum* and available for quantification of parasite numbers. Furthermore, eighteen patient samples infected with *P. falciparum* ($n = 9$) and *P. vivax* ($n = 9$) showed 100% of sensitivity (18/18). FLISA showed 96.3% of specificity (26/27) because one sample of patient blood infected with *T. gondii* gave a false positive reactivity among healthy donors ($n = 9$), *T. gondii*-infected patients ($n = 9$), and amoeba-infected patients ($n = 9$).

Conclusion: FLISA has a keen and high performance to detect malaria antigen, suggesting a potential assay as malaria immunodiagnostic.

Keywords: Coumarin-derived dendrimer, Fluorescence-linked immunosorbent assay (FLISA), Lactate dehydrogenase (LDH), ELISA, *Plasmodium falciparum*, *Plasmodium vivax*

* Correspondence: ethan@kangwon.ac.kr; hyunpk@wonkwang.ac.kr

²Department of Medical Environmental Biology and Tropical Medicine, School of Medicine, Kangwon National University, Chuncheon, Gangwon-do, Republic of Korea

¹Zoonosis Research Center, Department of Infection Biology, School of Medicine, Wonkwang University, Iksan, Jeonbuk 570-749, Republic of Korea

Full list of author information is available at the end of the article

Background

Five human infectious malarias, including *Plasmodium falciparum*, *Plasmodium vivax*, *Plasmodium ovale*, *Plasmodium malariae*, and *Plasmodium knowlesi*, remain a serious health issue and are widespread in developing countries, continuing to cause about 219 million cases and one million deaths annually [1]. National malaria mortality rates are, however, particularly difficult to assess reliably and are underestimated in some important endemic areas due to misunderstanding malaria infection [2-4]. Therefore, the establishment of accurate diagnostics of malaria based on quantification of malaria parasite number with high sensitivity and specificity is the top priority for the management of malaria.

As a gold standard of malaria diagnostics, thick and thin blood smears are reliable but it was claimed that in the case of low parasitaemia, aggregation in a specific area of the smear was noticed and blood smearing was less sensitive than immunodiagnostics and PCR methods [5,6]. As an alternative quantitative diagnosis of malaria, over the last few decades, numerous polymerase chain reactions (PCR)-based diagnostic tests targeting RNA or DNA have been developed to confirm the malaria infection in addition to microscopic observation [7-9]. Quantification of malaria infectious parasite numbers in patients with only RNA or DNA copy numbers has been implemented but the requirement of equipped laboratory facilities presents an obstacle to application to a field setting [10]. Due to the limitation of PCR, different immunological assays that use antibodies to detect parasites have been developed with greater potential for adaption to field application than previous approaches, even though immune assays provide sensitivity issues [11]. Therefore, immunological assays have become the basis of most commercial diagnostic test kits, with most interest focused on the use of monoclonal antibodies (mAbs). Using mAbs, enzyme-linked immunosorbent assay (ELISA) method and fluorescence-linked immunosorbent assay (FLISA) have established significant diagnostic biomarkers [12,13]. Resulting devices for the diagnosis of malaria based on malaria-specific antigens, such as histidine-rich protein 2 (HRP2) and lactate dehydrogenase (LDH), have been developed as alternative diagnostics to microscopy and PCR [14].

Microscopy has often been the routine diagnostic technology available in developing countries. However, it has been considered to show variable sensitivity depending on expertise of microscopist [15]. Despite continuous application as key diagnostic tests, microscopy techniques have limitations as universal or targeted donor screening tests due to lack of sensitivity at low parasitaemia [16]. As a higher throughput method of malaria diagnosis, ELISA is suitable for epidemiological surveys [12].

Previously, performance of two monoclonal antibodies (D2H and D7E) targeting conserved 31 amino acids of *Plasmodium* LDH was shown to be potential to be useful

for malaria diagnostics [17]. Besides, light-emitting diode (LED)-based novel organic fluorophore, coumarin-derived dendrimer was developed for malaria diagnostics [13]. As a light source of biosensor, LED has the advantage, because laser diode has various limitations such as expense, difficult operation, limited emission wavelength selections, and a short lifetime [13]. Therefore, development of LED light based diagnostics has advantage as further diagnostics. In this study, the enhanced coumarin-derived dendrimer-based malarial FLISA assay with novel monoclonal antibodies (D2H and D7E) was compared with ELISA.

Methods

In vitro culture and determination of parasitaemia

Plasmodium falciparum FCR-3 (ATCC 50005) were purchased from ATCC (Manassas, VA, USA) and strains used in this study were kept in -80°C as frozen stocks. *Plasmodium falciparum* was cultured by standard procedure as described previously [18], using a 5% haematocrit of type O-positive human red erythrocytes suspended in RPMI 1640 medium with 5% NaHCO₃, 0.5% Albumax, 25 µg/mL gentamicin and supplemented with heat-inactivated 10% type O-positive human serum. The six-well plates were placed in an incubator (5% CO₂, 5% O₂, and 90% N₂ atmosphere) at 37°C, and the medium was changed daily at least 5% parasitaemia. Blood smears were stained with Giemsa for counting parasite numbers with microscopy under 1,000 × magnifications. Parasite density was determined as a percentage of infected erythrocytes in fields of total 500 erythrocytes in the study. Animal experiments were performed at the experimental protocol, which was approved by the Animal Care and Use Committee at Wonkwang University.

Purification of coumarin-derived dendrimer-or horseradish peroxidase (HRP)-bioconjugates

Subcloned cells (D7E or D2H) secreting pLDH were kindly provided by Professor Ho-Joon Shin in Ajou University, which were used in the previous report [17]. To obtain mouse ascitic fluid, 1 × 10⁷ cells were injected in incomplete Freund's adjuvant-primed ICR mice. Ascitic fluids harvested were further processed for purification of antibodies by protein A agarose 4B column (Incospahrm, Korea) according to manufacturer's instruction.

Purified D2H (1 mg/ml) were gently mixed with conjugation buffer (0.1 M NaHCO₃, pH 8.5) for 30 min at room temperature (RT) with gentle rotation. After coumarin-derived dendrimer (a kind donation of Dr Hak Sung Kim, Wonkwang University, Korea) was dissolved in DMSO (1 mg/ml), antibody (D7E) in conjugated buffer was incubated with 6 moles per mole of antibody for overnight at 4°C. Then the mixture was filtered by using desalting column (Thermo Scientific™ Zeba™ Spin Desalting columns) to remove free coumarin-derived dendrimer. Coumarin-


derived dendrimer bioconjugate was collected with centrifugation at 1,000 × g for 2 min at 4°C.

D2H (1 mg/ml) was conjugated with horseradish peroxidase (HRP) in EZ-Link™ Plus Activated Peroxidase (Thermo Scientific, Rockford, IL, USA) according to manufacturer's instructions. Briefly, 1 mg/ml of D2H in carbonate-bicarbonate buffer (pH 9.4) were reacted with 1 mg of lyophilized peroxidase at RT for one hour. After additional reaction with 10 µL of sodium cyanoborohydride for 15 min at RT, reaction stopped by 20 µL of quenching buffer for 15 min at RT. The HRP bioconjugates were further purified with conjugate purification kit (Thermo Scientific, Rockford, IL, USA) to remove unbound HRP.

Fluorescence-linked immunosorbent assay (FLISA)

FLISA were performed as previously described [13]. *Plasmodium falciparum* and *Plasmodium vivax* recombinant LDH (rLDH) were purchased from C&K Bio Resource

Inc, Korea. Briefly, 1 µg/well D7E were prepared for coating a black 96-well microtitre plate (Greiner, Germany) overnight and next day, plate was blocked with 2% casein-based blocking buffer at 37°C. After two hours, plate was washed with 200 µL of phosphate buffered saline with 0.05% Tween 20 (PBS-T, pH 7.4) and 100 µL of sample were added to each well and incubated at 37°C. After two hours, plate was washed with 200 µL of PBS-T. Finally, 200 µL of coumarin-derived dendrimer-conjugated D2H was added to detect antigens for one hour. Stringent washing with PBS-T was performed six times to remove unbound antibodies and 100 µL of PBS was added to each well to measure the fluorescence. Fluorescence with 460 nm excitation and 560 nm emission was measured using an Infinite F200 microplate reader (TECAN, Männedorf, Switzerland). Relative fluorescence was determined by extracting fluorescence value of negative control.



Sandwich enzyme-linked immunosorbent assay (ELISA)

Polystyrene, 96-well plate (NUNC, Pasadena, Texas, USA) was coated with 100 µL (10 µg/mL) of D7E in coating buffer (50 mM bicarbonate/carbonate coating buffer, pH 9.6) and stored in an incubator overnight at 4°C. After incubation, the plate well was washed four times with 200 µL of PBS-T and blocked with 1% casein-based blocking buffer at 37°C for two hours. After washing with PBS-T buffer twice, differently diluted analytes were allowed to each well. The 100 µL of HRP conjugates D2H (2 µg/ml) was applied to each well at 27°C for one hour. After washing with PBS-T five times, plates were filled with 100 µL of 3,3',5,5'-tetra methyl benzidine (TMB) substrate solution (Invitrogen, Camarillo, CA, USA) in the dark for 30 min, following stopping with additional 100 µL of sulphuric acid (0.5 M) per well. Optical density (OD) was measured by a microplate reader at 450 nm. Data were expressed as mean ± standard deviation.

Ethical considerations

Professor Tai-Soon Yong (Yonsei University, College of Medicine, Seoul, Korea) kindly provided amoeba-infected specimens and Professor Ho-Woo Nam (Catholic University, College of Medicine, Seoul, Korea) kindly provided *Toxoplasma gondii*-infected samples. Malaria patient samples were provided from professor Eun-Taek Han (Kangwon National University, School of Medicine, Chuncheon, Korea). The study was approved by the Kangwon National University Hospital Institutional Review Board (Approval No. 10-041-07).

Results

Coumarin-derived dendrimer was chosen as labelling fluorophore, which was confirmed to have good spectroscopic properties in previous study [13]. In this study, two newly developed mAbs targeting pLDH were used for FLISA instead of commercialized mAbs of pLDH, and performance

in FLISA was defined to determine antigen-antibody reactivity. Figure 1A illustrates the schematic representations of the conjugation reactions to understand the principle of the coumarin-derived dendrimer-based FLISA with three analytes (*Plasmodium* recombinant LDH, *in vitro* *P. falciparum*-infected human RBC, and patients bloods). N-hydroxysuccinimide (NHS) esters of coumarin-derived dendrimers react with primary amine groups (-NH₂) of mAbs in buffer (pH 8.5) to form stable amide bonds upon release of NHS. Further FLISA procedure is followed on conventional sandwich ELISA and in FLISA, fluorescence intensity indicates the interaction between parasite antigen and coumarin-derived dendrimer-conjugated antibody. The wavelength of maximum absorption (usually the same as the excitation maximum) of coumarin-derived dendrimer is 460 nm and excitation spectrum of a given bioconjugate is 590 nm [13].

Plasmodium LDH is an enzyme of glycolytic pathway of different *Plasmodium* species at both the asexual and sexual stages and also correlated with the number of parasites present in the plasma of infected patients [19–21]. The LDH from *P. vivax*, *P. malariae*, and *P. ovale* has 87% sequence identity with pLDH from *P. falciparum* [22]. In current study, target epitope (1–31 amino acid) of *Plasmodium* LDH were aligned in BLAST, using BLASTP 2.2.29+ program and the epitope sequence of *Plasmodium* LDH shows 97–100% identities among seven *Plasmodium* species [*P. falciparum* (GenBank accession no. AEX28368.1), *P. vivax* (GenBank accession no. AAS77573.1), *P. malariae* (GenBank accession no. AAS77572.1), *P. ovale* (GenBank accession no. AAS77571.1), *P. knowlesi* (GenBank accession no. AEL88505.1), *Plasmodium yoelii* (GenBank accession no. XP724101.1), and *Plasmodium berghei* (GenBank accession no. XP679401.1)]. In contrast, the epitope of *Plasmodium* LDHs shows 76% identities of *Eimeria maxima* LDH (GenBank accession no. AAN38977.1) and 66% of *Toxoplasma gondii* LDH (GenBank

Table 1 Alignment of pLDH epitope among *Plasmodium* species

Species	Epitope region in LDH	Identity	GenBank accession no.
<i>P. vivax</i>	75 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 105	31/31(100%)	AAS77573.1
<i>P. falciparum</i>	61 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 91	31/31(100%)	AEX28368.1
<i>P. knowlesi</i>	82 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 112	31/31(100%)	AEL88505.1
<i>P. yoelii</i> 17XNL	82 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 112	31/31(100%)	XP724101.1
<i>P. berghei</i> ANKA	82 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 112	31/31(100%)	XP679401.1
<i>P. malariae</i>	75 FTKVPGKSDKEWNRRDDLLPLNNKIMIEIGGH 105	30/31(97%)	AAS77572.1
<i>P. ovale</i>	75 FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH 105	31/31(100%)	AAS77571.1
<i>Eimeria maxima</i>	88 TKIPGKSDKEWSRMDLLPVNIKIMREVGG 116	22/29(76%)	AAN38977.1
<i>Toxoplasma gondii</i>	86 LTKVPGKSDKEWSRNLLPFNAKIIREVA 114	19/29(66%)	XP002368488.1
<i>Homo sapiens</i>	98 AGARQQEGESRLNLVQRN123	7/31 (23%)	CAE11711

Plasmodium LDH epitope (FTKAPGKSDKEWNRRDDLLPLNNKIMIEIGGH) is highly conserved among seven *Plasmodium* spp. and the percentage of sequence identity is determined by epitope alignments across the listed in NCBI.

accession no. XP002368488.1). This epitope has only 26% identities in *Homo sapiens* LDH (GenBank accession no. CAE11711) (Table 1). Therefore, antibodies targeting this epitope may act as pan-specific anti-LDH, which is able to react with most *Plasmodium* spp.

To investigate the ability of FLISA to quantify malaria antigen, FLISA were performed with different concentration

of *P. falciparum*- and *P. vivax*-recombinant LDH antigens at six-point standard curve based on two-fold dilutions from 0.01 to 1,000 ng/mL. Well-defined linear regressions were produced at FLISA for both *P. falciparum*- ($R^2 = 0.972$) and *P. vivax* ($R^2 = 0.967$) LDH (Figure 2A and B). *Plasmodium* LDH antigen was detectable up to 10 ng/mL in ELISA at the same condition with FLISA (Figure 2C). As OD value of




Figure 2 Comparison of FLISA and ELISA detectability against *Plasmodium* LDH. Standard curve for a detectable range of FLISA were generated: *P. falciparum* rLDH in FLISA (A), *P. vivax* rLDH in FLISA (B), and *P. vivax* rLDH in ELISA (C). In FLISA, 0.01 ng/uL of antigen was detectable but 10 ng/uL of antigen was a minimum detectable antigen level in ELISA.




Figure 3 Comparison of FLISA (A) and ELISA (B) analyses of *Plasmodium falciparum*-infected human red blood cells. Human bloods were infected with *P. falciparum* FCR-3 for three days and when parasitaemia approached 3.2%, infected bloods were serially diluted and applied FLISA and ELISA.

0.1 is positive, the detection threshold of FLISA is at least 1,000-fold higher than ELISA, implying that FLISA would contribute to increase detectability of pLDH in early infection when it is hard to be detected in conventional ELISA. Furthermore, *in vitro*-cultured *P. falciparum* was prepared to determine FLISA performance to detect *P. falciparum*-infected blood RBC. *P. falciparum* infected human RBC and cultured for three days to induce 3.2% of parasitaemia in 1.6×10^8 total RBC. As a result of FLISA, fluorescent intensity and *P. falciparum* density was highly correlated and reliable ($R^2 = 0.97$), between 20 and 5,120 parasites (Figure 3A). In contrast, detection threshold of ELISA was 5,120 parasites (Figure 3B). Therefore, FLISA would be a useful diagnostic assay with high detectability to quantify low number of parasite density than ELISA.

To determine the sensitivity and specificity of FLISA, *P. falciparum* ($n = 9$) and *P. vivax* ($n = 9$) infected patient's bloods were applied to FLISA and FLISA showed 100% of sensitivity (18/18). In contrast, non-malaria disease samples among healthy donors ($n = 9$), *T. gondii*- ($n = 9$), and amoeba-infected patients ($n = 9$) showed one false positive reactivity with *T. gondii*-infected patient blood. Therefore, specificity of FLISA was 96.3% (26/27) (Figure 4).

Conclusion

Since elimination of malaria infection by WHO, malaria prevalence is getting low but still prevalent worldwide. In Republic of Korea (ROK), after reemergence of malaria near location of demilitarized zone (DMZ) in 1993, the annual incidence has increased rapidly [23-25]. Until 2010, ten cases of transfusion-transmitted malaria have been reported in ROK [26]. The incidence of transfusion-transmitted malaria among people residing in endemic areas is unknown and a substantial proportion of the

population in malaria-endemic countries has asymptomatic parasitaemia [16]. Therefore, development of enhanced *Plasmodium* blood screening test is advocated to prevent spread by transfusion in blood banks and in ROK, ELISA has been developed to screen and monitor transfusion-transmitted malaria [27,28]. There are two main aspects to keep in mind when considering malaria risk and transfusion, such as the malaria risk associated with any individual donor and the ability of the systems to identify and manage the donor [29]. The development of tests that are able to screen and identify donors potentially infectious for parasitic infections without causing the deferral




Figure 4 FLISA performance to diagnosis malaria infection in patient samples. *P. falciparum* ($n = 9$)- and *P. vivax* ($n = 9$)-infected patients' bloods were applied to FLISA to evaluate the sensitivity of FLISA. As non-malaria specimens, healthy donors ($n = 9$), *T. gondii* ($n = 9$)-, and amoeba ($n = 9$)-infected patient bloods were tested to determine the specificity. Red dotted line indicates the cut-off value between positive- and negative values of FLISA.

of a large number of non-infectious donors is considered as one goal for prevention [30]. For this reason, coumarin-derived dendrimer-based FLISA assay was developed and compared to conventional ELISA. Because fluorescent density was highly correlated with parasite numbers, FLISA is applicable for quantification of parasite numbers in addition to diagnosis of *Plasmodium* infection. Therefore, FLISA has a potential as malarial diagnostic to facilitate the accurate diagnosis of *Plasmodium* infection and it would be an apparent alternative assay to determine *Plasmodium* parasite density. Because fluorescent intensity is correlated with malaria density with high sensitivity, FLISA may be considered as a feasible solution for high-throughput analysis of malaria infection during blood transfusion in ROK. Further studies should assess the diagnostic performance for other *Plasmodium* spp. infection and asymptomatic parasitaemia patients.

Competing interests

The authors have declared that they have no competing interests.

Authors' contributions

SJ, DT, JH, JY, WJ, and HJ carried out the assays and analysed the results; SJ, ET and HP wrote the manuscript; SJ, WJ, HJ, ET, and HP conceived and designed the experiments. All authors read and approved the final manuscript.

Acknowledgements

This research was supported by a Research fund (2013E5400800) of the Korea Centers for Disease Control and Prevention and by the Public Welfare & Safety Research Program through the National Research foundation of Korea (NRF) funded by the Ministry of Science, ICT & Future Planning (2010-0020803).

Author details

¹Zoonosis Research Center, Department of Infection Biology, School of Medicine, Wonkwang University, Iksan, Jeonbuk 570-749, Republic of Korea.
²Department of Medical Environmental Biology and Tropical Medicine, School of Medicine, Kangwon National University, Chuncheon, Gangwon-do, Republic of Korea. ³Division of Malaria and Parasitic Disease, Korea National Institute of Health, Osong 363-951, Republic of Korea. ⁴Department of Microbiology, Ajou University School of medicine, Suwon 443-721, Republic of Korea.

Received: 20 May 2014 Accepted: 4 July 2014

Published: 10 July 2014

References

- WHO: World Malaria Report. [<http://www.who.int/campaigns/malaria-day/2013/en/index.html>]
- Dhingra N, Jha P, Sharma VP, Cohen AA, Jotkar RM: Adult and child malaria mortality in India: a nationally representative mortality survey. *Lancet* 2010, **376**:1768–1774.
- Jha P: Reliable mortality data: a powerful tool for public health. *Natl Med J India* 2001, **14**(suppl 3):129–131.
- Jha P, Gajalakshmi V, Gupta PC: Prospective study of one million deaths in India: rationale, design, and validation results. *PLoS Med* 2006, **3**(suppl 2):e18.
- Johnston SP, Pieniazek NJ, Xayavong MV, Slemenda SB, Wilkins PP, da Silva AJ: PCR as a confirmatory technique for laboratory diagnosis of malaria. *J Clin Microbiol* 2006, **44**:1087–1089.
- Bejon P, Andrews L, Hunt-Cooke A, Sanderson F, Gilbert SC, Hill AV: Thick blood film examination for *Plasmodium falciparum* malaria has reduced sensitivity and underestimates parasite density. *Malar J* 2006, **8**:104.
- Haanshuus CG, Mohn SC, Mørch K, Langeland N, Blomberg B, Hanevik K: A novel, single-amplification PCR targeting mitochondrial genome highly sensitive and specific in diagnosing malaria among returned travellers in Bergen, Norway. *Malar J* 2013, **12**:26.
- Milne LM, Kyi MS, Chiodini PL, Warhurst DC: Accuracy of routine laboratory diagnosis of malaria in the United Kingdom. *J Clin Pathol* 1994, **47**:740–742.
- Mangold KA, Manson RU, Koay ES, Stephens L, Regner M, Thomson RB Jr, Peterson LR, Kaul KL: Real-time PCR for detection and identification of *Plasmodium* spp. *J Clin Microbiol* 2005, **43**:2435–2440.
- Kamau E, Tolbert LS, Kortepeter L, Pratt M, Nyakoe N, Muringo L, Ongutu B, Waitumbi JN, Ockenhouse CF: Development of a highly sensitive genus-specific quantitative reverse transcriptase real-time PCR assay for detection and quantitation of plasmodium by amplifying RNA and DNA of the 18S rRNA genes. *J Clin Microbiol* 2011, **49**:2946–2953.
- Snounou G, Viriyakosol S, Jarra W, Thaithong S, Brown KN: Identification of the four human malaria parasite species in field samples by the polymerase chain reaction and detection of a high prevalence of mixed infections. *Mol Biochem Parasitol* 1993, **58**:283–292.
- Noedl H, Yingyuen K, Laoboonchai A, Fukuda M, Sirichaisinthop J, Miller RS: Sensitivity and specificity of an antigen detection ELISA for malaria diagnosis. *Am J Trop Med Hyg* 2006, **75**:1205–1208.
- Song HO, Lee B, Bhushal RP, Park B, Yu K, Chong CK, Cho P, Kim SY, Kim HS, Park H: Development of a novel fluorophore for real-time biomonitoring system. *PLoS One* 2012, **7**:e48459.
- Mendoza NM, Cucunubá ZM, Aponte S, González NE, Bernal SD: Field evaluation for diagnostic accuracy of the rapid test SD Bioline Malaria Antigen Pf/Pv® in Colombia. *Biomedica* 2013, **33**:587–597.
- Coleman RE, Maneechai N, Rachaphaew N, Kumpitak C, Miller RS, Soyseng V, Thimasarn K, Sattabongkot J: Comparison of field and expert laboratory microscopy for active surveillance for asymptomatic *Plasmodium falciparum* and *Plasmodium vivax* in western Thailand. *Am J Trop Med Hyg* 2002, **67**:141–144.
- Nansue JR, Noubiap JJ, Ndoula ST, Zeh AF, Monamele CG: What is the best strategy for the prevention of transfusion-transmitted malaria in sub-Saharan African countries where malaria is endemic? *Malar J* 2013, **12**:465.
- Kim JH, Lee J, Sohn HJ, Song HO, Kim JY, Lee WJ, Park H, Shin HJ: Production of monoclonal antibodies for *Plasmodium vivax* lactate dehydrogenase and patient sera screening using sandwich ELISA. *Parasitol Res* 2012, **111**:1645–1650.
- Miao J, Cui L: Rapid isolation of single malaria parasite-infected red blood cells by cell sorting. *Nat Protoc* 2011, **6**:140–146.
- Sousa LP, Mariuba LA, Holanda RJ, Pimentel JP, Almeida ME, Chaves YO, Borges D, Lima E, Crainey JL, Orlandi PP, Lacerda MV, Nogueira PA: A novel polyclonal antibody-based sandwich ELISA for detection of *Plasmodium vivax* developed from two lactate dehydrogenase protein segments. *BMC Infect Dis* 2014, **14**:49.
- Doderer C, Heschung A, Guntz P, Cazenave JP, Hansmann Y, Senegas A, Pfaff AW, Abdelrahman T, Candolfi E: A new ELISA kit which uses a combination of *Plasmodium falciparum* extract and recombinant *Plasmodium vivax* antigens as an alternative to IFAT for detection of malaria antibodies. *Malar J* 2007, **6**:19.
- Martin SK, Rajasekariah GH, Awinda G, Waitumbi J, Kifude C: Unified parasite lactate dehydrogenase and histidine-rich protein ELISA for quantification of *Plasmodium falciparum*. *Am J Trop Med Hyg* 2009, **80**:516–522.
- Brown WM, Yowell CA, Hoard A, Vander Jagt TA, Hunsaker LA, Deck LM: Comparative structural analysis and kinetic properties of lactate dehydrogenases from the four species of human malarial parasites. *Biochemistry* 2004, **43**:6219–6229.
- Korea Centers for Disease Control & Prevention. http://cdc.go.kr/CDC/together/CdcKrTogether0302.jsp?menulds=HOME001-MNU1154-MNU0004-MNU0088&fid=51&q_type=&q_value=&cid=9885&pageNum=1.
- Global Health Group Project team: Eliminating malaria in the Democratic People's Republic of Korea. San Francisco: UCSF; 2012.
- Park JW, Jun G, Yeom JS: *Plasmodium vivax* malaria: status in the Republic of Korea following reemergence. *Korean J Parasitol* 2009, **47**(Suppl):S39–50.
- Cho YH, Kwon SY, Seo DH, Kim SI: Transfusion transmitted malaria in Korea. *Korean J Blood Transfus* 2001, **12**:263–270.
- Nam MH, Kim JS, Cho CH, Han ET, Lee WJ, Lee HK, An SS, Lim CS, Lee KN: Evaluation of *Plasmodium vivax* ELISA for the blood screen. *Trop Med Int Health* 2010, **15**:1436–1441.

28. Oh JS, Kim JS, Lee CH, Nam DH, Kim SH, Park DW, Lee CK, Lim CS, Park GH: Evaluation of a malaria antibody enzyme immunoassay for use in blood screening. *Mem Inst Oswaldo Cruz* 2008, **103**:75–78.
29. Kitchen AD, Chiodini PL: Malaria and blood transfusion. *Vox Sang* 2006, **90**:77–84.
30. Singh G, Sehgal R: Transfusion-transmitted parasitic infections. *Asian J Transfus Sci* 2010, **4**:73–77.

doi:10.1186/1475-2875-13-266

Cite this article as: Yeo et al.: Performance of coumarin-derived dendrimer-based fluorescence-linked immunosorbent assay (FLISA) to detect malaria antigen. *Malaria Journal* 2014 13:266.

**Submit your next manuscript to BioMed Central
and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

