

Research Article

Inability of Some Commercial Assays to Measure Suppression of Glucagon Secretion

**Nicolai J. Wewer Albrechtsen,^{1,2} Simon Veedfald,^{1,2}
Astrid Plamboeck,^{1,2} Carolyn F. Deacon,^{1,2} Bolette Hartmann,^{1,2}
Filip K. Knop,^{1,2,3} Tina Vilsboll,³ and Jens J. Holst^{1,2}**

¹*Department of Biomedical Sciences, Faculty of Health and Medical Sciences, University of Copenhagen, 2200 Copenhagen, Denmark*

²*Novo Nordisk Foundation Center for Basic Metabolic Research, Faculty of Health and Medical Sciences, University of Copenhagen, 2200 Copenhagen, Denmark*

³*Center for Diabetes Research, Gentofte Hospital, University of Copenhagen, 2900 Hellerup, Denmark*

Correspondence should be addressed to Jens J. Holst; jjholst@sund.ku.dk

Received 4 September 2015; Revised 21 October 2015; Accepted 22 October 2015

Academic Editor: Raffaele Marfella

Copyright © 2016 Nicolai J. Wewer Albrechtsen et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Glucagon levels are increasingly being included as endpoints in clinical study design and more than 400 current diabetes-related clinical trials have glucagon as an outcome measure. The reliability of immune-based technologies used to measure endogenous glucagon concentrations is, therefore, important. We studied the ability of immunoassays based on four different technologies to detect changes in levels of glucagon under conditions where glucagon levels are strongly suppressed. To our surprise, the most advanced technological methods, employing electrochemiluminescence or homogeneous time resolved fluorescence (HTRF) detection, were not capable of detecting the suppression induced by a glucose clamp (6 mmol/L) with or without atropine in five healthy male participants, whereas a radioimmunoassay and a spectrophotometry-based ELISA were. In summary, measurement of glucagon is challenging even when state-of-the-art immune-based technologies are used. Clinical researchers using glucagon as outcome measures may need to reconsider the validity of their chosen glucagon assay. The current study demonstrates that the most advanced approach is not necessarily the best when measuring a low-abundant peptide such as glucagon in humans.

1. Introduction

Glucagon, a 29-amino-acid peptide secreted from the pancreatic alpha cells in response to hypoglycemia [1], is derived from the proglucagon molecule, which is also expressed in the intestine and brain [2]. Glucagon has stimulatory effect on hepatic glucose production, and dysregulation of its secretion may contribute to the development of diabetes [3–6]. Glucagon measurements are, therefore, often an important study outcome; according to clinicaltrials.gov, it is included as an endpoint in more than 400 clinical studies. However, measurement of glucagon is a delicate matter and the validity of the data relies on sufficient specificity and sensitivity of the assay. Differential tissue-specific processing of proglucagon results in molecular heterogeneity, meaning

that assay specificity with respect to the different molecular forms is important. Thus, in addition to glucagon itself, proglucagon gives rise to several peptides containing the glucagon sequence, including oxyntomodulin, glicentin, and proglucagon 1–61, as well as molecules with some sequence homology to glucagon, including glucagon-like peptide-1 (GLP-1) and glucagon-like peptide-2 (GLP-2) and major proglucagon fragment [7]. Furthermore, each of these molecular forms may occur in extended or truncated forms, which may or may not be biologically active [2]. The immediate specificity problem is therefore of considerable magnitude. Sensitivity is equally important, since glucagon occurs in low picomolar concentrations in the circulation. Its concentration rises in response to hypoglycemia and falls in response to rising glucose (e.g., after carbohydrate meals), with the rate

of as well as the absolute magnitude of the decrease being of considerable importance for the resulting glucose tolerance. The ability of assays to register these decreases from already low levels is, therefore, critical [8].

In the current study, we investigated assays based on four widely applied immune-based technologies: a radioimmunoassay (RIA), a spectrophotometric enzyme-linked immunoassay (ELISA), and ELISAs based on electrochemiluminescence (ECL), and homogeneous time-resolved fluorescence (HTRF) detection. We hypothesized that the assay type might influence measured glucagon concentrations. To address this, we analyzed glucagon levels during a glucose clamp with or without atropine (atropine blocks cholinergic signaling through the muscarinic receptors and leads to further suppression of glucagon secretion) in five healthy male participants using these four different approaches; previous measurements indicated that the clamp + atropine protocol resulted in pronounced suppression of glucagon levels [9].

2. Methods

2.1. Participants, Procedures, and Samples. Samples were derived from a previously published study by Plamboeck et al. [9]. The study was conducted in accordance with the Helsinki Declaration II and was approved by the Scientific-Ethical Committee of the Capital Region of Denmark (registration number: H-2-2011-062) and by the Danish Data Protection Agency (journal number: 2011-41-6381) and registered at clinicaltrials.gov (ID: NCT01534442). Oral and written informed consent was obtained from all participants. Glucose clamps (6 mmol/L) were performed in five healthy male participants (age: 25 ± 1 years, body mass index: 24 ± 0.5 kg/m², and HbA_{1c}: $5.1 \pm 1\%$) with or without blocking efferent muscarinic activity by infusion of atropine (1 mg bolus + an 80 ng/kg/min infusion). Samples were collected and stored using optimal conditions for glucagon analysis as described previously [8].

2.2. Measurement of Glucagon. We used four immune-based assays for measurement of glucagon: (A) an in-house C-terminal RIA (codename 4305) [6, 8, 10]; (B) Mercodia sandwich ELISA (spectrophotometry) (cat# 10-1271-01, Uppsala, Sweden); (C) sandwich ELISA from MSD (chemiluminescence) (cat# K151HCC-1, MD 21201, USA); and (D) sandwich ELISA from Cis-Bio (homogeneous time-resolved fluorescence) (cat# 62GLCPEK, Codolet, France). Assays were carried out as per protocol according to the manufacturers' instructions. Samples were kept cold (ice-bath) at all times, and all samples were measured simultaneously in a single run to eliminate interassay variance.

2.3. Statistics. To analyze changes in glucagon levels over time, a one-way ANOVA for repeated measurements followed by a Bonferroni post hoc analysis was performed for each of the four assays. To compare the ability of the assays to detect changes in glucagon levels, we created a generalized regression model (ANCOVA) with glucagon as dependent variable and time (minutes) and method (assay) as independent variables. Net area under the curve (delta changes from

time zero to 160 minutes relative to the individual baselines) (nAUC) was calculated using the trapezoidal rule and differences were tested using a two-sided test. A power calculation was made based on the following assumptions: normality of data distribution, homoscedasticity, one-sample *t*-test, quantification limits and coefficient of variations *provided by the manufacturers*, an alpha value of 0.05, and a sample size of five. The calculation showed that the power to detect significant changes (of 5%) in glucagon levels ranged from 79% to 84% across the four assays. $P < 0.05$ was considered significant. Calculations were made using GraphPad Prism version 6.04 for Windows, GraphPad Software, La Jolla, California, USA, <http://www.graphpad.com/>, and STAT14, Boston, MA, USA. For illustrations we used the Adobe CS6 software suite (California, USA).

3. Results

The recoveries of synthetic glucagon in pooled human plasma ($N = 4$) were $95 \pm 11\%$ (assay A), $104 \pm 5\%$ (assay B), $75 \pm 15\%$ (assay C), and $67 \pm 21\%$ (assay D). Glucagon levels dropped significantly compared to baseline (time = 0 min) in both saline and atropine treated groups ($P < 0.01$) when samples were measured using assay A (Figure 1(a)) and assay B (Figure 1(b)) but not with assay C (Figure 1(c), $P = 0.31$) and assay D (Figure 1(d), $P = 0.24$). Assay A was significantly different ($P < 0.05$) from assays C and D but not assay B ($P = 0.43$). Assay B was significantly different from assays C and D ($P < 0.05$) whereas there were no differences between assays C and D ($P = 0.27$). nAUCs during infusion of saline and atropine, respectively, for assay A and assay B were significantly different ($P < 0.001$), indicating further suppression of glucagon secretion with atropine addition. For assays C and D, nAUCs were significantly different between atropine ($P < 0.01$) and saline, indicating that atropine weakly suppressed glucagon levels compared to the clamp alone, where nAUCs did not show significance compared to baseline (zero) by one-side *t*-test ($P = 0.11$ and $P = 0.17$).

4. Discussion

Immune-based detection methods utilize the extreme binding energy of antibodies which may have equilibrium constants reaching values of 10^{12} L/mol, providing these methods with a potential to measure very low concentrations. However, the use of antibodies relies on their specificity and the antigen-antibody reaction may also be sensitive to the so-called matrix effects, that is, interference from components in plasma including a variety of high-abundant plasma molecules or proteins (e.g., albumin and immunoglobulins), leading to unspecific interference in antibody-antigen interaction [11].

Assay D uses the homogeneous time-resolved fluorescence technology which combines fluorescence resonance energy transfer technology (FRET) with time-resolved measurement (TR) [12]. HTRF is mainly used in (*in vitro*) primary




FIGURE 1: Plasma glucagon levels of five healthy participants during a 6 mmol/L glucose clamp with simultaneous infusion of either saline (black) or atropine (red). (a) depicts assay A, a radioimmunoassay; (b) depicts assay B, a spectrophotometrically based ELISA; (c) depicts assay C, a chemiluminescence based ELISA; and (d) depicts assay D, a homogeneous time-resolved fluorescence based ELISA. Net area under the curve (nAUC) is depicted separately at upper right quadrant on (a), (b), (c), and (d). * represents a significant two-sided *t*-test comparing saline nAUC to atropine nAUC. Data illustrated as mean \pm standard deviation.

and secondary screening phases of drug development. However, its usefulness in highly sensitive immunoassays required for detection of 1 pmol/L differences in glucagon levels may be questioned. In contrast, assay C applies an electrochemiluminescence approach: when excited by electrical stimulation, labeled molecules emit light, which then is detected by cameras. The most generic ELISA used in our study is assay B, involving spectrophotometry detection; a chromogenic substrate is added to sample wells, which is then cleaved by an enzyme, for example, horseradish peroxidase, coupled to the detection antibody. Assay A is a radioimmunoassay utilizing competition between radioactively labeled antigen and unlabeled antigen (peptide standard or sample with unknown concentration) for binding to a limited number of specific antibody binding sites. Although it is the most simple, with regard to technology, the data clearly shows that assays A and B perform significantly better than assays C

and D. However, a requirement for such a performance is the application of antibodies with sufficient binding energy (and specificity), which is often the crucial step in assay development.

In this study, we highlight the crucial importance of choosing the “right” immune-based method when analyzing endogenous glucagon. We have exemplified the challenge by demonstrating that changes in measured glucagon levels depend on the assay used (Figure 1); where assays A and B clearly register the attenuation of glucagon levels during a glucose clamp, assays C and D did not. The basal levels measured were comparable, around 15 pmol/L (although assay C showed slightly higher levels, around 20 pmol/L). This may reflect a specificity problem in assay C; for example, this assay could be cross-reacting with glucagon-like molecules (oxyntomodulin, glicentin, or glucagon-like peptide-1 [7, 11]) although not stated by the manufacturers. Otherwise,

the difference between the assays, although not formally tested here, most likely reflects differences in sensitivity and therefore ability to detect dynamic changes in the very low picomolar range. In a previous study [6], a sensitivity analysis was carried out for assays A and B (where assay A was more sensitive), and, in another study of other commercially available assays [7], sensitivity was clearly insufficient to allow analysis in this concentration range. In addition, we recently demonstrated that measured glucagon levels in subjects with renal dysfunction may erroneously appear elevated, probably due to cross-reactions with N-terminal elongated inactive isoforms of the glucagon molecule (1–61) when analyzed with conventional single antibody C-terminal radioimmunoassays [6]. Importantly for interpretation of clinical studies, the potential instability of glucagon during inappropriate sample preparation and storage, such as multiple freeze-thaw cycles or storing plasma samples at room temperature for more than 1 hour, should also be considered [8].

Novel mass-spectrometry based detection (e.g., selected reaction monitoring (SRM)) of low-abundant peptides as glucagon [13] may in the future facilitate validation of immune-based detection methods. Unfortunately, current mass-spectrometry based methods still depend on 2D-gel extraction techniques or bead-coupled antibodies [14] both of which have questionable recoveries and specificity. However, in the future mass-spectrometry based detection methods may involve label-free (be it chemical or antibody based) purification steps as recently demonstrated [15] and may therefore provide accurate validation.

In conclusion, levels of glucagon are increasingly being used as outcome measures in clinical trials and the reliability of the glucagon assays employed is therefore critical for appropriate interpretation of the data.

Conflict of Interests

Cis-Bio provided assay kits without imposing restrictions on the study design or interpretation of results. Otherwise the authors have no dualities of interest to declare.

Authors' Contribution

Nicolai J. Wewer Albrechtsen, Simon Veedfald, Astrid Plamboeck, Carolyn F. Deacon, Filip K. Knop, Tina Vilsboll, Bolette Hartmann, and Jens J. Holst provided substantial contribution to the concept and design; Nicolai J. Wewer Albrechtsen, Simon Veedfald, and Jens J. Holst substantially contributed to analysis and interpretation of data; Nicolai J. Wewer Albrechtsen, Simon Veedfald, and Jens J. Holst drafted the paper; Astrid Plamboeck, Carolyn F. Deacon, Bolette Hartmann, Tina Vilsboll, and Filip K. Knop revised the paper critically for important intellectual content. All authors have provided final approval of the version to be published. Jens J. Holst is responsible for the integrity of the work as a whole. Nicolai J. Wewer Albrechtsen and Simon Veedfald contributed equally to the study and are co-first authors.

Acknowledgments

The authors acknowledge support from NNF Center for Basic Metabolic Research, University of Copenhagen, NNF Application no. 13563 (Novo Nordisk Foundation, Denmark), The Danish Council for Independent Research (DFR, 1333-00206A), Augustinus Foundation 14-0962, European Molecular Biology Organisation, and the European Foundation for the Study of Diabetes.

References

- [1] P. E. Cryer, "Mechanisms of hypoglycemia-associated autonomic failure in diabetes," *The New England Journal of Medicine*, vol. 369, no. 4, pp. 362–372, 2013.
- [2] J. E. Campbell and D. J. Drucker, "Islet α cells and glucagon—critical regulators of energy homeostasis," *Nature Reviews Endocrinology*, vol. 11, no. 6, pp. 329–338, 2015.
- [3] F. K. Knop, T. Vilsbøll, S. Madsbad, J. J. Holst, and T. Krarup, "Inappropriate suppression of glucagon during OGTT but not during isoglycaemic i.v. glucose infusion contributes to the reduced incretin effect in type 2 diabetes mellitus," *Diabetologia*, vol. 50, no. 4, pp. 797–805, 2007.
- [4] P. Raskin and R. H. Unger, "Hyperglucagonemia and its suppression," *The New England Journal of Medicine*, vol. 299, no. 9, pp. 433–436, 1978.
- [5] P. Shah, A. Vella, A. Basu, R. Basu, W. F. Schwenk, and R. A. Rizza, "Lack of suppression of glucagon contributes to postprandial hyperglycemia in subjects with type 2 diabetes mellitus," *The Journal of Clinical Endocrinology & Metabolism*, vol. 85, no. 11, pp. 4053–4059, 2000.
- [6] N. J. Wewer Albrechtsen, B. Hartmann, S. Veedfald et al., "Hyperglucagonaemia analysed by glucagon sandwich ELISA: nonspecific interference or truly elevated levels?" *Diabetologia*, vol. 57, no. 9, pp. 1919–1926, 2014.
- [7] M. J. Bak, N. W. Albrechtsen, J. Pedersen et al., "Specificity and sensitivity of commercially available assays for glucagon and oxyntomodulin measurement in humans," *European Journal of Endocrinology*, vol. 170, no. 4, pp. 529–538, 2014.
- [8] N. J. Wewer Albrechtsen, M. J. Bak, B. Hartmann et al., "Stability of glucagon-like peptide 1 and glucagon in human plasma," *Endocrine Connections*, vol. 4, no. 1, pp. 50–57, 2015.
- [9] A. Plamboeck, S. Veedfald, C. F. Deacon et al., "The role of efferent cholinergic transmission for the insulinotropic and glucagonostatic effects of GLP-1," *American Journal of Physiology—Regulatory, Integrative and Comparative Physiology*, vol. 309, no. 5, pp. R544–R551, 2015.
- [10] J. J. Holst, "Evidence that glicentin contains the entire sequence of glucagon," *The Biochemical Journal*, vol. 187, no. 2, pp. 337–343, 1980.
- [11] R. E. Kuhre, N. J. W. Albrechtsen, B. Hartmann, C. F. Deacon, and J. J. Holst, "Measurement of the incretin hormones: glucagon-like peptide-1 and glucose-dependent insulinotropic peptide," *Journal of Diabetes and Its Complications*, vol. 29, no. 3, pp. 445–450, 2015.
- [12] F. Degorce, A. Card, S. Soh, E. Trinquet, G. P. Knapik, and B. Xie, "HTRF: a technology tailored for drug discovery—a review of theoretical aspects and recent applications," *Current Chemical Genomics*, vol. 3, no. 1, pp. 22–32, 2009.
- [13] J. W. Howard, R. G. Kay, T. Tan et al., "Development of a high-throughput UHPLC–MS/MS (SRM) method for the

quantitation of endogenous glucagon from human plasma,” *Bioanalysis*, vol. 6, no. 24, pp. 3295–3309, 2014.

- [14] A. Y. H. Lee, D. L. Chappell, M. J. Bak et al., “Multiplexed quantification of proglucagon-derived peptides by immunoaffinity enrichment and tandem mass spectrometry after a meal tolerance test,” *Clinical Chemistry*, 2015.
- [15] N. J. Wewer Albrechtsen, D. Hornburg, S. Torang, R. E. Kuhre, C. F. Deacon, F. Meissner et al., “A novel immune-based approach for measurement of the anorectic gut hormone oxyntomodulin: changes after gastric bypass surgery,” *Diabetes*, 2015.