



Since January 2020 Elsevier has created a COVID-19 resource centre with free information in English and Mandarin on the novel coronavirus COVID-19. The COVID-19 resource centre is hosted on Elsevier Connect, the company's public news and information website.

Elsevier hereby grants permission to make all its COVID-19-related research that is available on the COVID-19 resource centre - including this research content - immediately available in PubMed Central and other publicly funded repositories, such as the WHO COVID database with rights for unrestricted research re-use and analyses in any form or by any means with acknowledgement of the original source. These permissions are granted for free by Elsevier for as long as the COVID-19 resource centre remains active.



Diagnosis of SARS-CoV-2 infection from breath - a proof-of-concept study

Bulemba Katende , Lucia González Fernández ,  
Josephine Muhairwe , Tracy R Glass , Irene Ayakaka ,  
Niklaus D Labhardt , Morten Ruhwald , Margaretha de Vos ,  
Moniek Bresser , Klaus Reither

PII: S2772-7076(22)00133-3  
DOI: <https://doi.org/10.1016/j.ijregi.2022.10.007>  
Reference: IJREGI 178

To appear in: *IJID Regions*

Received date: 27 October 2022  
Accepted date: 28 October 2022

Please cite this article as: Bulemba Katende , Lucia González Fernández , Josephine Muhairwe , Tracy R Glass , Irene Ayakaka , Niklaus D Labhardt , Morten Ruhwald , Margaretha de Vos , Moniek Bresser , Klaus Reither , Diagnosis of SARS-CoV-2 infection from breath - a proof-of-concept study, *IJID Regions* (2022), doi: <https://doi.org/10.1016/j.ijregi.2022.10.007>

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2022 Published by Elsevier Ltd on behalf of International Society for Infectious Diseases.  
This is an open access article under the CC BY-NC-ND license  
(<http://creativecommons.org/licenses/by-nc-nd/4.0/>)

**Short communication**

**Diagnosis of SARS-CoV-2 infection from breath - a proof-of-concept study**

Bulemba Katende<sup>1\*</sup>, Lucia González Fernández<sup>2,3,4\*</sup>, Josephine Muhairwe<sup>1</sup>, Tracy R Glass<sup>3,4</sup>,  
Irene Ayakaka<sup>1</sup>, Niklaus D Labhardt<sup>3,4,5</sup>, Morten Ruhwald<sup>6</sup>, Margaretha de Vos<sup>6</sup>, Moniek  
Bresser<sup>3,4</sup>, Klaus Reither<sup>3,4</sup>

<sup>1</sup> SolidarMed, Partnerships for Health, Maseru, Lesotho

<sup>2</sup> SolidarMed, Partnerships for Health, Lucerne, Switzerland

<sup>3</sup> Swiss Tropical and Public Health Institute, Basel, Switzerland

<sup>4</sup> University of Basel, Basel, Switzerland

<sup>5</sup> Department of Infectious Diseases and Hospital Epidemiology, University Hospital Basel,  
Basel, Switzerland

<sup>6</sup> FIND, the global alliance for diagnostics, Geneva, Switzerland

\*contributed equally

**Correspondence to:**

Klaus Reither, Clinical Research Unit, Swiss Tropical and Public Health Institute, Kreuzstrasse  
2, 4123 Allschwil, Switzerland, +41 61284 8967

**Keywords:**

SARS-CoV-2, COVID-19, bioaerosol capture device, Lesotho

**Highlights**

- Bioaerosol capture might become a new way of detecting SARS-CoV-2
- SARS-CoV-2 can be detected with the novel AL2 bioaerosol capture device
- Larger studies are needed to fully assess its diagnostic performance

**Abstract:**

Bioaerosol capture and analysis is emerging as a non-invasive diagnostic method for the detection of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2). In this proof-of-concept study conducted in Lesotho, we evaluated the novel and simple AL2 bioaerosol detection device in comparison to conventional nasopharyngeal sampling methods. We demonstrated for the first time that SARS-CoV-2 can be detected using the AL2 bioaerosol capture device. However, studies with a larger sample size are needed to further evaluate this bioaerosol capture device for the detection of SARS-CoV-2.

**Introduction:**

Bioaerosol capture and analysis has the potential to become a non-invasive diagnostic method for the detection of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) (Ricco et al., 2022) and other respiratory pathogens that are transmitted through respiratory droplets and aerosol particles (Drossinos et al., 2021, Jayaweera et al., 2020). While tests from nasopharyngeal or nasal mucosal swabs may be negative at an initial, already infectious stage of

infection, bioaerosol-based tests might allow more timely detection, treatment and isolation of infectious patients (Hawks et al., 2021, Jarvis and Kelley, 2021, Kucirka et al., 2020, Ma et al., 2021). We conducted a study in Lesotho evaluating a novel bioaerosol collection device in comparison to conventional nasopharyngeal sampling methods to investigate whether SARS-CoV-2 can be detected using this device.

### **Methods:**

This study took place during two waves of the COVID-19 pandemic (study periods: Sep 8-23, 2021 and Jan 6 – Feb 3, 2022) at St. Charles Mission Hospital Seboche, Lesotho. Surveillance data from South Africa, Lesotho's neighbour, strongly suggest that the Delta variant was predominant during the first study period and the Omicron variant during the second (Viana et al., 2022).

Persons aged  $\geq 18$  years with body temperature  $\geq 38^{\circ}\text{C}$  or at least one of 10 symptoms (fever/chills, cough, tiredness, dyspnea, sore throat, body pain, diarrhea, loss of taste/smell, recent weight loss, night sweats) or close contact to a probable or confirmed COVID-19 case in the last 14 days were eligible.

A novel bioaerosol capture device (Reconogen<sup>TM</sup>; AL2 Impact, Inc, US), consisting of a blow tube containing a removable capture disc made of inert polymers, served as index test. With the device pressed against the mouth or nostril, participants performed 20 deep breaths, 10 coughs, and a count to 20 and 10 exhalations through each nostril while closing the opposite nostril. Thereafter, the capture disc was pushed into a vial with 1mL Universal Transport Medium

buffer, Triton X-45 was added (only in first study period), vortexed and real-time SARS-CoV-2 PCR (Alinity m System, Abbott, USA) was performed.

As a reference standard, real-time SARS-CoV-2 PCR (SARS-CoV-2/2019-nCoV assay by Daan Gene /ABI 7500 platform (target: N gene and ORF1ab), Applied Biosystem, USA, first study period; Alinity m System (target: N gene and RdRp), Abbott, USA, second study period; Limit of detection: Daan Gene assay: 500 virus copies/ml; Alinity *m* assay: 100 virus copies/ml) was done from a nasopharyngeal swab (NP-PCR). Comparator was an antigen rapid diagnostic tests (STANDARD Q COVID-19 Ag Test, SD Biosensor, Republic of Korea) from a second nasopharyngeal swab (NP-RDT). Patients were asked to blow their nose before the nasopharyngeal swab.

## **Results:**

The median age of the 131 participants was 46 years (IQR 33-62), 63.4% (83) were female, 96.2% (126) had at least one symptom, and 3.1% (4) COVID-19 exposure only. None of the participants were severely ill (i.e. altered mental status, tachypnea, SpO<sub>2</sub><94%, or systolic blood pressure <100mmHg).

### ***Test performance during the first study period***

During this period, we recruited 55 participants. The bioaerosol capture device showed positive results in 9 participants, with 3 of these participants being NP-PCR-negative and NP-RDT-positive, and 3 being both NP-PCR- and NP-RDT-negative (Table 1). Overall, only 5 participants had a positive NP-PCR (positivity rate 9.1%). Using NP-PCR as reference standard

(as per protocol), the sensitivity of the bioaerosol capture device is 60.0% (95%CI: 14.7-94.7) and the specificity 88.0% (95%CI: 75.7-95.5).

### ***Test performance during the second study period***

During this period, we recruited 76 participants, 32 had a positive NP-PCR (positivity rate 42.1%), 10 participants had a positive result from the bioaerosol capture device, of which 7 matched the total of 8 positive NP-RDT results (Table 1). With NP-PCR as a reference standard (as per protocol), the sensitivity and the specificity of PCR from bioaerosol capture device is 31.3% (95%CI: 16.1-50.0) and 100.0% (95%CI: 92.0-100.0).

The performance results in relation to time from symptom onset in both study periods is shown in table 2.

### **Discussion:**

In this study, we provided evidence that SARS-CoV-2 can be detected with the AL2 bioaerosol capture device. Of note, the bioaerosol capture device identified 6 patients during the first study period who were not identified by NP-PCR, suggesting that negative NP-PCR results do not always rule out SARS-CoV-2 infection and it is a potentially imperfect reference standard (Mardian et al., 2021). Due to the small sample size, generalizable conclusions about diagnostic performance and relationship to other variables (symptom duration, Ct or CN score) or a recommendation for clinical use cannot be made at this time, and further studies with a larger sample size are needed to evaluate this bioaerosol capture device for SARS-CoV-2 detection.”

**Conflict of interest:**

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper. AL2 Impact and 3M Company provided the bioaerosol capture devices and gave technical advice, but were not involved in data analysis.

**Funding Source:**

This work was supported by Botnar Research Centre for Child Health (BRCCH) as part of the Multi-Investigator Project/Fast Track Call for Acute Global Health Challenges [grant numbers: DZX2167, DZX2168] and Foundation for Innovative New Diagnostics (FIND) [FIND/Swiss TPH Project Agreement 2/2021].

**Ethical Approval statement:**

The National Research and Ethics Committee of Lesotho (ID-107-2020 Modify 03) approved the study. All participants provided written informed consent.

**References:**

- Drossinos Y, Weber TP, Stilianakis NI. Droplets and aerosols: An artificial dichotomy in respiratory virus transmission. *Health Sci Rep* 2021;4(2):e275.
- Hawks SA, Prussin AJ, 2nd, Kuchinsky SC, Pan J, Marr LC, Duggal NK. Infectious SARS-CoV-2 Is Emitted in Aerosol Particles. *mBio* 2021;12(5):e0252721.
- Jarvis KF, Kelley JB. Temporal dynamics of viral load and false negative rate influence the levels of testing necessary to combat COVID-19 spread. *Sci Rep* 2021;11(1):9221.
- Jayaweera M, Perera H, Gunawardana B, Manatunge J. Transmission of COVID-19 virus by droplets and aerosols: A critical review on the unresolved dichotomy. *Environ Res* 2020;188:109819.



- Kucirka LM, Lauer SA, Laeyendecker O, Boon D, Lessler J. Variation in False-Negative Rate of Reverse Transcriptase Polymerase Chain Reaction-Based SARS-CoV-2 Tests by Time Since Exposure. *Ann Intern Med* 2020;173(4):262-7.
- Ma J, Qi X, Chen H, Li X, Zhang Z, Wang H, et al. Coronavirus Disease 2019 Patients in Earlier Stages Exhaled Millions of Severe Acute Respiratory Syndrome Coronavirus 2 Per Hour. *Clin Infect Dis* 2021;72(10):e652-e4.
- Mardian Y, Kosasih H, Karyana M, Neal A, Lau CY. Review of Current COVID-19 Diagnostics and Opportunities for Further Development. *Front Med (Lausanne)* 2021;8:615099.
- Riccò M, Zaniboni A, Satta E, Ranzieri S, Marchesi F. Potential Use of Exhaled Breath Condensate for Diagnosis of SARS-CoV-2 Infections: A Systematic Review and Meta-Analysis. *Diagnostics (Basel)* 2022;12(9).
- Viana R, Moyo S, Amoako DG, Tegally H, Scheepers C, Althaus CL, et al. Rapid epidemic expansion of the SARS-CoV-2 Omicron variant in southern Africa. *Nature* 2022;603(7902):679-86.

Bioaerosol capture device (PCR)	NP-RDT	NP-PCR	First study period	Second study period	Total
positive	positive	positive	3	7	10
positive	positive	negative	3	0	3
positive	negative	positive	0	3	3
positive	negative	negative	3	0	3
negative	positive	positive	2	1	3
negative	positive	negative	1	0	1
negative	negative	positive	0	21	21
negative	negative	negative	43	44	87

**Table 1: Overview of performance results in both study periods**

Days from symptom onset	Patients tested positive by any test	Bioaerosol-positive patients	NP-RDT-positive patients	NP-PCR-positive patients
0-3	8	5 (62.5%)	3 (37.5%)	6 (75.0%)

4-7	20	9 (45.0%)	8 (40.0%)	17 (85.0%)
8-14	8	4 (50.0%)	4 (50.0%)	7 (87.5%)
$\geq 15$	7	1 (14.3%)	1 (14.3%)	6 (85.7%)

**Table 2: Performance results in relation to time from symptom onset in both study periods. N=44 had at least one positive test but 1 of these was missing information on the time since onset of symptoms.**

Journal Pre-proof