

# Chaos-Based Analysis of Heart Rate Variability Time Series in Obstructive Sleep Apnea Subjects

## Abstract

Obstructive sleep apnea (OSA) is a common disorder which can cause periodic fluctuations in heart rate. To diagnose sleep apnea, some studies analyze electrocardiogram (ECG) signals by adopting chaos-based analysis. This research is going to specifically focus on whether it is possible to use chaos-based analysis of heart rate variability (HRV) signals rather than using chaotic analysis of ECG signals to diagnose OSA. While conventional studies mostly use chaos-based analysis of ECG signals to detect OSA, here, we apply correlation dimension (CD) as a chaotic index to analyze HRV data in OSA patients. For this purpose, 17 patients with OSA and 9 healthy individuals referred to a sleep clinic in Isfahan/Iran are studied, and their HRV time series were extracted from 1-h ECG signals recorded overnight. The preliminary step to calculate CD is phase-space reconstruction of the system based on HRV time series. Corresponding parameters, including embedding dimension and lag time, are estimated optimally using enhanced related methods, and then CD is calculated using Grassberger–Procaccia algorithm. Moreover, to evaluate our results, detrended fluctuation analysis (DFA), one of the well-known nonlinear methods in HRV analysis to detect OSA, is also applied to our data and the result is compared with those obtained from CD analysis of HRV. CD index with  $P < 0.005$  indicates a significant difference in nonlinear dynamics of HRV signals detected from OSA patients and healthy individuals.

**Keywords:** Chaotic indexes, correlation dimension, detrended fluctuation analysis, heart rate variability, obstructive sleep apnea

Submitted: 08-May-19

Revision: 13-Jun-19

Accepted: 13-Jul-19

Published: 06-Feb-20

## Introduction

Sleep apnea or pauses in breathing for more than 10 s followed by awakening, is one of the common sleep disorders. Sleep apnea is divided into three categories: obstructive sleep apnea (OSA), central sleep apnea, and mixed (complex) sleep apnea. OSA, the most common one, is caused by a local blockage or obstruction of the upper airway and is usually accompanied by exhale and deep inhale at the end of blockage. Obesity, aging, narrowed airways, smoking, and alcohol consumption are considered as the certain risk factors of OSA, and men are more likely to have sleep apnea.<sup>[1]</sup> Sleep apnea reduces the quality of life since it may cause depression, irritation, and lack of sleep. Investigations reveal that OSA is the reason of many car accidents and work events;<sup>[2,3]</sup> furthermore, according to the prevalence of cardiovascular diseases

among those suffering from sleep apnea, many studies have been done to explore the relation of hypertension and heart diseases to sleep apnea. Studies have shown that untreated OSA can lead to high blood pressure, cardiac arrhythmias, heart failure, and brain stroke as well.<sup>[4-6]</sup> Studies on OSA reveal that apnea is accompanied by periodic variations in heart rate.<sup>[7]</sup> Irregular heartbeats appear in the form of bradycardia (decrease in heart rate to  $<60$  times/min) during apnea, and in the form of tachycardia (increase in heart rate to more than 100 times/min) after that; moreover, according to the research done on sympathetic neural activity, sympathetic activation increases during apnea.<sup>[8]</sup> In patients with OSA, sympathetic activation continues during the day that can increase the risk of getting cardiovascular diseases.<sup>[9]</sup> There have always been many interests in finding simple and accurate methods for the study of or detection of OSA; accordingly, some research on detection of sleep

**Shiva Naghsh<sup>1</sup>,  
Mohammad Ataei<sup>1</sup>,  
Mohammadreza  
Yazdchi<sup>2</sup>,  
Mohammad  
Hashemi<sup>3</sup>**

Departments of <sup>1</sup>Electrical Engineering and <sup>2</sup>Biomedical Engineering, Faculty of Engineering, University of Isfahan, <sup>3</sup>Department of Cardiology, School of Medicine, Isfahan University of Medical Sciences, Isfahan, Iran

## Address for correspondence:

Dr. Mohammad Ataei,  
Department of Electrical Engineering, Faculty of Engineering, University of Isfahan, Isfahan, Iran.  
E-mail: ataei@eng.ui.ac.ir

## Access this article online

Website: [www.jmssjournal.net](http://www.jmssjournal.net)

DOI: 10.4103/jmss.JMSS\_23\_19

## Quick Response Code:



**How to cite this article:** Naghsh S, Ataei M, Yazdchi M, Hashemi M. Chaos-based analysis of heart rate variability time series in obstructive sleep apnea subjects. J Med Signals Sens 2020;10:53-9.

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: [reprints@medknow.com](mailto:reprints@medknow.com)

apnea has been already done using different biological signals such as electroencephalography signal<sup>[10]</sup> and electrocardiogram (ECG) signal with or without respiratory signal.<sup>[11,12]</sup> Furthermore, blood oxygen measurements<sup>[13]</sup> and feature assessments of acoustic signal produced by snoring<sup>[14]</sup> have been already used for sleep apnea detection. According to the studies,<sup>[7]</sup> heart rate variability (HRV), variation in the time intervals between successive heartbeats, has been also used efficiently to detect sleep apnea. HRV has become a conventional marker to detect different failures in autonomous nervous and cardiovascular system.<sup>[15]</sup> HRV signals are typically extracted from ECG signals by measuring the R–R intervals. Like other physiological signals, HRV can be analyzed using time/frequency-domain, nonlinear, and chaotic methods. Time-domain methods calculate the statistical features of R–R intervals. They can only provide the simplest insight into the cardiovascular system. Frequency-domain indexes, which are mostly based on spectral bands, can reveal the autonomic nervous system activities. However, nonlinear methods can describe hidden dynamics and some nonstationary conditions of the cardiovascular system.<sup>[16]</sup> The human heart acts like a nonlinear oscillator, and HRV is one observed variable from a complex system (cardiovascular system); therefore, complexities in the cardiovascular system can be revealed using nonlinear analysis of HRV.<sup>[17,18]</sup> Based on the study done by Voss *et al.*<sup>[19]</sup> due to nonlinear nature of biological signals, nonlinear techniques can be efficiently used for the analysis of these signals; moreover, in biological signal analysis, nonlinear methods can provide complementary information along with linear methods (time/frequency-domain methods). The conventional nonlinear methods used to analyze HRV in sleep studies include detrended fluctuation analysis (DFA), approximate entropy (ApEn), and sample entropy (SampEn) as well as chaotic methods such as recurrence quantification analysis (RQA), correlation dimension (CD), and Lyapunov exponent (LE). Some sleep studies focused on comparing the use of nonlinear methods with linear techniques for analysis of HRV in sleep apnea patients. In a study performed by Penzel *et al.*,<sup>[20]</sup> frequency-domain analysis and DFA were used to investigate the effect of sleep apnea on HRV. Their results suggest that in comparison with spectral analysis, DFA as a nonlinear method can better depict fluctuations caused by sleep apnea in HRV signal. In another study, Al-Angari and Sahakian<sup>[21]</sup> have applied SampEn as a nonlinear method to assess the complexity of HRV behavior due to OSA. According to their results, in healthy individuals, HRV pattern is considerably more complex than in patients with OSA. They also compared the results of the spectral analysis as a linear method with those obtained from SampEn. They claimed that although both methods are accurate, SampEn calculations are simpler in comparison with spectral methods.

Some noticeable nonlinear methods in sleep research are chaos-based analysis. CD is a well-known chaotic index

used in some sleep apnea studies. In a study, Miyata *et al.*<sup>[22]</sup> used CD to investigate nonlinear features and chaotic behavior of respiratory movement and breath to breath fluctuations in patients with OSA/hypopnea syndrome during wakefulness with eyes closed. They claimed that applying CD to respiratory movement can be used to efficiently identify patients with this syndrome during wakefulness. In their previous study,<sup>[23]</sup> they also showed using chaotic analysis of respiratory movement cannot be useful to detect OSA/hypopnea syndrome during apneic sleep. In a recent work, Moeynoi and Kitjaidure<sup>[24]</sup> explored RQA, CD, and DFA to extract nonlinear features from HRV signals and ECG derived respiratory signals to detect sleep apnea. However, their main focus of this work is on introducing canonical correlation analysis as a method for dimensional reduction of sleep apnea features extracted from ECG for better classification. In such works, discrimination abilities of chaotic methods have been rarely investigated in a specific way. These methods have been only used for classification purposes. In one study focused on chaotic analysis, Acharya *et al.*<sup>[25]</sup> applied nonlinear parameters including ApEn, fractal dimension, CD, largest LE, and Hurst exponent to ECG signals to distinguish sleep apnea patients. Nevertheless, in this study, phase-space reconstruction which is the primary step to calculate CD was performed using ECG signals as time series. In another study, Jafari<sup>[26]</sup> used nonlinear methods consisting of DFA, CD, large LE, and spectral entropy to detect OSA. He found that detrended fluctuation and CD are more significant for OSA detection among other methods he applied, but like the work done by Acharya *et al.*,<sup>[25]</sup> he used ECG signals for phase-space reconstruction. LE is another well-known chaotic index with the ability to detect chaotic behavior of biological signals. Jafari showed large LE is also a significant index for OSA detection. However, Zapanta *et al.*<sup>[27]</sup> introduced a new method called numerical titration technique instead of LE to detect chaos in HRV of children with OSA. They claimed that numerical titration technique could provide robust support for the presence of chaotic behavior in HRV. In spite of such studies worked on chaotic analysis in sleep research, up to our knowledge, sleep apnea studies have rarely focused on calculation chaotic indexes using HRV time series. As above mentioned, in the study by Jafari,<sup>[26]</sup> ECG signals are used as time series for the phase-space reconstruction as well as estimation of embedding dimension and lag time parameters, which are necessary parameters to calculate chaotic indexes; hence, this question arises that whether it is possible to use HRV of OSA patients rather than their ECG signals as time series to calculate chaotic indexes or not. Thus, the main aim of this study is specific concentration on the investigation of CD calculation using HRV as time series. In fact, while conventional studies such as the study by Acharya *et al.*<sup>[25]</sup> or by Jafari,<sup>[26]</sup> mostly use chaos-based analysis of ECG signals to study OSA, in this research, we concentrate to calculate CD using HRV signals instead of

using ECG signals. We are going to find out whether there is a significant difference in nonlinear dynamics of HRV signals between OSA patients and healthy individuals. We are also going to compare our results of CD calculated by HRV data as time series with the results obtained from DFA, which is one of the most commonly used nonlinear methods in sleep apnea studies.<sup>[20]</sup>

## Materials and Methods

### Data

In this research, the exploited data set is obtained from a cross-sectional study<sup>[28]</sup> on patients suffering from OSA. This data set contains both patients and normal individuals. The patient group consisted of individuals with untreated OSA who were referred to the Bamdad Sleep Clinic in Isfahan, Iran, and were diagnosed as mild-to-severe OSA, based on Apnea-Hypopnea Index ( $\geq 5$  events/h), and the age- and sex-matched control group consisted of healthy individuals without any sleep or cardiac complaints. All patients and normal individuals were visited and examined by a cardiologist, and those with a history or clinical evidence of heart failure, ischemic heart disease, cardiomyopathy, valvular heart disease, arrhythmia, persistent atrial fibrillation, bundle branch block, pericarditis, electrolyte abnormalities, renal failure, pulmonary disease, thyroid dysfunction, hypertension, diabetes, and use of medications affecting the electrocardiographic parameters, were not included into the study. The study protocol was approved by the Ethics Committee of the Isfahan University of Medical Sciences (Grant # 392101), and informed consent was provided by all patients and normal individuals. Patients had undergone an overnight polysomnographic study at the sleep clinic while all of them had effective sleep.

In this study, a 24-h electrocardiography was performed for patients and normal individuals by using a digital Holter recorder (H200 recorder, Kavoshgaran Teb Kharazmi Co, Iran), with six channels (three analogs I, II, and III and three digital, aVR, aVL, and aVF channels). Software was developed by the Biomedical Engineering Department of the University of Isfahan to measure ECG parameters. This software detects Q, R, S, P, and T points using gradient-based methods; then calculates important time intervals in ECG signal. RR intervals of ECGs achieved by H200 recorder for both patient and normal groups were used as an HRV data set in this research. The HRV time series extracted from 1-h ECG signals (recorded during the night) related to 17 patients with OSA and 9 healthy controls are analyzed.<sup>[7]</sup>

### Preprocessing

To analyze HRV related to sleep apnea, Penzel et al.<sup>[20]</sup> have selected an experimental feature for automated preprocessing of HRV time series or RR intervals. The feature is as follows:

0.33 s < each RR interval < 1.5 s.

The interval between two successive RR intervals is just 0.66 s.

If >1% of RR time intervals of a signal do not meet the above conditions, the signal cannot be used for future analysis. Of course, those HRV samples that exceed the above limit have to be discarded from the accepted signals.

Since the resulted signals were too noisy, after the evaluation of the above conditions, a median filter was applied to the signal to preserve the latent dynamics in noisy data.

### Correlation dimension

CD is a measure of the complexity of the system that shows variability and irregularity of the process. For calculation of CD as a chaotic index, reconstruction of a phase-space equivalent to the original phase space in topological aspects is the first important step. Takens' theory is often used for phase-space reconstruction,<sup>[29]</sup> that is called as the method of delays. In this method, for a time series derived from a deterministic dynamical system, the delay coordinates are used to form an  $m$ -dimensional vector space as follows:

$$x(t) = f(x(t-\tau), x(t-2\tau), \dots, x(t-(m-1)\tau)) \quad (1)$$

Where scalar  $m$  called embedding dimension and scalar  $\tau$  called time delay.

This theory only explains the possibility of phase-space reconstruction with two parameters, embedding dimension and time delay, but it does not provide any information about the way of acquiring these parameters. Time delay or  $\tau$  expresses the distance between components of the delay vector and embedding dimension or  $m$  expresses the components in each delay vector.

The success of CD analysis depends on suitable reconstruction of phase space that is based on the calculation of embedding dimension and lag time. There are several ways to obtain these parameters. In this study, the mutual information method is selected to achieve the optimized value for time delay;<sup>[30]</sup> Indeed, the mutual information ( $M(\tau)$ ) is obtained for different  $\tau$  values from (2).

$$M(\tau) = - \sum_{ij} p_{ij} \ln \frac{p_{ij}}{p_i p_j} \quad (2)$$

Where  $P_i$  is the probability of finding a data point in the distance subscripted by  $i$  and  $P_{ij}$  is the joint probability of an observation in the distance subscripted by  $i$  while the next observation lies in the distance subscripted by  $j$  after the time delay  $\tau$ . The first minimum of the  $M(\tau)$  is considered as the optimum time delay for phase-space reconstruction. Figure 1 shows  $M(\tau)$  calculated for several  $\tau$  values for the HRV signal of a patient with OSA.

For determination of the optimized embedding dimension, the false nearest neighbor method is exploited which is one of the most precise methods for the selection of optimum embedding dimension in reconstructing a low-dimension phase space.<sup>[31]</sup> In this method, optimum embedding dimension is selected as the first dimension for which the amount on nearest neighbors equals or approaches to zero.

After reconstruction of phase space, CD can be calculated using the Grassberger–Procaccia algorithm which in this regard is one of the well-known and optimized approaches. Relation or correlation between delay vectors extracted from phase-space reconstruction is measured with correlation integral. Correlation integral is an estimation of the probability of two delay vectors being at a distance less than  $r$ . Correlation integral is resulted from (3).

$$C(r) \approx \lim_{N \rightarrow \infty} \left( \frac{\sum_{i=1, j>i}^N \Theta(r - |x_i - x_j|)}{\frac{1}{2} N(N-1)} \right) \quad (3)$$

Where  $\Theta$  shows Heaviside function,  $N$  is the number of points of the time series, and  $r$  is the distance under study;  $x_i$  and  $x_j$  refer to the points located on the trajectories of the reconstructed phase space. Heaviside function is shown in (4).

$$\Theta(s) = \begin{cases} 1 & \text{if } s \geq 0 \\ 0 & \text{if } s < 0 \end{cases} \quad (4)$$

Then, CD is calculated using (5).

$$CD = \lim_{r \rightarrow \infty} \frac{\log C(r)}{\log r} \quad (5)$$

When  $\log C(r)$  is plotted as a function of  $\log r$ , three distinct regions appear: Null region, the region in which we could not find two points, located at a distance less than  $r$ ; Saturation region, the region in which  $C(r)$  does not grow

by increasing  $r$ ; and Scaling region, the region between null and saturation regions. CD equals to the slope of this curve in the third region.<sup>[32,33]</sup> Figure 2 shows  $\log C(r)$  as a function of  $\log r$  extracted from the HRV signal related to a patient with OSA.

### Detrended fluctuation analysis

DFA is used to determine the self-similarity and fractal structure in a time-series. In DFA, root mean square (RMS) fluctuation resulted from a detrended and integrated time series is measured using different frames. These values are plotted as a function of applied frames in a log-log plot. First, the time series  $y(k)$  with the total length of  $N$ , is integrated with the length of  $K \in N$  depicted in (6).

$$Y(k) = \sum_{i=1}^k [y(i) - y_m] \quad (6)$$

Where  $y(i)$  is data point number  $i$  from  $y(k)$  time series and  $y_m$  is the mean value of the time series. The integration process reveals that the main signal is nonstationary. After integration, the integrated time series is equally divided into subsets; each subset with length  $L$ . Then, a line obtained using the least square error method fits to data of each section and shows the trend of that section. After that, local trends related to each section are removed from the integrated time series  $Y(k)$ . For higher-order trend removal, a higher-order polynomial should fit to the time series. At last, RMS of the final time series is calculated as in (7) and (8).

$$f_L(v) = \frac{1}{L} \sum_{i=1}^L (Y((v-1)L + i) - Y_v(i))^2 \quad (7)$$

$$v = 1, 2, \dots, N_L$$

$$F(L) = \left( \frac{1}{N_L} \sum_{v=1}^{N_L} f_L(v) \right)^{\frac{1}{2}} \quad (8)$$

Where  $Y_v(k)$  shows lines fitted to each section (with length  $L$ ) of the time series  $Y(k)$ ,  $N_L$  shows the number of sections




Figure 1: Mutual information diagram for a heart rate variability signal from a patient with obstructive sleep apnea




Figure 2: The diagram of  $\log C(r)$  as a function of  $\log r$  for a heart rate variability signal from a patient with obstructive sleep



after division of time series into sections with length  $L$ . The above calculations should be continued until a meaningful relevance between  $F(L)$  and  $L$  is defined.  $F(L)$  typically grows in proportion with  $L$ . Slope of  $\log F(L)$  as a function of  $\log L$  achieves the self-similarity parameter called  $\alpha$ , which is the measurable parameter in DFA.<sup>[34]</sup>  $\alpha$  is calculated using (9).

$$\alpha = \frac{\partial F(L)}{\partial L} \quad (9)$$

## Results

The primary step toward calculations of CD is phase-space reconstruction of the system based on HRV time series. Two corresponding important parameters, including embedding dimension and lag time, are estimated optimally using enhanced related methods. The time delay is acquired through mutual information method. The first minimum of the mutual information diagram is considered as the optimum time delay for the phase-space reconstruction.  $M(\tau)$  is calculated for several  $\tau$  values for the HRV signal of a patient with OSA and is shown in Figure 1. Time delay for all the HRV signals are calculated and the dominant value ( $\tau = 20$ ) is selected for the phase-space reconstruction. Then, embedding dimension is achieved using the nearest neighbor method. The first dimension, for which the number of nearest neighbors approaches to 0, is selected as the embedding dimension. After the evaluation of embedding dimensions obtained from all HRV signals, the dominant value ( $m = 5$ ) is used for the phase-space reconstruction; therefore, CD is calculated using  $\tau = 20$  as the time delay and  $m = 5$  as the embedding dimension. Table 1 shows CDs calculated for HRV signals recorded from both patients with OSA and healthy individuals. The results of  $t$ -test analysis performed on CD values are statistically significant.<sup>[35]</sup> The CD with  $P < 0.005$  indicates that there is significant difference in nonlinear dynamics of HRV signals related to OSA patients and healthy individuals; indeed, our result indicates the possibility of using HRV data in sleep apnea as time series to calculate chaotic indexes.

To evaluate our results with another method, we also apply DFA to our data. In DFA, RMS should be repeatedly calculated for at least 30 frames. In this study, according to the work done by Penzel *et al.*,<sup>[20]</sup> 60 frames between  $L = 4$  and  $L = 1000$  are taken into account. Table 1 shows  $\alpha$  values obtained from DFA on HRV signals for patients

with OSA and healthy controls. As mentioned above,  $\alpha$  is the measurable parameter resulted from DFA. The results of  $t$ -test analysis performed on  $\alpha$  values are statistically significant. Our obtained results shown in Table 1 indicate that in comparison with DFA, CD as a chaotic method which is calculated using HRV time series can better quantify HRV changes.

## Discussion and Conclusion

Chaos-based methods can assist to realize hidden dynamical features of biological time series, so they can be used to detect serious disease such as sleep apnea. In most chaotic analysis methods such as CD, phase-space reconstruction is the first step. Currently, most of the sleep apnea studies calculating CD exploit ECG signals as time series for phase-space reconstruction. Up to our knowledge, studies on sleep apnea disorder rarely focused on independent investigation of the use of HRV as time series to perform phase-space reconstruction and to calculate CD. To this end, we attempted to use HRV data as time series to calculate CD. Statistical tests such as  $t$ -test typically are necessary to evaluate the discrimination ability of every feature extracted from time series analysis. Our results of using  $t$ -test on CD obtained from HRV time series with  $P = 0.003$  show a significant difference in nonlinear dynamics of HRV signals detected from OSA patients and healthy individuals, so this method has a very significant ability to distinguish OSA patients from normal people. A few studies on OSA using chaotic analysis can obtain practically significant results; for instance, Miyata *et al.*<sup>[22]</sup> obtained results with  $P < 0.01$  by applying CD to respiratory movement for diagnosis of OSA/hypopnea syndrome during wakefulness, or Zapanta *et al.*<sup>[27]</sup> got results with  $P < 0.05$  to detect chaos in HRV of children with OSA. Faust *et al.*<sup>[36]</sup> in their study claim that from research experience, features with low  $P$  value are more practical to discriminate patients from healthy people. Based on their claim, having features with low  $P$  value is necessary to design an automatic diagnosis system. Therefore, our statistical results with practical significance prove that the CD obtained from HRV time series can be used to explain an appropriate classifier for automatic detection of OSA. Design of a classifier based on distinguish ability of our method is another work that we already aim to propose. Moreover, we also used DFA (a commonly used method for OSA detection) to examine our HRV data. Comparing our obtained results from CD method with our results from DFA also indicates that HRV changes of sleep apnea patients can be efficiently quantified using CD. As shown in Table 1, our findings indicate more discrimination ability for CD than DFA. In conclusion, we attempted to specifically investigate HRV as time series to perform the phase-space reconstruction and to calculate CD as well, and our results as discussed above can be efficiently useful in study of OSA.

**Table 1: Results**

Features	Mean±SD		P
	Normal	OSA	
CD	2.696±0.194	2.933±0.184	0.003
$\alpha$	1.159±0.11	1.063±0.132	0.044

OSA – Obstructive sleep apnea; CD – Correlation dimension

## Financial support and sponsorship

None.

## Conflicts of interest

There are no conflicts of interest.

## References

- Epstein LJ, Kristo D, Strollo PJ Jr., Friedman N, Malhotra A, Patil SP, *et al.* Clinical guideline for the evaluation, management and long-term care of obstructive sleep apnea in adults. *J Clin Sleep Med* 2009;5:263-76.
- Findley LJ, Unverzagt ME, Suratt PM. Automobile accidents involving patients with obstructive sleep apnea. *Am Rev Respir Dis* 1988;138:337-40.
- Terán-Santos J, Jiménez-Gómez A, Cordero-Guevara J. The association between sleep apnea and the risk of traffic accidents. Cooperative group burgos-santander. *N Engl J Med* 1999;340:847-51.
- Shahar E, Whitney CW, Redline S, Lee ET, Newman AB, Nieto FJ, *et al.* Sleep-disordered breathing and cardiovascular disease: Cross-sectional results of the sleep heart health study. *Am J Respir Crit Care Med* 2001;163:19-25.
- Peppard PE, Young T, Palta M, Skatrud J. Prospective study of the association between sleep-disordered breathing and hypertension. *N Engl J Med* 2000;342:1378-84.
- Adlakha A, Shepard JW Jr. Cardiac arrhythmias during normal sleep and in obstructive sleep apnea syndrome. *Sleep Med Rev* 1998;2:45-60.
- Guilleminault C, Connolly S, Winkle R, Melvin K, Tilkian A. Cyclical variation of the heart rate in sleep apnoea syndrome. Mechanisms, and usefulness of 24 h electrocardiography as a screening technique. *Lancet* 1984;1:126-31.
- Narkiewicz K, Somers VK. Sympathetic nerve activity in obstructive sleep apnoea. *Acta Physiol Scand* 2003;177:385-90.
- Narkiewicz K, Pesek CA, Kato M, Phillips BG, Davison DE, Somers VK, *et al.* Baroreflex control of sympathetic nerve activity and heart rate in obstructive sleep apnea. *Hypertension* 1998;32:1039-43.
- Sforza E, Grandin S, Jouny C, Rochat T, Ibanez V. Is waking electroencephalographic activity a predictor of daytime sleepiness in sleep-related breathing disorders? *Eur Respir J* 2002;19:645-52.
- O'Brien C, Heneghan C. A comparison of algorithms for estimation of a respiratory signal from the surface electrocardiogram. *Comput Biol Med* 2007;37:305-14.
- Mendez MO, Ruini DD, Villantieri OP, Matteucci M, Penzel T, Cerutti S, *et al.* Detection of Sleep Apnea from Surface ECG Based on Features Extracted by an Autoregressive Model. 29<sup>th</sup> Annual International Conference of the IEEE; 2007. p. 6105-8.
- Alvarez D, Hornero R, Marcos JV, del Campo F. Multivariate analysis of blood oxygen saturation recordings in obstructive sleep apnea diagnosis. *IEEE Trans Biomed Eng* 2010;57:2816-24.
- Goldshtein E, Tarasiuk A, Zigel Y. Automatic detection of obstructive sleep apnea using speech signals. *IEEE Trans Biomed Eng* 2011;58:1373-82.
- Dingli K, Assimakopoulos T, Wraith PK, Fietze I, Witt C, Douglas NJ, *et al.* Spectral oscillations of RR intervals in sleep apnoea/hypopnoea syndrome patients. *Eur Respir J* 2003;22:943-50.
- Steven V. Heart Rate Variability: Linear and Non-Linear Analysis with Applications in Humans Physiology. Doctoral Thesis of Faculty of Electrical Engineering Kasteelpark Arenberg. Belgium; 2010.
- Rajendra Acharya U, Paul Joseph K, Kannathal N, Lim CM, Suri JS. Heart rate variability: A review. *Med Biol Eng Comput* 2006;44:1031-51.
- Goldberger AL, West BJ. Applications of nonlinear dynamics to clinical cardiology. *Ann N Y Acad Sci* 1987;504:195-213.
- Voss A, Schulz S, Schroeder R, Baumert M, Caminal P. Methods derived from nonlinear dynamics for analysing heart rate variability. *Philos Trans A Math Phys Eng Sci* 2009;367:277-96.
- Penzel T, Kantelhardt JW, Grote L, Peter JH, Bunde A. Comparison of detrended fluctuation analysis and spectral analysis for heart rate variability in sleep and sleep apnea. *IEEE Trans Biomed Eng* 2003;50:1143-51.
- Al-Angari HM, Sahakian AV. Use of sample entropy approach to study heart rate variability in obstructive sleep apnea syndrome. *IEEE Trans Biomed Eng* 2007;54:1900-4.
- Miyata M, Burioka N, Sako T, Suyama H, Fukuoka Y, Tomita K, *et al.* A short daytime test using correlation dimension for respiratory movement in OSAHS. *Eur Respir J* 2004;23:885-90.
- Miyata M, Burioka N, Suyama H, Sako T, Nomura T, Takeshima T, *et al.* Non-linear behaviour of respiratory movement in obstructive sleep apnoea syndrome. *Clin Physiol Funct Imaging* 2002;22:320-7.
- Moeynoi P, Kitjaidure Y. Canonical Correlation Analysis for Dimensionality Reduction of Sleep Apnea Features Based on ECG Single Lead. 2016 9<sup>th</sup> Biomedical Engineering International Conference (BMEiCON). Laung Prabang. Laos; 2016. p. 1-5.
- Acharya UR, Chua EC, Faust O, Lim TC, Lim LF. Automated detection of sleep apnea from electrocardiogram signals using nonlinear parameters. *Physiol Meas* 2011;32:287-303.
- Jafari A. Sleep apnoea detection from ECG using features extracted from reconstructed phase space and frequency domain. *Biomed Signal Process Control* 2013;8:551-8.
- Zapanta L, Poon CS, White DP, Marcus CL, Katz ES. Heart Rate Chaos in Obstructive Sleep Apnea in Children. In Proceedings of the 26<sup>th</sup> Annual IEEE Engineering in Medicine and Biology Conference. New York, USA. IEEE EMBS; 2006. p. 3565-8.
- Jazi MH, Amra B, Yazdchi MR, Jahangiri M, Tabesh F, Gholamrezaei A, *et al.* P wave duration and dispersion in holter electrocardiography of patients with obstructive sleep apnea. *Sleep Breath* 2014;18:549-54.
- Takens F. Detecting strange attractors in turbulence. *Dynamical Systems and Turbulence*. Warwick, Germany. Springer; 1980. p. 366-81.
- Fraser AM, Swinney HL. Independent coordinates for strange attractors from mutual information. *Phys Rev A Gen Phys* 1986;33:1134-40.
- Ataei M, Sedigh AK, Lohmann B, Lucas C. Determining Embedding Dimension from Output Time Series of Dynamical Systems- Scalar and Multiple Output Cases. Proceedings of the International Conference on System Identification and Control Problems. SICPRO03. Moscow; 2003.
- Grassberger P, Procaccia I. Measuring the strangeness of strange attractors. *Physica D* 1983;9:189-208.
- Franszek M. Optimized algorithm for the calculation of correlation integrals. *Phys Rev A Gen Phys* 1989;39:5440-3.
- Peng CK, Havlin S, Stanley HE, Goldberger AL. Quantification of scaling exponents and crossover phenomena in nonstationary heartbeat time series. *Chaos* 1995;5:82-7.
- Boneau CA. The effects of violations of assumptions underlying the test. *Psychol Bull* 1960;57:49-64.
- Faust O, Acharya UR, Molinari F, Chattopadhyay S, Tamura T. Linear and non-linear analysis of cardiac health in diabetic subjects. *Biomed Signal Process Control* 2012;7:295-302.

---

## BIOGRAPHIES



**Shiva Naghsh** was born in 1989. She received her B.Sc. degree in biomedical engineering in 2011 and the M.Sc. degree in electrical engineering in 2014 from the University of Isfahan, Isfahan, Iran. She is currently working as an R&D engineer. Her research interests include biological signal processing, time series analysis, and data analysis.

**Email:** sh.naghsh@gmail.com



**Mohammad Ataei** received the B.Sc. degree from the Isfahan University of Technology, Iran, in 1994, the M.Sc. degree from the Iran University of Science and Technology, Iran, in 1997, and the Ph.D. degree from the K. N. Toosi University of Technology, Iran, in 2004 (joint project with the University of Bremen, Germany-DAAD scholarship), all in control Engineering. He is currently a full Professor with the Department of Electrical Engineering, University of Isfahan, Iran. His main areas of research interest are control theory and applications, nonlinear control, and chaotic systems' analysis and control.

**Email:** ataei@eng.ui.ac.ir



**Mohammadreza Yazdchi** obtained his B.Sc. in Electrical Engineering from the Isfahan University of Technology in 1997. He received his M.Sc. And Ph.D. in Biomedical Engineering from Amirkabir University of Technology in 2000 and 2006 respectively. He is currently an Associate Professor in the Department of Biomedical

Engineering at the University of Isfahan since 2007. His research interests include Biomedical Signal Processing, Medical Image Processing, Biomedical Instrumentation, Biologically Inspired Computing and Bio-Inspired Engineering.

**Email:** yazdchi@eng.ui.ac.ir



**Mohammad Hashemi** was born in Isfahan in 1965. He studied medicine at Isfahan University of Medical Sciences (IUMS) and graduated in 1990. He continued his studies to become specialized in Cardiology in IUMS until 1994 and began fellowship of Interventional Cardiology in 1997 in Iran University Of Medical Sciences. In 1999,

he joined the department of Internal Medicine of IUMS until now. He is currently a full professor of Cardiology. He is interested on Endothelial Function as academic research.

**Email:** mohammad\_hashemi@med.mui.ac.ir

---