


Article

Genotyping of Multiple Clinical Samples with a Combined Direct PCR and Magnetic Lateral Flow Assay



Chao Zhang,
Xiaonan Liu, Yao
Yao, ..., Zuankai
Wang, Alphonsus
J.M. Vermorken,
Yali Cui

zuanwang@cityu.edu.hk (Z.W.)
yalicui@nwu.edu.cn (Y.C.)

HIGHLIGHTS

Multiple samples (blood,
DBS, saliva and buccal
swab) for PCR template
directly

Combined direct PCR and
magnetic lateral flow
assay for *ALDH2*
genotyping

Genotypes can be
obtained within 90 min

Zhang et al., iScience 7, 170–
179
September 28, 2018 © 2018
The Author(s).
[https://doi.org/10.1016/
j.isci.2018.09.005](https://doi.org/10.1016/j.isci.2018.09.005)

Article

Genotyping of Multiple Clinical Samples with a Combined Direct PCR and Magnetic Lateral Flow Assay

Chao Zhang,^{1,5} Xiaonan Liu,^{1,5} Yao Yao,² Kewu Liu,¹ Wenli Hui,¹ Juanli Zhu,² Yaling Dou,³ Kai Hua,¹ Mingli Peng,² Zuankai Wang,^{4,6,*} Alphonsus J.M. Vermorken,¹ and Yali Cui^{1,2,*}

SUMMARY

Developing a sensitive, low-cost, and easy-to-use point-of-care testing system for genotyping is important for informing treatment decisions and predicting the risk of underlying diseases. Conventional methods normally require complex operational procedures as well as expensive and sophisticated instruments. Here, we report a general approach that enables us to detect the genotype of multiple sample types directly without DNA purification. Moreover, the PCR results can be further quantitatively analyzed based on a magnetic lateral flow assay (MLFA) system, which avoids multiple steps needed for conventional nucleic acid biosensors. As a demonstration, we show that three genotypes of aldehyde dehydrogenase 2 (ALDH2) can be identified using a small volume of sample with an accuracy of 100% and a sensitivity of 1.0×10^2 cells/ μ L, which are better than those of the gold standard methods. We believe that the direct PCR-MLFA system represents a significant advance toward the development of portable, sensitive biomedical platforms.

INTRODUCTION

SNPs represent the most frequent type of variation (approximately 90%) in the human genome (Wang et al., 1998; Ye et al., 2001), the detection of which is of great significance to association studies of complex diseases (Ngo et al., 2016), pharmacogenomics (Li et al., 2015), population genetics (Liu and Fu, 2015; Mccarroll et al., 2008), and physical mapping (Charlier et al., 2008). There are three possible genotypes for each biallelic SNP locus, wild-type and heterozygous and homozygous mutation types, which correspond to different phenotypes and lead to different clinical manifestations. For genotyping, three distinct operational steps are typically required: sample preparation, target amplification, and signal readout. Although each step can be considered individually, it is important to emphasize that a key challenge for the development of such nucleic acid detection methods is streamlining the steps as much as possible. Existing methods usually require sample purification via a labor-intensive, time-consuming, and expensive sample preparation step due to the stringent requirements for DNA quality and quantity (Aljanabi and Martinez, 1997; Golenberg et al., 1996), which also enhances the risk of cross-contamination between samples (Brouwershagen et al., 2010; Kelly and Woolley, 2005). Nevertheless, the amplification products are usually analyzed by agarose gel electrophoresis, a further step such as hybridization, or professional software in a signal readout step. Therefore, a universal platform that can visually or automatically read out the PCR results is lacking. Moreover, existing analytical methods are usually too slow to support immediate treatment decisions or to be implemented as point-of-care diagnostics (Schoepp et al., 2017). Scientists have developed various methods to increase the timeliness, sensitivity, and specificity of genotyping using more sophisticated techniques and instruments. Currently, DNA sequencing is considered to be the gold standard for laboratory-based SNP detection (Hawkins, 2017); however, high operating expenses and high-precision equipment requirements have impeded its wide application, particularly in resource-poor settings.

Establishing a genotyping system that is adapted to a wide variety of sample types and avoids DNA isolation is challenging because current strategies for genotyping based on PCR require relatively high purity of DNA. However, samples commonly used in clinical settings, such as blood, saliva, dried blood spot (DBS), and buccal swab, are not easy to genotype without DNA purification because cell membranes prevent cell damage and DNA release. Thus, developing a general strategy for a genotyping system that analyzes multiple sample types without DNA isolation has remained elusive. Although several protocols and

¹College of Life Sciences, Northwest University, Xi'an, China

²Shaanxi Provincial Engineering Research Center of Nano-Biomedical Detection, Xi'an, China

³Department of Clinical Laboratory, Peking Union Medical College Hospital, Chinese Academy of Medical Sciences, Beijing, China

⁴Department of Mechanical Engineering, City University of Hong Kong, Hong Kong, China

⁵These authors contributed equally

⁶Lead Contact

*Correspondence: zuanwang@cityu.edu.hk (Z.W.), yalicui@nwu.edu.cn (Y.C.)

<https://doi.org/10.1016/j.isci.2018.09.005>






Figure 1. Schematic Diagram of the Direct PCR-MLFA System

(A) Sample collection and treatment with NaOH.

(B) Target amplification using specifically designed primers (corresponding to M and WT).

(C) Signal readout: the composition and working principle of the magnetic lateral flow assay. When the amplification products are added to the MLFA strip, they are driven forward by capillary action, and magnetic signals and color appear at the control line (C line) or both the control and test lines (T line). The results can be read visually, or they can be detected automatically with a magnetic signal reader.

commercial kits that perform PCR amplification without DNA extraction from whole blood (Nanayakkara et al., 2017) and saliva (Ambers et al., 2018) have been developed, they still require complex reagents, large volume of samples, and substantial power consumption.

To overcome this challenge, we have established a new approach that makes direct PCR feasible by leveraging a simple process that enables the efficient release of nucleic acids and inactivates the natural PCR inhibitors in a sample. Many methods have offered sensitive and specific approaches for genotyping. However, the tremendous potential of gene detection in clinical practice has been limited thus far by multiple time-consuming steps and the need for professionally trained staff and costly equipment to obtain genotyping results. Furthermore, the results are usually analyzed by agarose gel electrophoresis (most PCR-based methods) (Ramire-Zexpósito et al., 2016; Wu et al., 2015), a further reaction step, and professional software (e.g., biochip, sequencing, and mass array) (Nijveen et al., 2013; Xu et al., 2012; Yi et al., 2014). To address these issues, we used a magnetic lateral flow assay (MLFA) system to interpret the results through visualization or a magnetic signal reader. The MLFA is derived from traditional lateral flow strips using our own synthetic gold magnetic nanoparticles (GMNPs) as labels, which can be harnessed to develop simple and portable devices that enable both the visual and quantitative interpretation of data for point-of-care SNP detection.

In this study, we developed an integrated strategy for miniaturizing simple process laboratory assays to shorten their complex steps and demonstrated that the entire contiguous sample-to-answer workflow could enable the genotyping of a clinical sample in less than 90 min. We combined two improved technologies into one system: a direct PCR assay for DNA amplification without purification and an MLFA system to read the results rapidly and automatically. These contributions address several bottlenecks of current methods while providing the advantages of simplicity, cost, portability, and quantitative genotyping. This technology is one step closer to realizing the ubiquitous availability of gene tests, which can ultimately aid rapid medical decisions.

RESULTS

Direct PCR-MLFA System Design

The proof-of-concept scheme described in this article uses direct PCR followed by an MLFA (direct PCR-MLFA) for SNP detection in two steps: target amplification and signal readout. Figure 1A shows the

collection and treatment of clinical samples. Designed as a universal genotyping platform, the direct PCR-MLFA system could be used for a variety of sample types. Four frequently used clinical samples, including whole blood, DBS, buccal swab, and saliva, were selected and successfully applied to the established platform, which indicates that the direct PCR-MLFA system is a platform with general applications that could be applied to type other genes and sample types. The collected sample is treated with NaOH solution followed by direct PCR amplification with specifically designed primers as shown in [Figures 1B](#) and [S1](#) and [Table S1](#). For convenience, we refer to the mutation type as “M” and the wild-type as “WT.” For each sample to be analyzed, two separate reactions (M tube and WT tube) are run simultaneously using the same treated sample. Two sets of allele-specific primers (for the WT and M sequences) are added to two different tubes for amplification, and target PCR products are acquired only when the 3' end of the specific primer is complementary with the template. The direct PCR procedure takes only about 5 min for sample collection and treatment and about 75 min for amplification. The three possible outcomes are presented in [Figure 1B](#): (a) homozygous mutation type, (b) heterozygous mutation type, and (c) wild-type.

As shown in [Figure 1C](#) (signal readout step), the MLFA is performed using a strip that is composed of five components, a sample pad, a conjugate pad, a nitrocellulose membrane, an absorbent pad, and a plastic cushion. Amplicons are added to the sample pad and then read automatically with a magnetic reader or visualized with the unaided eye. The PCR products added to the sample pads of the MLFA strips can be combined with anti-digoxin (Dig) antibody-conjugated GMNPs (GMNPs-anti-Dig) on the conjugate pad due to the specific combination of a Dig label at the 5'-end of the specific primers and an anti-Dig antibody. As the products migrate along the strip, the PCR product-GMNP-anti-Dig conjugates can be captured by streptavidin immobilized at the test line (T line) due to the biotin label at the 5' end of the common primers, which produces a red band at the T line. The rest of the GMNPs-anti-Dig keep moving and are captured by goat anti-mouse IgG at the control line (C line), which forms another red band that confirms the efficacy of the MLFA system. In the absence of target PCR products, no red band is observed at the T line. The results can be obtained using a magnetic reader to detect the magnetic signals at the T and C lines, or they can be read visually according to the presence of a red band at the T line. The MLFA system operating procedure can be completed within 5 min from amplicon loading to result readout.

The assay control system must be capable of verifying the amplification step and testing the lateral flow step. For the amplification step, only when at least one of the two test strips has a red band on the test line after lateral flow detection the amplification test is treated as valid. This is because one SNP can only have one of the three genotypes: wild-type, heterozygous mutation type, and homozygous mutation type. Therefore, the dual strips themselves can be applied as a control to ensure that the amplification process is effective. For the signal readout step, the control line is included for testing the GMNPs and the proper reagent flow in each strip. Meanwhile, a positive control using sequenced genomic DNA (wild-type and homozygous type) as templates and a negative control using NaOH-treated double-distilled water (ddH₂O) as template are performed simultaneously. To sum up, both amplification and lateral flow procedure are under appropriate control.

Biological Sample Treatment

Technically, the factors that cause challenges for genotyping clinical samples directly include the enclosure of nucleic acids by a membrane and the presence of commensal or contaminating organisms in the clinical samples, which can inhibit PCR amplification. If the sample treatment method does not lyse the membrane or inactivate the inhibitors, then the genotype cannot be determined accurately. To facilitate the release of DNA from cells, various methods have been applied, such as using expensive special materials or complex solvents ([Gong and Li, 2014](#); [Saavedra-Matiz et al., 2013](#)). Despite their efficacy, the utility of these procedures is questionable. Therefore, the search for an inexpensive and nontoxic procedure is ongoing. Herein, we present a simple, inexpensive yet powerful chemical procedure devoid of costly materials and complex solutions for DNA lysis. Briefly, NaOH solution was used for cell lysis, which allowed rapid DNA release from the collected samples.

First, we tested the effects of the NaOH solution on cells in the samples. As shown in [Figure 2A](#), a large amount of cellular debris was found in the whole blood samples treated with NaOH, whereas the cellular integrity of the blood cells was preserved in physiological saline. Similar results were found with saliva and buccal swab solutions treated with NaOH and physiological saline. The results indicated that cell lysis occurs during alkaline treatment, releasing DNA from the cells into the sample solution. The success of direct




Figure 2. Effect of NaOH in the Direct PCR-MLFA System

(A) Image of blood cells treated with physiological saline or NaOH.

(B) Transmission electron microscopy images of GMNPs-anti-digoxin antibodies treated with ddH₂O or direct PCR product.

(C) Images and average size (diameter) of GMNPs-anti-digoxin antibodies under different treatment conditions.

PCR amplification may be affected by the sample dilution factor, which tends to significantly reduce the concentration of PCR inhibitors present in the sample. We have failed to obtain any amplification products in NaOH-based direct PCR at a 1:1 dilution, suggesting that dilution may play a critical role in the success of direct PCR.

As the pH of the NaOH solution can be as high as 13.0, we next evaluated the pH value of the reaction system, which affects the PCR efficiency of the target amplification step as the optimal working pH for Taq DNA polymerase is 7.4–8.3, according to Min et al. (Min et al., 2005). We tested the pH value at each stage of the direct PCR process after NaOH addition. As shown in Figure S2, the pH value of the NaOH solution was 13.0; when the NaOH solution was added to the sample, the pH value of the mixture was 11.8–12.0. However, the pH value drops dramatically to 7.6–8.1 after mixing the NaOH-treated sample with the PCR reaction buffer, and no significant difference in pH value was observed between the PCR reaction buffer with and without the NaOH-treated sample. Thus, the Tris-HCl buffer in the PCR reaction played a determinant role in pH control and provided an optimal working environment for DNA polymerase.

Direct PCR Amplification

Our next goal was to optimize the direct PCR procedure to improve the specificity while eliminating background noise (non-specific amplification or primer dimerization). We designed allele-specific primer pairs and optimized their concentrations in the direct PCR system. By testing various conditions (100, 75, 50, 25,

and 10 nM) simultaneously, the optimized primer concentrations were determined to be 50 nM for each primer (Figure S3A). In the current study, to assure specificity, the allele-specific primers were designed to contain an additional mismatch at the penultimate 3' nucleotide to increase the strength of allelic discrimination according to the principle of amplification-refractory mutation system (Little, 1995; Newton et al., 1989). Moreover, we also studied the effect of sample concentration on the direct PCR-MLFA system signal. The response signal (T line) increased when the sample solution in the PCR mix was raised from 3 to 5 μ L but then leveled off, indicating that the optimal volume was attained (Figure S3B).

To assess the ability of the established direct PCR to eliminate false-positive results due to carryover contaminants during SNP detection, we carried out contamination prevention reactions using dUTP and uracil-DNA glycosylase (UDG). Using dUTPs during amplification and pretreating subsequent samples with UDG to remove carryover PCR products from prior reactions are common practices (Laird, 2010). To verify the effectiveness of the UDG system in the direct PCR-MLFA system, the sample solution and/or a few previous PCR products were combined into tubes as template before direct PCR amplification, and a digestion step (2 min at 50°C) was executed. The UDG enzyme specifically cleaves uracil bases from any uracil-containing PCR products. As a result, the carryover contaminants generated from previous PCR reactions are effectively eliminated (as shown in Figure S3C), and they cannot be used as templates for re-amplification. Hence, UDG is able to prevent the amplification of carryover contaminant templates, which significantly decreases the likelihood of false-positive results during direct PCR-MLFA analysis.

Once the direct PCR products are loaded onto the sample pad of the MLFA strip, the stability of the GMNPs-anti-Dig is invariably challenging for lateral flow. To verify the morphological stability of the GMNPs-anti-Dig, transmission electron microscopic (TEM, Figure 2B) images were observed. As shown in the TEM image, the GMNPs-anti-Dig were well dispersed in ddH₂O and the direct PCR product solution had no effect on the GMNPs-anti-Dig. The good dispersion of the GMNPs-anti-Dig was further confirmed by measuring their average size (diameter) under different conditions (Figure 2C). A single, sharp peak at 100–150 nm was observed with GMNPs-anti-Dig in ddH₂O (blue), NaOH solution (yellow), and PCR product solution (with purified DNA as template; black). However, in the direct PCR product solution with NaOH-treated sample, there are two peaks at 100 and 3,100+ nm (green), likely due to the cellular debris in the treated sample.

When direct PCR-MLFA was performed with collected samples, the best results were obtained at a final reaction volume of 50 μ L, which contained 10 \times PCR buffer, 0.2 mM of each dNTP (dATP, dUTP, dCTP, and dGTP), 3 mM of MgCl₂, 0.5 U of HotMaster Taq DNA polymerase (TIANGEN Biotech Co., Ltd., Beijing, China), 0.5 U of UDG (Shanghai ShineGene Molecular Biotechnology Co., Ltd., Shanghai, China), 50 nM of forward and reverse primers (forward M primer in M tube and forward WT primer in WT tube, reverse primer added in both tubes), and 5 μ L of sample-treated solution as template. All the amplifications were performed according to the following parameters: two initial denaturation steps for 2 min at 50°C and 3 min at 94°C; 32 cycles of 5 s at 94°C, 10 s at 60°C, and 30 s at 65°C; and one step of 10 min at 65°C.

Performance of the Direct PCR-MLFA System

To test the detection limit of the direct PCR-MLFA system, we assayed serial dilutions of whole blood, saliva, and buccal swab samples using a magnetic signal reader to obtain accurate parameters. For the direct, quantitative measure of the limit of detection (LOD), serial dilutions of blood, saliva, and buccal swab samples were prepared in physiological saline and analyzed with the direct PCR-MLFA system. As shown in Figure S4, whole-blood samples containing different densities of white blood cells (WBCs), ranging from 0.06×10^3 – 5.96×10^3 cells/ μ L (Figure S4A); saliva samples containing different densities of oral epithelial cells, ranging from 0.03×10^3 – 0.46×10^3 cells/ μ L (Figure S4C); and buccal swab samples containing different densities of oral epithelial cells, ranging from 0.04×10^3 – 1.35×10^3 cells/ μ L (Figure S4E), were examined with our system. The PCR amplification efficiency (relative magnetic units [RMUs]) significantly improved as the concentration of WBCs and oral epithelial cells increased. Figures S4B, S4D, and S4F show the magnetic signal peak value of the T line and C line at different concentrations of WBCs and oral epithelial cells, which simulated the output of the magnetic reader. Even when the dose was as low as 1.0×10^2 cells/ μ L, the specificity of the test remained high, with no false-negative results. Normally, the number of WBCs in whole blood is in the range of 4.0×10^3 – 10.0×10^3 cells/ μ L. Therefore, the LOD of the direct PCR-MLFA system is fully applicable for clinical testing. Moreover, as shown in Figure 3, a standard curve for detection was plotted as the average magnetic signal at the T line relative to the




Figure 3. Standard Curve for the Direct PCR-MLFA Detection

Standard curve for the detection of the average magnetic signal at the T line relative to the respective WBC count in whole blood (A) or the oral epithelial cell density in (B) saliva, and (C) buccal swab samples.

respective concentrations of WBCs or oral epithelial cells. The high correlation coefficient ($R^2 > 0.997$) indicated that the direct PCR-MLFA system could be applied for DNA quantification.

We further assessed the reproducibility of the direct PCR-MLFA system by analyzing amplicons with wild type, heterozygous mutation, and homozygous mutation genotypes (verified by DNA sequencing) of aldehyde dehydrogenase 2 (*ALDH2*2*). Three amplifications per sample were carried out, followed by visual detection using lateral flow strips. The three pairs of strips were prepared using three different batches of reagents under the same protocol. The results presented in Figure S5 indicated that no differences were detected among the three pairs of strips. The present study therefore illustrates the high reproducibility, sensitivity, and reliability of our direct PCR-MLFA system, providing a useful method for direct SNP detection in clinical samples.

Evaluation of the Direct PCR-MLFA System Using Clinical Samples

Having established the direct PCR-MLFA system, we next tested its accuracy with clinical samples (matched 200 whole blood and 200 DBS samples; matched 50 buccal swab and 50 saliva samples). For each sample, the magnetic signal value (RMUs) of the T line was obtained. All blood samples were also sequenced by BGI (Beijing Genomic Institute, Beijing, China). As shown in Figure 4, the RMU of WT strip was defined as a negative value, and the M strip was defined as a positive value. Of the 200 whole blood samples and DBS samples, 2 samples had only an M RMU, 134 samples had only a WT RMU, and the remaining 64 samples had both M and WT RMUs (Figures 4A and 4B), indicating 2 homozygous mutation type samples, 134 wild-type samples, and 64 heterozygous mutation type samples among the 200 whole blood samples and DBS samples. Of the 50 saliva samples and buccal swab samples, 1 sample had only an M RMU, 16 samples had only a WT RMU, and the remaining 33 samples had both M and WT RMUs (Figures 4C and 4D), indicating 1 homozygous mutation type sample, 16 wild-type samples, and 33 heterozygous mutation type samples among the 50 saliva samples and buccal swab samples. No discrepancies were observed when the results of the direct PCR-MLFA system and DNA sequencing were compared (Table S2). The observed allele frequencies were 83% and 17% for *2G and *2A, respectively (calculated from *2G: $F1 + F2/2$ and *2A: $F2/2 + F3$). The frequencies of the *ALDH2*2* allele measured with the direct PCR-MLFA system were not significantly different from those reported by Eng. et al. in the Chinese population (Eng et al., 2007). As shown in Table S3, the one-to-one correspondence between the genotyping results of the buccal swab and saliva samples using the direct PCR-MLFA system also indicated successful SNP detection using buccal swab and saliva samples.

Limitation of Study

Here, we have established a sensitive, low cost, and easy-to-use SNP detection platform using the combination of the direct PCR and lateral flow assay, which enables us to obtain genotyping results within 90 min by using the clinically relevant samples (whole blood, DBS, buccal swab, and saliva) directly without DNA purification. It is envisaged that the direct PCR-MLFA system could be quickly adapted for the detection of other DNA mutation types, such as recombination, deletion, and insertion. Despite many advantages, it still remains elusive for us to detect the variation in the DNA copy number. Further efforts on the design of the amplification principle and primer sets will be explored to make the MLFA system applicable to copy number variation detection.




Figure 4. Genotyping Results for Multiple Sample Types (the Magnetic Signal Value of the Wild-Type Was Defined as a Negative Value)

- (A) Genotyping results of 200 whole blood samples.
 (B) Genotyping results of 200 dried blood spot samples.
 (C) Genotyping results of 50 saliva samples.
 (D) Genotyping results of 50 buccal swab samples.

DISCUSSION

To date, most of the PCR-based genotyping methods rely on the use of the purified DNA, rather than the use of the clinical samples (Cavanaugh and Bathrick, 2018). Moreover, only a few conventional PCR-based genotyping methods are capable of achieving a sample-to-answer result directly from a clinical sample in less than 90 min (Broccanello et al., 2018; Di et al., 2017; White and Cantsilieris, 2017). Remarkably, Gomez-Martinez et al. have developed a multiplex linear-after-the-exponential (LATE)-PCR for visual detection of blood group genotype with turnaround time of approximately 1 hr by using a KAPA2G Fast HotStart DNA Polymerase (Gomez-Martinez et al., 2018). However, it is not clear whether this LATE-PCR method can be applicable to different sample types. By leveraging NaOH to release nucleic acids and inactivate the natural PCR inhibitors in a sample, our PCR-MLFA system is capable of genotyping target DNA with different clinical sample types directly by using a common Taq DNA polymerase without undergoing the conventional DNA isolation step. Thus, compared with LATE-PCR, which involves the use of genetically engineered HotStart DNA polymerase, our method is more affordable and versatile.

The treatment of samples with NaOH plays an important role in the direct PCR-MLFA system, making the whole testing procedure convenient and usable. First, compared with other SNP detection methods, the DNA purification step was eliminated in our assay with the help of NaOH treatment, which shortened the processing time from 1 or 2 hr to a mere 5 min. Second, by using NaOH-treated blood samples for PCR, the problem of cross-contamination, which exists in traditional blood DNA purification processes, is eliminated. Furthermore, various natural PCR inhibitors in the samples such as hemoglobin, IgG, lactoferrin, and proteases (Adams, 2005) are inactivated by NaOH treatment.

Lateral flow assay was chosen for signal readout because it is a well-established technique with several readout methods (Nayak et al., 2016; Sajid et al., 2015). By further leveraging on the GMNPs, the MLFA system was established, allowing for the qualitative analysis of genotype using an automatic magnetic reader. The major advantage of this assay is its rapid qualitative output of "Yes" or "No." Compared with a previous study on SNP detection using PCR-based methods or DNA sequencing, our assay has significant advantages (Hui et al., 2016; Ngo et al., 2016; Yun et al., 2015; Zhang et al., 2016). Specifically, (1) existing genotyping methods based on PCR, which require purified DNA as template, such as PCR microarray, qPCR, and DNA sequencing, are time consuming because conventional DNA purification usually takes more than 1 hr, whereas the present assay system requires less than 5 min for sample preparation. Moreover, (2) the previously developed techniques usually require a complex operational procedure as well as expensive and sophisticated instruments that may not be available in many laboratories, whereas the present method provides an easy-to-operate and affordable on-site technique for genotyping with high efficiency. (3) The MLFA is ideal for quantification, because it is made with magnetic nanoparticles and there is no optical interfering with magnetic signal readout, which is distinct from traditional lateral flow assay.

Our MLFA chip provides an alternative to sequencing or real-time PCR for rapid genotyping while maintaining the advantages of lateral flow assay in terms of simplicity, affordability, and rapid qualitative/quantitative nucleic acid readout (Chen et al., 2016). For simplicity and affordability, the MLFA chip is designed with simple construction, materials, and reagents and the cost is quite low and can be further reduced when mass-produced. Combined with direct PCR, we have shown that it is possible to perform PCR amplification using multiple sample types directly without DNA purification. This is significantly lower cost and faster readout than conventional genotyping methods.

Human *ALDH2*, which is responsible for the oxidation of aldehydes in the liver, was utilized as a model for SNP detection. Differences in *ALDH2* expression may contribute to a wide variety of human diseases, including cardiovascular disease, diabetes, and cancer (Chen et al., 2014). In addition, genetic polymorphisms of *ALDH2* alter susceptibility to ethanol intake as well as the risk of alcoholism and alcoholic complications, and *ALDH2* may possess important therapeutic potential against alcoholism and other forms of myocardial damage (Hou et al., 2017). The application of this rapid and direct PCR-MLFA system in additional SNP studies is an important next step that, if successful, would further validate and demonstrate its clinical utility. In addition, the standard curve obtained in this study had a high correlation coefficient, indicating that the direct PCR-MLFA assay could also be applied for human DNA quantification. Therefore, the assay could be used to determine trace amounts of DNA of interest among abundant background DNA, such as specific mutation detection in circulating tumor DNA (Thierry et al., 2014), and for complex gene quantification, both of which are clinically valuable (Gonçalves et al., 2016; Sidstedt et al., 2017).

METHODS

All methods can be found in the accompanying [Transparent Methods supplemental file](#).

SUPPLEMENTAL INFORMATION

Supplemental Information includes Transparent Methods, five figures, and three tables and can be found with this article online at <https://doi.org/10.1016/j.isci.2018.09.005>.

ACKNOWLEDGMENTS

This research was supported by the National Natural Science Foundation of China (NO. 81772289) and Shaanxi Provincial Nano-Biomedical Detection Innovation Team Found (No. 2017KCT-25).

AUTHOR CONTRIBUTIONS

C.Z. and X.L. carried out assays and analyzed the results; C.Z., X.L., and K.L. supported the optimization of assays and prepared data; W.H. developed the strip; X.L. and K.L. collected specimens, C.Z., Y.D., M.P., Z.W., and A.J.M.V. performed analyses of assay results; Y.C., Z.W., J.Z., and C.Z. conceived and designed the experiments; C.Z., Y.Y., K.H., Y.C., and Z.W. wrote the manuscript.

DECLARATION OF INTERESTS

The authors declare no competing interests.

Received: July 17, 2018

Revised: August 19, 2018

Accepted: September 5, 2018

Published: September 28, 2018

REFERENCES

- Adams, D.N. (2005). Shortcut method for extraction of staphylococcus aureus DNA from blood cultures and conventional cultures for use in real-time PCR assays. *J. Clin. Microbiol.* **43**, 2932–2933.
- Aljanabi, S.M., and Martinez, I. (1997). Universal and rapid salt-extraction of high quality genomic DNA for PCR-based techniques. *Nucleic Acids Res.* **25**, 4692–4693.
- Ambers, A., Wiley, R., Novroski, N., and Budowle, B. (2018). Direct PCR amplification of DNA from human bloodstains, saliva, and touch samples collected with microFLOQ® swabs. *Forensic Sci. Int. Genet.* **32**, 80–87.
- Broccanello, C., Chiodi, C., Funk, A., Mcgrath, J.M., Panella, L., and Stevanato, P. (2018). Comparison of three PCR-based assays for SNP genotyping in plants. *Plant Methods* **14**, 28.
- Brouwershaven, I.R.V., Bruil, M.L., Leeuwen, G.C.M.V., and Kox, L.F.F. (2010). A real-time (TaqMan) PCR assay to differentiate *Monilinia fructicola* from other brown rot fungi of fruit crops. *Plant Pathol.* **59**, 548–555.
- Cavanaugh, S.E., and Bathrick, A.S. (2018). Direct PCR amplification of forensic touch and other challenging DNA samples: a review. *Forensic Sci. Int. Genet.* **32**, 40–49.
- Charlier, C., Coppieters, W., Rollin, F., Desmecht, D., Agerholm, J.S., Cambisano, N., Carta, E., Dardano, S., Dive, M., and Fasquelle, C. (2008). Highly effective SNP-based association mapping and management of recessive defects in livestock. *Nat. Genet.* **40**, 449.
- Chen, C.H., Ferreira, J.C.B., Gross, E.R., and Mochlyrosen, D. (2014). Targeting aldehyde dehydrogenase 2: new therapeutic opportunities. *Physiol. Rev.* **94**, 1–34.
- Chen, Y., Sun, J., Xianyu, Y., Yin, B., Niu, Y., Wang, S., Cao, F., Zhang, X., Wang, Y., and Jiang, X. (2016). A dual-readout chemiluminescent-gold lateral flow test for multiplex and ultrasensitive detection of disease biomarkers in real samples. *Nanoscale* **8**, 15205–15212.
- Di, F.R., Amtrano, F., and De, L.D. (2017). Evaluation of genotyping methods and costs for VDR polymorphisms. *Eur. Rev. Med. Pharmacol. Sci.* **21**, 1–3.
- Eng, M.Y., Luczak, S.E., and Wall, T.L. (2007). ALDH2, ADH1B, and ADH1C genotypes in Asians: a literature review. *Alcohol Res. Health* **30**, 22–27.
- Golenberg, E.M., Bickel, A., and Weihs, P. (1996). Effect of highly fragmented DNA on PCR. *Nucleic Acids Res.* **24**, 5026–5033.
- Gomez-Martinez, J., Silvy, M., Chiaroni, J., Fournier-With, C., Roubinet, F., Bailly, P., and Brès, J. (2018). Multiplex lateral flow assay for rapid visual blood group genotyping. *Anal. Chem.* **90**, 7502–7509.
- Gonçalves, H.M., Moreira, L., Pereira, L., Jorge, P., Gouveia, C., Martins-Lopes, P., and Fernandes, J.R. (2016). Biosensor for label-free DNA quantification based on functionalized LPGs. *Biosens. Bioelectron.* **84**, 30–36.
- Gong, R., and Li, S. (2014). Extraction of human genomic DNA from whole blood using a magnetic microsphere method. *Int. J. Nanomed.* **9**, 3781–3789.
- Hawkins, G.A. (2017). Chapter 5-Analysis of human genetic variations using DNA sequencing. In *Basic Science Methods for Clinical Researchers*, M. Jalali, F.Y.L. Saldanha, and M. Jalali, eds. (Academic Press), pp. 77–98.
- Hou, G., Chen, L., Liu, G., Li, L., Yang, Y., Yan, H.X., Zhang, H.L., Tang, J., Yang, Y.C., and Lin, X. (2017). Aldehyde dehydrogenase-2 (ALDH2) opposes hepatocellular carcinoma progression by regulating AMP-activated protein kinase signaling in mice. *Hepatology* **65**, 1628–1644.
- Hui, W., Zhang, S., Zhang, C., Wan, Y., Zhu, J., Zhao, G., Wu, S., Xi, D., Zhang, Q., and Li, N. (2016). A novel lateral flow assay based on GoldMag nanoparticles and its clinical applications for genotyping of MTHFR C677T polymorphisms. *Nanoscale* **8**, 3579–3587.
- Kelly, R.T., and Woolley, A.T. (2005). Microfluidic systems for integrated, high-throughput DNA analysis. *Anal. Chem.* **77**, 252–257.
- Laird, P.W. (2010). Principles and challenges of genome-wide DNA methylation analysis. *Nat. Rev. Genet.* **11**, 191–203.
- Li, P., He, H., Wang, Z., Feng, M., Jin, H., Yun, W., Zhang, L., Zhang, L., and Tang, X. (2015). Sensitive detection of single-nucleotide mutation in the BRAF mutation site (V600E) of human melanoma using phosphate-pyrene-labeled DNA probes. *Anal. Chem.* **88**, 883–889.
- Little, S. (1995). Amplification-refractory mutation system (ARMS) analysis of point mutations. In *Curr. Protoc. Hum. Genet.*, **7** (John Wiley & Sons), pp. 9.8.1–9.8.12.
- Liu, X., and Fu, Y.X. (2015). Exploring population size changes using SNP frequency spectra. *Nat. Genet.* **47**, 555–559.
- Mccarroll, S.A., Kuruvilla, F.G., Korn, J.M., Cawley, S., Nemesi, J., Wysoker, A., Shaper, M.H., de Bakker, P.I., Maller, J.B., and Kirby, A. (2008). Integrated detection and population-genetic analysis of SNPs and copy number variation. *Nat. Genet.* **40**, 1166–1174.
- Min, L., Lin, Y.C., Wu, C.C., and Liu, H.S. (2005). Enhancing the efficiency of a PCR using gold nanoparticles. *Nucleic Acids Res.* **33**, e184.
- Nanayakkara, I.A., Cao, W., and White, I.M. (2017). Simplifying nucleic acid amplification from whole blood with direct PCR on chitosan microparticles. *Anal. Chem.* **89**, 3773–3779.
- Nayak, S., Blumenfeld, N.R., Laksanasopin, T., and Sia, S.K. (2016). Point-of-care diagnostics: recent developments in a connected age. *Anal. Chem.* **89**, 102–123.
- Newton, C.R., Graham, A., Heptinstall, L.E., Powell, S.J., Summers, C., Kalsheker, N., Smith, J.C., and Markham, A.F. (1989). Analysis of any point mutation in DNA. The amplification refractory mutation system (ARMS). *Nucleic Acids Res.* **17**, 2503.
- Ngo, H.T., Gandra, N., Fales, A.M., Taylor, S.M., and Vo-Dinh, T. (2016). Sensitive DNA detection and SNP discrimination using ultrabright SERS nanorattles and magnetic beads for malaria diagnostics. *Biosens. Bioelectron.* **81**, 8–14.
- Nijveen, H., Kaauwen, M.V., Esselink, D.G., Hoegen, B., and Vosman, B. (2013). QualitySNPng: a user-friendly SNP detection and visualization tool. *Nucleic Acids Res.* **41**, W587–W590.
- Ramirez-Zepósito, M.J., Martinez-Martos, J.M., Palomeque, T., and Lorite, P. (2016). A PCR-RFLP method for detection of the hNPEP encoding human insulin-regulated aminopeptidase (IRAP) rs4869317 polymorphism. *Indian J. Med. Res.* **144**, 120–123.
- Saavedra-Matiz, C.A., Isabelle, J.T., Biski, C.K., Duva, S.J., Sweeney, M.L., Parker, A.L., Young, A.J., Diantonio, L.L., Krein, L.M., and Nichols, M.J. (2013). Cost-effective and scalable DNA extraction method from dried blood spots. *Clin. Chem.* **59**, 1045–1051.
- Sajid, M., Kawde, A.N., and Daud, M. (2015). Designs, formats and applications of lateral flow assay: a literature review. *J. Saudi Chem. Soc.* **19**, 689–705.
- Schoepp, N.G., Schlappi, T.S., Curtis, M.S., Butkovich, S.S., Miller, S., Humphries, R.M., and Ismagilov, R.F. (2017). Rapid pathogen-specific phenotypic antibiotic susceptibility testing using digital LAMP quantification in clinical samples. *Sci. Transl. Med.* **9**, eal3693.
- Sidstedt, M., Romsos, E.L., Hedell, R., Ansell, R., Steffen, C.R., Vallone, P.M., Rådström, P., and Hedman, J. (2017). Accurate digital polymerase chain reaction quantification of challenging samples applying inhibitor-tolerant DNA polymerases. *Anal. Chem.* **89**, 1642–1649.
- Thierry, A.R., Mouliere, F., El, M.S., Mollevi, C., Lopezcrapez, E., Rolet, F., Gillet, B., Gongora, C., Dechelotte, P., and Robert, B. (2014). Clinical validation of the detection of KRAS and BRAF mutations from circulating tumor DNA. *Nat. Med.* **20**, 430–435.

Wang, D.G., Fan, J.B., Siao, C.J., Berno, A., Young, P., Sapolsky, R., Ghandour, G., Perkins, N., Winchester, E., and Spencer, J. (1998). Large-scale identification, mapping, and genotyping of single-nucleotide polymorphisms in the human genome. *Science* 280, 1077–1082.

White, S.J., and Cantsilieris, S. (2017). Genotyping: methods and protocols. In *Methods in Molecular Biology*, 1492, S.J. White and S. Cantsilieris, eds. (Humana Press).

Wu, Z.K., Zhou, D.M., Wu, Z., Chu, X., Yu, R.Q., and Jiang, J.H. (2015). Single-base mismatch discrimination by T7 exonuclease with target

cyclic amplification detection. *Chem. Commun.* 51, 2954–2956.

Xu, F., Wang, W., Wang, P., Li, M.J., Sham, P.C., and Wang, J. (2012). A fast and accurate SNP detection algorithm for next-generation sequencing data. *Nat. Commun.* 3, 1258.

Ye, S., Dhillon, S., Ke, X., Collins, A.R., and Day, I.N.M. (2001). An efficient procedure for genotyping single nucleotide polymorphisms. *Nucleic Acids Res.* 29, 88.

Yi, M., Zhao, Y., Jia, L., He, M., Kebebew, E., and Stephens, R.M. (2014). Performance comparison of SNP detection tools with illumina exome

sequencing data—an assessment using both family pedigree information and sample-matched SNP array data. *Nucleic Acids Res.* 42, e101.

Yun, K.J., Kim, J., Mathies, R.A., and Chem, A. (2015). Microfluidic linear hydrogel array for multiplexed single nucleotide polymorphism (SNP) detection. *Anal. Chem.* 87, 3165–3170.

Zhang, C., Yao, Y., Zhu, J.L., Zhang, S.N., Zhang, S.S., Wei, H., Hui, W.L., and Cui, Y.L. (2016). Establishment and application of a real-time loop-mediated isothermal amplification system for the detection of CYP2C19 polymorphisms. *Sci. Rep.* 6, 26533.

ISCI, Volume 7

Supplemental Information

Genotyping of Multiple Clinical Samples

with a Combined Direct PCR

and Magnetic Lateral Flow Assay

Chao Zhang, Xiaonan Liu, Yao Yao, Kewu Liu, Wenli Hui, Juanli Zhu, Yaling Dou, Kai Hua, Mingli Peng, Zuankai Wang, Alphonsus J.M. Vermorken, and Yali Cui




Figure S1, Related to Figure 1. Optimal primer set selection. (M=Mutant, WT=Wild Type)

(A) The primer sets were designed according to the principle of Amplification Refractory Mutation System (ARMS). We tested six sets of mismatched primers (as showed in Table S1) by agarose gel to select for the most specific primer sets. As showed in the figure, the results indicated that, except for primer set 1 and set 3, the other groups result in non-specific bands on the agarose gel.

(B) When the priemr set 1 and primer set 3 were applied to the lateral flow strip, the mutant primer of the primer set 3 showed a non-specific band.




Figure S2, Related to Figure 2. The pH value at each stage of direct PCR assay.




Figure S3, Related to Figure 2. Optimization of the direct PCR-MLFA system (M=M tube, WT=WT tube, NC=negative control, using NaOH treated ddH₂O as template instead of NaOH treated sample).

(A) Optimized primer concentrations were determined to be 50 nM.

(B) Optimized sample solution amounts were determined to be 5 μ L.

(C) The effect of UDG used to prevent contamination. Line 1 and line 4: homozygous mutation sample; line 2 and line 5: homozygous wild type sample; line 3 and line 6: heterozygous mutation sample.




Figure S4, Related to Figure 3. Sensitivity and specificity of direct PCR-MLFA system with different amount of cells.

Different density of (A) white blood cells in whole blood, oral epithelial cells in (C) saliva and (E) buccal swab sample were determined. The magnetic signal peak value of T line and C line of (B) blood sample, (D) saliva sample and (F) buccal swab sample at different cell concentration were measured by the magnetic signal reader.




Figure S5, Related to Figure 3. The reproducibility of whole blood PCR-MLFA system for three genotypes.

Table S1, Related to Figure 1. Primer sets used for *ALDH2* genotyping.

Primer sets	Introduced mismatch	Sequences (5'→3')
Primer set 1	Medium mismatch at the penultimate position	WT: TCCCACACTCACAGTTTTCACTa <u>T</u>
		M: TCCCACACTCACAGTTTTCACTa <u>C</u>
Primer set 2	Strong mismatch at the penultimate position	WT: TCCCACACTCACAGTTTTCACTc <u>T</u>
		M: TCCCACACTCACAGTTTTCACTc <u>C</u>
Primer set 3	Weak mismatch at the antepenultimate position	WT: TCCCACACTCACAGTTTTCACg <u>AT</u>
		M: TCCCACACTCACAGTTTTCACg <u>AC</u>
Primer set 4	Medium mismatch at the antepenultimate position	WT: TCCCACACTCACAGTTTTCACa <u>AT</u>
		M: TCCCACACTCACAGTTTTCACa <u>AC</u>
Primer set 5	Strong mismatch at the antepenultimate position	WT: TCCCACACTCACAGTTTTCACc <u>AT</u>
		M: TCCCACACTCACAGTTTTCACc <u>AC</u>
Primer set 6	No mismatch nucleotide introduced	WT: TCCCACACTCACAGTTTTCACTt <u>T</u>
		M: TCCCACACTCACAGTTTTCACTt <u>C</u>

Specific nucleotides are underlined. Mismatch nucleotides are in lowercase letters.

Table S2, Related to Figure 4. Gene test results and frequency of whole blood and dried blood spot samples of *ALDH22 polymorphism (type-specific concordance between direct PCR-LFA and sequencing)**

Direct PCR-LFA (blood, n=200)	Direct PCR-LFA (DBS, n=200)				Sequencing (n=200)				Total	Agreement	Frequency
	*2GG	*2GA	*2AA	Discrepant	*2GG	*2GA	*2AA	Discrepant			
*2GG	134	0	0	0	134	0	0	0	134	100%	67% (F1)
*2GA	0	64	0	0	0	64	0	0	64	100%	32% (F2)
*2AA	0	0	2	0	0	0	2	0	2	100%	1% (F3)
Total	134	64	2	0	134	64	2	0	200	100%	100%

Table S3, Related to Figure 4. Gene test results of buccal swab and saliva samples of *ALDH22 polymorphism (type-specific concordance between buccal swab samples and saliva samples using direct PCR-LFA system)**

Direct PCR-LFA (buccal swab, n=50)	Direct PCR-LFA (saliva, n=50)				Total	Agreement
	*2GG	*2GA	*2AA	Discrepant		
*2GG	16	0	0	0	16	100%
*2GA	0	33	0	0	33	100%
*2AA	0	0	1	0	1	100%
Total	16	33	1	0	50	100%

Transparent Methods

Oligonucleotides and reagents.

In this study, the *ALDH2**2 (rs671, G>A, Glu504Lys) polymorphism was selected as a target. All oligonucleotide primers were synthesized by Invitrogen Biotechnology Ltd. (Shanghai, China). For *ALDH2* polymorphism genotyping, specific primers were designed according to Little's report (Little, 1995). The ARMS Oligonucleotides that are complementary to a given DNA sequence except for a mismatched 3' terminus will not function as PCR primers under appropriate conditions. Additional deliberate mismatches should normally be introduced at the penultimate base of the primer to increase the specificity of the PCR reaction. Because different mismatches have been found to have different destabilizing effect, it is necessary to consider both terminal and penultimate mismatches together. If the mutation-induced terminal mismatch is strong, a weak additional mismatch should be selected, and vice versa. As shown in Figure S1, six primer sets were designed to select the optimal primer set. Finally, we chose the primer set 1 for further experiment: a 5' biotin-labeled common forward primer: 5'-Biotin-AC TTTGGGGCAATACAGGGG-3'; a 5' digoxin-labeled specific reverse primer for WT: 5'-Digoxin-TCCCACACTCACAGTTTTTCACTAT-3'; and a 5' digoxin-labeled specific reverse primer for M: 5'-Digoxin-TCCCACACTCACAGTTTTTCACTAC-3'. Meanwhile, two primers were used for sequencing: forward primer 5'-CTTTGGGGCAATACAGGGGG-3' and reverse primer 5'-AGAGGCTGGGTCTTTACCCT-3'.

All chemicals were of analytical grade and were purchased from reputable vendors. Buffers were prepared according to standard laboratory procedures. Anti-digoxin antibodies were purchased from Meridian Life Science, Inc. (Saco, ME, USA). Streptavidin was obtained from

Promega Biotech, Inc. (Madison, WI, USA). Goat anti-mouse IgG was from Joey Bioscience, Inc. (Shanghai, China). HotMaster Taq DNA polymerase and 10×PCR buffer were purchased from TIANGEN Biotech Co., Ltd. (Beijing, China). dNTPs (including dATP, dUTP, dCTP, and dGTP) and UDG were obtained from Shinegene Molecular Biotechnology Co., Ltd. (Shanghai, China).

Synthesis of anti-digoxin antibody-conjugated gold magnetic nanoparticles.

Gold magnetic nanoparticles (GMNPs, GoldMag) were synthesized and characterized according to methods described previously (Cui et al., 2001; Hui et al., 2012). The synthesized nanoparticles were functionalized using cetyltrimethylammonium bromide (CTAB) surfactant, modified with poly-acrylic acid (PAA) and conjugated with anti-digoxin antibodies using an EDC chemistry method according to a previous report (Yang et al., 2013). A Hitachi H-600 transmission electron microscope (TEM) (Hitachi, Tokyo, Japan) was employed to acquire images of the particles for structural assessment, whereas the size (diameter) was characterized using a Malvern Zetasizer ZS instrument (Malvern Instruments Ltd., Worcestershire, UK).

Assembly of the gold magnetic nanoparticle-based lateral flow strips.

The MLFA was performed over a strip composed of five components, a sample pad, conjugate pad, nitrocellulose membrane, absorbent pad and plastic cushion purchased from Merck Millipore (Darmstadt, Germany). Briefly, streptavidin (2 mg/mL) and goat anti-mouse IgG (2 mg/mL) were pre-immobilized using a BioJet HM3010 dispenser (BioDot Inc.,

California, USA) to make a defined T line and C line on a porous nitrocellulose membrane. Then, the solution containing GMNPs-anti-Dig was dispensed on the conjugate pad. The strips were dried via a 202-2AB Electrothermal constant temperature drying oven (Tianjin Taisite Instrument Co., Ltd., Tianjin, China) at 60 °C for 3 h and stored in a sealed aluminum foil bag at room temperature.

Sample preparation.

Blood samples or saliva samples: Peripheral blood samples were collected in EDTA-coated tubes, and saliva samples were obtained and placed in Eppendorf tubes. The collected samples (5 μ L for each) were mixed with 100 mM NaOH in a 1:2 ratio, and 5 μ L of the mixed solution was taken as a sample. Buccal swab samples: The buccal swab (purchased from Jiangsu Kangjian Medical Apparatus Co., Ltd., Jiangsu, China) head was cut off (approximately 5 mm underneath the head) and placed in 100 μ L of 100 mM NaOH, which was then heated at 65°C for 5 min; 5 μ L of the mixed solution was taken as a sample. DBS samples: One spot of dried blood was put into an Eppendorf tube and mixed with 100 mM NaOH, and 5 μ L of the mixed solution was taken as a sample. The sample solutions were prepared fresh and used as soon as possible. An inverted optical microscope (IX71, Olympus Optical Co., Ltd., Tokyo, Japan) with a 100 W halogen light source (U-LH100L-3, Olympus) was used to confirm the effect of the NaOH solution on blood cells. Cells were stained using a rapid Wright-Giemsa Staining Solution Kit (Sangon Biotech (Shanghai) Co., Ltd., Shanghai, China). A automated blood hematology analyzer (Beckman Coulter LH785, Beckman Coulter, Inc, Florida, USA) was used for cell counting.

Direct PCR amplification.

For each sample to be detected, two separate PCR reactions (M tube and WT tube) were run simultaneously using the same template. Each PCR reaction was performed using a final volume of 50 μ L, including 10 \times reaction buffer (10 mM Tris-HCl and 50 mM KCl), 0.2 mM of each dNTP (dATP, dUTP, dCTP and dGTP), 3 mM of MgCl₂, 0.5 U of HotMaster Taq DNA polymerase, 0.5 U of UDG polymerase, 50 nM common primer, 50 nM allele-specific primer (M primer in M tube and WT primer in WT tube), and 5 μ L of the prepared sample solution. All amplifications were performed using a 2720 Thermal Cycler (Applied Biosystems, Foster City, USA) according to the following parameters: two initial denaturation steps for 2 min at 50 $^{\circ}$ C and 3 min at 94 $^{\circ}$ C; 33 cycles of 5 s at 94 $^{\circ}$ C, 10 s at 60 $^{\circ}$ C and 30 s at 65 $^{\circ}$ C; and one step of 10 min at 65 $^{\circ}$ C.

Detection of PCR products with MLFA strips.

For each reaction tube, all of the PCR solution was pipetted onto the sample pad after amplification. The reference samples with *2GG, *2GA and *2AA genotypes (confirmed by sequencing) were used to validate the method. The sensitivity of the system was evaluated by detecting samples with a gradient of cell numbers. The pH change of the solution during the reaction was measured using a pHS-3C pH-meter (Shanghai Precision & Scientific Instrument Co., Ltd., Shanghai, China). The magnetic signal at the T line and C line of the strips was determined using a magnetic reader (Magna Bio Sciences, USA).

Clinical application and statistical analysis.

Matched fresh human whole blood and DBS samples were collected from 200 unrelated Chinese volunteers using EDTA-coated tubes at the Shaanxi Provincial People's Hospital (Xi'an, China). Matched saliva and buccal swab samples were obtained from 50 Chinese volunteers at Northwest University (Xi'an, China). Each individual provided informed consent. The study was approved by the Ethics Committee at the College of Life Sciences at Northwest University (Xi'an, China). All methods were performed in accordance with these approved guidelines. The genotype of each sample was analyzed via the direct PCR-MLFA assay and was compared with the results determined by DNA sequencing. Based on statistical analysis, the coincidence rate of the three genotypes and their total agreements were calculated to evaluate the accuracy of our method.

References

- Cui, Y., Hu, D., Fang, Y., and Jianbiao, M.A. (2001). Preparation and mechanism of Fe₃O₄/Au core/shell super-paramagnetic microspheres. *Sci. China-Chem.* *44*, 404-410.
- Hui, W., Shi, F., Yan, K., Peng, M., Cheng, X., Luo, Y., Chen, X., Roy, V.A., Cui, Y., and Wang, Z. (2012). Fe₃O₄/Au/Fe₃O₄ nanoflowers exhibiting tunable saturation magnetization and enhanced bioconjugation. *Nanoscale* *4*, 747-751.
- Little S. (1995). Amplification-refractory mutation system (ARMS) analysis of point mutations, In *Curr. Protoc. Hum. Genet.* *7*, (John Wiley & Sons), 9.8.1-9.8.12.
- Yang, D., Ma, J., Zhang, Q., Li, N., Yang, J., Raju, P.A., Peng, M., Luo, Y., Hui, W., and Chen, C. (2013). Polyelectrolyte-coated gold magnetic nanoparticles for immunoassay development: toward point of care diagnostics for syphilis screening. *Anal. Chem.* *85*, 6688-6695.