



Guideline

2015 HRS/EHRA/APHRS/SOLAECE expert consensus statement on optimal implantable cardioverter-defibrillator programming and testing ☆, ☆ ☆



Bruce L. Wilkoff, MD, FHRS, CCDS, (Chair)^a, Laurent Fauchier, MD, PhD, (Co-Chair)^{b,1}, Martin K. Stiles, MBChB, PhD, (Co-Chair)^{c,2}, Carlos A. Morillo, MD, FRCPC, FHRS, (Co-Chair)^{d,3}, Sana M. Al-Khatib, MD, MHSc, FHRS, CCDS^e, Jesøes Almendral, MD, PhD, FESC^{f,1}, Luis Aguinaga, MD, PhD, FACC, FESC^{g,3}, Ronald D. Berger, MD, PhD, FHRS^h, Alejandro Cuesta, MD, PhD, FESC^{i,3}, James P. Daubert, MD, FHRS^e, Sergio Dubner, MD, FACC^{j,ai,3}, Kenneth A. Ellenbogen, MD, FHRS^k, N.A. Mark Estes III, MD^{l,4}, Guilherme Fenelon, MD, PhD^{m,3}, Fermin C. Garcia, MD^{n,3}, Maurizio Gasparini, MD^{o,1}, David E. Haines, MD, FHRS^p, Jeff S. Healey, MD, MSc, FRCPC, FHRS^d, Jodie L. Hurtwitz, MD^{q,5}, Roberto Keegan, MD^{r,3}, Christof Kolb, MD^{s,1}, Karl-Heinz Kuck, MD, FHRS^{t,1}, Germanas Marinskis, MD, FESC^{u,1}, Martino Martinelli, MD, PhD^v, Mark McGuire, MBBS, PhD^{w,2}, Luis G. Molina, MD, DSc^{x,3}, Ken Okumura, MD, PhD^{y,2}, Alessandro Proclemer, MD^{z,1}, Andrea M. Russo, MD, FHRS^{aa}, Jagmeet P. Singh, MD, DPhil, FHRS^{ab}, Charles D. Swerdlow, MD, FHRS^{ac}, Wee Siong Teo, MBBS, FHRS^{ad,2}, William Uribe, MD, FHRS^{ae,3}, Sami Viskin, MD^{af,1}, Chun-Chieh Wang, MD^{ag,2}, Shu Zhang, MD^{ah,2}

☆☆Document Reviewers: Giuseppe Boriani, MD, PhD (Italy); Michele Brignole, MD, FESC (Italy); Alan Cheng, MD, FHRS (USA); Thomas C. Crawford, MD, FACC, FHRS (USA); Luigi Di Biase, MD, PhD, FACC, FHRS (USA); Kevin Donahue, MD (USA); Andrew E. Epstein, MD, FAHA, FACC, FHRS (USA); Michael E. Field, MD, FACC, FHRS (USA); Bulent Gorenek, MD, FACC, FESC (Turkey); Jin-Long Huang, MD, PhD (China); Julia H. Indik, MD, PhD, FACC, FAHA, FHRS (USA); Carsten W. Israel, MD (Germany); Mariell L. Jessup MD, FACC, FAHA, FESC (USA); Christophe Leclercq, MD, PhD (France); Robert J. MacFadyen, MD, PhD (UK); Christopher Madias, MD, FHRS (USA); Manlio F. Marquez, MD, FACC (Mexico); Brian Olshansky, MD, FACC, FAHA, FHRS (USA); Kristen K. Patton, MD (USA); Marwan M. Refaat, MD, mMBA, FACC, FAHA, FHRS, FASE, FESC, FACP, FAAMA (USA); Cynthia M. Tracy, MD, FACC, FAHA (USA); Gaurav A. Upadhyay, MD (USA); Diego Vanegas, MD, FHRS (Colombia); Paul J. Wang, MD, FHRS, CCDS (USA).

^a Cleveland Clinic, Cleveland, Ohio

^b Centre Hospitalier Universitaire Trousseau, Tours, France

^c Waikato Hospital, Hamilton, New Zealand

^d Department of Medicine, Cardiology Division, McMaster University–Population Health Research Institute, Hamilton, Canada

^e Duke University Medical Center, Durham, North Carolina

^f Grupo HM Hospitales, Universidad CEU San Pablo, Madrid, Spain

^g Centro Privado De Cardiología, Tucuman, Argentina

^h Johns Hopkins University, Baltimore, Maryland

ⁱ Servicio de Arritmias, Instituto de Cardiología Infantil, Montevideo, Uruguay

^j Clínica y Maternidad Suizo Argentina

Abbreviations: aCRT, adaptive cardiac resynchronization therapy; AF, atrial fibrillation; ATP, antitachycardia pacing; CI, confidence interval; CL, cycle length; CRT, cardiac resynchronization therapy; CRT-D, cardiac resynchronization therapy–defibrillator; DT, defibrillation testing; EEG, electroencephalography; EGM, electrogram; HF, heart failure; HR, hazard ratio; ICD, implantable cardioverter-defibrillator; LV, left ventricle; LVEF, left ventricular ejection fraction; MI, myocardial infarction; MVP, managed ventricular pacing; NCDR, National Cardiovascular Data Registry; NYHA, New York Heart Association; OR, odds ratio; PEA, peak endocardial acceleration; PVC, premature ventricular contraction; RCT, randomized clinical trial; RV, right ventricle; SCD, sudden cardiac death; S-ICD, subcutaneous implantable cardioverter-defibrillator; SVT, supraventricular tachycardia; TIA, transient ischemic attack; VF, ventricular fibrillation; VT, ventricular tachycardia (Heart Rhythm 2015;0:1–37)

^{*}Developed in partnership with and endorsed by the European Heart Rhythm Association (EHRA), the Asia Pacific Heart Rhythm Society (APHRS), and the Sociedad Latinoamericana de Estimulación Cardíaca y Electrofisiología (SOLAECE)–Latin American Society of Cardiac Pacing and Electrophysiology. Endorsed by the American College of Cardiology (ACC), American Heart Association (AHA), Canadian Heart Rhythm Society (CHRS), and Cardiac Arrhythmia Society of Southern Africa (CASSA).

Address reprint requests and correspondence: Ms. Emily Senerth, Heart Rhythm Society, 1325G Street NW, Suite400, Washington, DC 20005. E-mail address: esenerth@hrsonline.org

1547-5271/\$–see front matter © 2016 Heart Rhythm Society, European Heart Rhythm Association, a registered branch of the European Society of Cardiology, the Asia Pacific Heart Rhythm Society, and the Sociedad Latinoamericana de Estimulación Cardíaca y Electrofisiología (SOLAECE) <http://dx.doi.org/10.1016/j.hrthm.2015.11.018>

<http://dx.doi.org/10.1016/j.joa.2015.12.001>

1880-4276/© 2016 Japanese Heart Rhythm Society. Published by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

- ^k Virginia Commonwealth University Medical Center, Richmond, VA
- ^l New England Medical Center, Boston, Massachusetts
- ^m Federal University of S^o Paulo, S^o Paulo, Brazil
- ⁿ Hospital of the University of Pennsylvania, Philadelphia, Pennsylvania
- ^o Humanitas Research Hospital, Milan, Italy
- ^p William Beaumont Hospital Division of Cardiology, Royal Oak, Michigan
- ^q North Texas Heart Center, Dallas, Texas
- ^r Hospital Privado del Sur, Bahia Blanca, Argentina
- ^s Deutsches Herzzentrum Munchen, Munich, Germany
- ^t Allgemeines Krankenhaus St. Georg, Hamburg, Germany
- ^u Vilnius University, Clinic of Cardiac and Vascular Diseases, Lithuania
- ^v Instituto do Coraç^o, Universidade de S^o Paulo, S^o Paulo, Brazil
- ^w Royal Prince Alfred Hospital, Sydney, Australia
- ^x Mexico's National University, Mexico's General Hospital, Mexico City, Mexico
- ^y Hirosaki University Graduate School of Medicine, Hirosaki, Aomori, Japan
- ^z Azienda Ospedaliero Universitaria S. Maria della Misericordia– Udine, Udine, Italy
- ^{aa} Cooper University Hospital, Camden, New Jersey
- ^{ab} Massachusetts General Hospital, Harvard Medical School, Boston, Massachusetts
- ^{ac} Cedars-Sinai Medical Center, Beverly Hills, California
- ^{ad} National Heart Centre Singapore, Singapore, Singapore
- ^{ae} CES Cardiología and Centros Especializados San Vicente Fundación, Medellín y Rionegro, Colombia
- ^{af} Tel Aviv Sourasky Medical Center and Sackler School of Medicine, Tel Aviv University, Tel Aviv, Israel
- ^{ag} Chang Gung Memorial Hospital, Taipei, Taiwan
- ^{ah} National Center for Cardiovascular Disease and Beijing Fu Wai Hospital, Peking Union Medical College and China Academy of Medical Sciences, Beijing, China
- ^{ai} De Los Arcos Sanatorio, Buenos Aires, Argentina

ARTICLE INFO

Available online 1 February 2016

Keywords:

- Implantable cardioverter-defibrillator
- Bradycardia mode and rate
- Tachycardia detection
- Tachycardia therapy
- Defibrillation testing
- Programming

Contents

1. Introduction	3
2. Bradycardia Mode and Rate Programming	3
2.1. Single- or Dual-Chamber Pacing Mode	4
2.1.1. Evidence	4
3. Programming of Rate Modulation	4
4. Sinus Node Disease	4
5. Atrial Fibrillation and Atrioventricular Block	4
6. Intact Atrioventricular Conduction	5
6.1. Right Ventricular Pacing	5
7. Non-CRT Devices: Algorithms to Reduce Right Ventricular Stimulation	5
8. Cardiac Resynchronization Therapy: Consistent Delivery of Ventricular Pacing	6
9. Tachycardia Detection Programming	7
10. Duration Criteria for the Detection of Ventricular Arrhythmia	8
11. Limitations of Data on the Duration of Tachycardia Required for Detection	9
12. Rate Criteria for the Detection of Ventricular Arrhythmia	9
13. Single- or Multi-Zone Detection	10
14. Discrimination Between Supraventricular and Ventricular Arrhythmia	10
15. SVT-VT Discriminator Components	10
16. Rejection of Sinus Tachycardia by Onset	10
16.1. Rejection of AF by Ventricular Interval Regularity	11
16.2. Diagnosis of VT by Dual-Chamber Components: Atrial vs Ventricular Rate and Atrioventricular Association	11
16.3. The Ventricular Electrogram Morphology Discriminator	11
17. SVT-VT Discrimination Algorithms	11

¹ Representative of the European Heart Rhythm Association (EHRA).
² Representative of the Asia-Pacific Heart Rhythm Society (APHRs).
³ Representative of the American College of Cardiology (ACC).
⁴ Representative of the American Heart Association (AHA).
⁵ Representative of the Sociedad Latinoamericana de Estimulacion Cardiaca y Electrofisiologia (SOLAECE).

18.	Assessing Clinical Benefits and Risks	11
18.1.	What Evidence Supports a Benefit?	11
18.2.	Which Patients are Most Likely to Benefit, and Which are Least Likely to Benefit?	11
18.3.	What are the Risks?	11
19.	Additional Considerations	12
19.1.	SVT Limit	12
19.2.	Duration-Based “Safety-Net” Features to Override Discriminators	12
19.3.	Dual-Chamber vs Single-Chamber Algorithms	12
19.4.	Ventricular Oversensing	12
20.	Programming to Reduce T-Wave Oversensing	12
21.	Lead-Related Oversensing	13
22.	The Subcutaneous Defibrillator (S-ICD)	13
23.	Integrating Tachycardia Detection Data Into Programming Recommendations	13
24.	Tachycardia Therapy Programming	15
25.	Benefits and Risks	15
26.	Classification of Therapy	15
27.	Appropriate	15
28.	Inappropriate	16
29.	Avoidable	16
30.	Phantom	16
31.	Unintended Consequences of ICD Therapy and ICD Therapy Programming	16
31.1.	ATP	16
31.2.	Customized vs Strategic Programming	16
31.3.	Secondary Prevention	16
32.	Intraprocedural Testing of Defibrillation Efficacy	17
33.	Periprocedural Mortality	17
34.	DT-Related Complications	17
35.	Contraindications to Defibrillation Threshold Testing	20
36.	S-ICD	20
37.	Conclusion	20
Appendix		
Disclosures		
	References	24

1. Introduction

Implantable cardioverter-defibrillator (ICD) therapy is clearly an effective therapy for selected patients in definable populations. The benefits and risks of ICD therapy are directly impacted by programming and surgical decisions. This flexibility is both a great strength and a weakness, for which there has been no prior official discussion or guidance. It is the consensus of the 4 continental electrophysiology societies that there are 4 important clinical issues for which there are sufficient ICD clinical and trial data to provide evidence-based expert guidance. This document systematically describes the greater than 80% (83%–100%, mean 96%) required consensus achieved for each recommendation by official balloting in regard to the programming of (1) bradycardia mode and rate, (2) tachycardia detection, (3) tachycardia therapy, and (4) the intraprocedural testing of defibrillation efficacy. Representatives nominated by the Heart Rhythm Society (HRS), European Heart Rhythm Association (EHRA), Asian Pacific Heart Rhythm Society (APHRS), and the *Sociedad Latinoamericana de Estimulación Cardíaca y Electrofisiología* (SOLAECE-Latin American Society of Cardiac Pacing and Electrophysiology) participated in the project definition, the literature review, the recommendation development, the writing of the document, and its approval. The 32 recommendations were balloted by the 35 writing committee members and were approved by an average of 96%.

The classification of the recommendations and the level of evidence follow the recently updated ACC/AHA standard [1,2]. Class I is a strong recommendation, denoting a benefit greatly exceeding risk. Class IIa is a somewhat weaker recommendation, with a benefit probably exceeding risk, and Class IIb denotes a benefit equivalent to or possibly exceeding risk. Class III is a

recommendation against a specific treatment because either there is no net benefit or there is net harm. Level of Evidence A denotes the highest level of evidence from more than 1 high-quality randomized clinical trial (RCT), a meta-analysis of high-quality RCTs, or RCTs corroborated by high-quality registry studies. Level of evidence B indicates moderate-quality evidence from either RCTs with a meta-analysis (B-R) or well-executed nonrandomized trials with a meta-analysis (B-NR). Level of evidence C indicates randomized or nonrandomized observational or registry studies with limited data (C-LD) or from expert opinions (C-EO) based on clinical experience in the absence of credible published evidence. These recommendations were also subject to a 1-month public comment period. Each society then officially reviewed, commented, edited, and endorsed the final document and recommendations. All author and peer reviewer disclosure information is provided in [Appendix A](#).

The care of individual patients must be provided in context of their specific clinical condition and the data available on that patient. Although the recommendations in this document provide guidance for a strategic approach to ICD programming, as an individual patient's condition changes or progresses and additional clinical considerations become apparent, the programming of their ICDs must reflect those changes. Remote and in-person interrogations of the ICD and clinical monitoring must continue to inform the programming choices made for each patient. The recommendations in this document specifically target adult patients and might not be applicable to pediatric patients, particularly when programming rate criteria.

Please consider that each ICD has specific programmable options that might not be specifically addressed by the 32 distinctive recommendations in this document. Appendix B,

published online (<http://www.hrsonline.org/appendix-b>), contains the writing committee's translations specific to each manufacturer and is intended to best approximate the recommended behaviors for each available ICD model.

2. Bradycardia Mode and Rate Programming

2.1. Single- or Dual-Chamber Pacing Mode

2.1.1. Evidence

Because the ICD is primarily indicated for tachycardia therapy, there might be some uncertainty regarding optimal bradycardia management for ICD patients. Data from clinical studies adequately address only the programmed mode rather than the number of leads implanted, the number of chambers stimulated, or how frequently the patients required bradycardia support. It is of note that most information on pacing modes has been collected from pacemaker patients, and these patients are clinically distinct from ICD recipients. Dual-chamber pacing (atrial and ventricular) has been compared with single-chamber pacing (atrial or ventricular) in patients with bradycardia in 5 multicenter, parallel, randomized trials, in 1 meta-analysis of randomized trials, and in 1 systematic review that also included 30 randomized crossover comparisons and 4 economic analyses [3–9]. Meta analyses comparing dual-chamber to single-chamber ICDs did not evaluate pacing modes [10,11]. Compared with single-chamber pacing, dual-chamber pacing results in small but potentially significant benefits in patients with sinus node disease and/or atrioventricular block. No difference in mortality has been observed between ventricular pacing modes and dual-chamber pacing modes. Dual-chamber pacing was associated with a lower rate of atrial fibrillation (AF) and stroke [12]. The benefit in terms of AF prevention was more marked in trials comprised of patients with sinus node disease. Although trends in favor of dual-chamber pacing have been observed in some trials, there was no benefit in terms of heart failure (HF). In patients without symptomatic bradycardia, however, the Dual Chamber and VVI Implantable Defibrillator (DAVID) trial in ICD recipients showed that one specific choice of dual-chamber rate-responsive (DDDR) programming parameters led to poorer outcomes than VVI backup pacing, most likely secondary to unnecessary right ventricular (RV) pacing. The fact that RV stimulation was responsible was reinforced in the DAVID II trial, in which AAI pacing was demonstrated to be non-inferior to VVI backup pacing [13].

Approximately a quarter of patients with either sinus node disease or atrioventricular block develop “pacemaker syndrome” with VVI pacing usually associated with retrograde (ventricular to atrial) conduction, which in turn is associated with a reduction in the quality of life [14]. In crossover trials, symptoms of pacemaker syndrome (dyspnea, dizziness, palpitations, pulsations, and chest pain) were reduced by reprogramming to a dual-chamber mode [14]. Dual-chamber pacing is associated with better exercise performance compared with single-chamber VVI pacing without rate adaptation, but produces similar exercise performance when compared with rate-responsive VVIR pacing. Because of the additional lead, dual-chamber devices involve longer implantation times, have a higher risk of complications, and are more expensive. However, because of the additional clinical consequences of pacemaker syndrome and AF (and its sequelae), the overall cost difference between single- and dual-pacing systems is moderated.

In patients with persistent sinus bradycardia, atrial rather than ventricular dual-chamber pacing is the pacing mode of choice. There is evidence for superiority of atrial-based pacing over ventricular pacing for patients who require pacing for a significant proportion of the day. The evidence is stronger for patients with

sinus node disease, in whom dual-chamber pacing confers a modest reduction in AF and stroke, but not in hospitalization for HF or death compared with ventricular pacing. In patients with acquired atrioventricular block, large randomized parallel trials were unable to demonstrate the superiority of dual-chamber pacing over ventricular pacing with regard to hard clinical endpoints of mortality and morbidity [4,6–8]. The benefit of dual-chamber over ventricular pacing is primarily due to the avoidance of pacemaker syndrome and to improved exercise capacity [14]. Even if it is a softer endpoint, pacemaker syndrome is associated with a reduction in quality of life that justifies the preference for dual-chamber pacing when reasonable; thus, there is strong evidence for the superiority of dual-chamber pacing over ventricular pacing that is limited to symptom improvement. Conversely, there is strong evidence of nonsuperiority with regard to survival and morbidity. The net result is that the indications for programming the dual-chamber modes are weaker and the choice regarding the pacing mode should be individualized, taking into consideration the increased complication risk and costs of dual-chamber devices. Because ICD patients usually do not require bradycardia support, with the exception of patients who require cardiac resynchronization, programming choices should avoid pacing and in particular avoid single ventricular pacing, if possible [15,16].

3. Programming of Rate Modulation

The benefit of rate response programming has been evaluated in patients with bradycardia in 5 multicenter, randomized trials and in 1 systematic review that also included 7 single-center studies [17–22]. Most of these data were obtained from pacemaker studies and must be interpreted in that light.

Although there is evidence of the superiority of VVIR pacing compared with VVI pacing in improving quality of life and exercise capacity, improvements in exercise capacity with DDDR compared with DDD have been inconsistent. In 2 small studies on patients with chronotropic incompetence comparing DDD and DDDR pacing, the latter had improved quality of life and exercise capacity; however, a larger, multicenter randomized trial (Advanced Elements of Pacing Randomized Controlled Trial [ADEPT]) failed to show a difference in patients with a modest blunted heart rate response to exercise [17–19]. In addition, DDDR programming in cardiac resynchronization therapy (CRT) patients has the potential to impair AV synchrony and timing. It should be noted that trials evaluating CRT generally did not use rate-responsive pacing, and many in fact avoided atrial stimulation using atrial sensed and ventricular paced pacing modes with a lower base rate. However, the Pacing Evaluation- Atrial Support Study in Cardiac Resynchronization Therapy (PEGASUS CRT) trial is the exception and did not demonstrate adverse impact on mortality and HF events [23].

4. Sinus Node Disease

In patients with persistent or intermittent sinus node dysfunction or chronotropic incompetence, the first choice is DDDR with algorithms responding to intermittent atrioventricular conduction. There is sufficient evidence for the superiority of VVIR compared with VVI in improving quality of life and exercise capacity. The evidence is much weaker in dual-chamber pacing (DDDR vs DDD).

Although only an issue when there is some concomitant AV block, the upper rate limit should be programmed higher than the fastest spontaneous sinus rhythm to avoid upper rate limit behavior. To avoid symptomatic bradycardia, the lower rate should be

programmed on an individual basis, according to the clinical characteristics and the underlying cardiac substrate of the patient.

5. Atrial Fibrillation and Atrioventricular Block

Patients with permanent AF and either spontaneous or AV junctional ablation-induced high-degree atrioventricular block have little to no chronotropic response to exercise; thus, VVIR pacing is associated with better exercise performance, improved daily activities, improved quality of life, and decreased symptoms of shortness of breath, chest pain, and heart palpitations, compared with VVI [20–22,24–26]. Therefore, rate-adaptive pacing is the first choice of pacing mode; fixed-rate VVI pacing should be abandoned in patients with permanent AF and atrioventricular block. It is the experts' opinion that the minimum rate can be programmed higher (e. g., 70 bpm) than for sinus rhythm patients, in an attempt to compensate for the loss of active atrial filling. In addition, the maximum sensor rate should be programmed restrictively (e. g., 110–120 bpm) to avoid “overpacing” (i.e., pacing with a heart rate faster than necessary), which can be symptomatic, particularly in patients with coronary artery disease. In a small study, however, it was found that rate-responsive pacing could be safe and effective in patients with angina pectoris, without an increase in subjective or objective signs of ischemia [25]. The lower rate should be programmed on an individual basis, according to the clinical characteristics and the underlying cardiac substrate of the patient. The clinical benefit of programming a lower resting rate at night based on internal clocks has not been evaluated in ICD patients. There is some concern that atrioventricular junction ablation and permanent ventricular pacing might predispose the patient to an increased risk of sudden cardiac death (SCD) related to a bradycardia-dependent prolongation of the QT interval. This risk might be overcome by setting the ventricular pacing rate to a minimum of 80 or 90 bpm for the first 1–2 months following the atrioventricular junction ablation, then reducing it to a conventional 60–70 bpm [27,28]. Not all patients with AF and milder forms of atrioventricular block will require a high percentage of ventricular pacing or have a wide QRS. Physicians should consider the risk of increasing preexisting left ventricular (LV) dysfunction with RV pacing vs improved chronotropic responsiveness and the potential value of CRT.

6. Intact Atrioventricular Conduction

6.1. Right Ventricular Pacing

The results of a number of large-scale, prospective randomized trials demonstrated a significant reduction in AF in pacemaker patients with atrial-based pacing (AAI or DDD) compared with patients with ventricular-based pacing [4,8,29]. In the Mode Selection Trial, which enrolled 2010 patients with sick sinus syndrome, the risk of AF increased linearly with the increasing percentage of RV pacing [30]. At the same time, deleterious effects of RV pacing in patients with LV dysfunction (left ventricular ejection fraction [LVEF] \leq 40%) implanted with dual-chamber ICD systems were observed in the Dual Chamber and VVI Implantable Defibrillator (DAVID) trial, which included 506 ICD patients without indications for bradycardia pacing. Patients within the DDDR-70 group (with paced and sensed atrioventricular delays of 170 and 150 ms, respectively, in most of the DDDR group patients) showed a trend toward higher mortality and an increased incidence of HF compared with the patients programmed to ventricular backup

pacing—the VVI-40 group. Within the DDDR-70 group, there were more cardiac events when the percentage of ventricular pacing exceeded 40% ($P=.09$) compared with patients with 95% RV stimulation (DDDR-70) or [31,32]. However, a more detailed *post hoc* analysis of the Inhibition of Unnecessary RV Pacing With Atrial-Ventricular Search Hysteresis in ICDs (INTRINSIC RV) trial revealed that the most favorable clinical results were not in the VVI groups with the least percentage of RV pacing but in the subgroup that had DDD pacing with longer atrioventricular delays and 11%–19% of ventricular pacing. This parameter selection probably helped patients to avoid exceedingly low heart rates while preserving intrinsic atrioventricular conduction most of the time [31,33]. In the Second Multicenter Automated Defibrillator Implantation Trial (MADIT II), a higher risk of HF was observed in patients who had a greater than 50% burden of RV pacing [34]. In another large observational study of 456 ICD patients without HF at baseline, a high RV pacing burden (RV pacing more than 50% of the time) was associated with an increased risk of HF events and appropriate ICD shocks [35]. Optimally, RV stimulation should be avoided, but the precise tradeoff between the percentage of ventricular pacing and atrioventricular timing is unclear in non-CRT patients.

7. Non-CRT Devices: Algorithms to Reduce Right Ventricular Stimulation

The importance of reducing or avoiding RV pacing in ICD patients with LV dysfunction was illustrated in the DAVID trial [31]. The feasibility of algorithms designed to decrease the burden of unnecessary ventricular pacing has been demonstrated in patients with dual-chamber pacemakers [36–38]. These algorithms usually provide functional AAI pacing with monitoring of atrioventricular conduction and an automatic mode switch from AAI to DDD during episodes of atrioventricular block. Some studies directly compared various algorithms to decrease ventricular pacing, showing that a “managed ventricular pacing” (MVP) algorithm resulted in greater ventricular pacing reduction than an “atrioventricular search” algorithm [39,40]; however, no randomized studies comparing these two algorithms with respect to important cardiovascular endpoints (e.g., HF, cardiac death) have been performed. The results of the studies on these pacing algorithms are summarized in Table 1.

Unnecessary RV pacing should be minimized by using specific algorithms or programming longer atrioventricular delays, and this process is more important for patients with a higher risk of AF or who already have poorer LV function [49]. Patients with longer baseline PR intervals have a higher risk of AF regardless of the percentage of ventricular pacing or the length of the programmed atrioventricular interval [50]. Use of the AAIR pacing mode with exceedingly long atrioventricular conduction times can lead to “AAIR pacemaker syndrome” and actually increases the risk of AF compared with the DDDR mode, as was shown in the Danish Multicenter Randomized Trial on Single Lead Atrial versus Dual-Chamber Pacing in Sick Sinus Syndrome (DANPACE) [3,51]. Therefore, excessively long atrioventricular delays resulting in nonphysiologic atrioventricular contraction patterns should be avoided. The potential harm of atrial pacing with a prolonged atrioventricular delay was also demonstrated in the MVP trial. In the MVP trial, dual-chamber pacing with the MVP algorithm was not superior to ventricular backup pacing (VVI 40 bpm) with respect to HF events. After a follow-up of 2.4 years, there was an apparent increase in HF events that was limited primarily to patients with a baseline PR interval of > 230 ms (mean PR of 255–260 ms) [42]. Long atrioventricular intervals also predispose the

Table 1
Influence of pacing modes and algorithms on clinical endpoints.

Study	Patients (PM/ICD)	Results and remarks
SAVE PACE, randomized multicenter (2007) [41]	1065 (PM)	40% relative risk reduction of AF in the MVP group compared with DDD pacing (4.8% absolute risk reduction).
MVP, randomized multicenter (2011) [42]	1030 (ICD)	No superiority of MVP over VVI-40 in terms of AF, VT/VF, quality of life, HF.
Steinbach et al, retrospective single-center (2011) [43]	102 (PM)	In patients over 75 years of age, MVP showed lower rates of HF episodes and all-cause mortality than conventional DDD pacing
long-MinVPACE, randomized single-center (2011) [44]	66 (PM)	Less RV pacing, less AF burden in MinVP group patients compared with DDDR (mean 12.8 vs 47.6%). Chosen AV/PV delay (150/130 ms) was probably too short in the DDDR (control) group.
Generation MVP, observational multicenter (2012) [45]	220 (PM)	Significantly fewer atrial arrhythmias when programmed to MVP compared with DDD.
PreFER MVP, randomized multicenter (2014) [46]	605 (556PM, 49 ICD)	No difference between cardiovascular hospitalization, AF, and the composite of death and hospitalization between the MVP and DDD groups. The authors stated that “patients were enrolled upon elective replacement of the device, and were healthy enough to survive the first device without experiencing a significant decrease in LV function.”
MINERVA, randomized multicenter (2014) [47]	1300 (PM)	AF burden: no superiority of MVP pacing compared with the DDDR mode (AV/ PV delay > 180/210 ms in greater than 60% of patients, 53% of RV pacing). MVP in combination with atrial antitachycardia pacing was superior to both DDDR and MVP-only.
COMPARE, randomized multicenter (2014) [48]	385 (PM)	Lower percentage of ventricular pacing (%VP) in the MVP group compared with the SearchAV+ group. A trend in the correlation between %VP and AT/AF burden.

AT=atrial tachycardia; HF=heart failure; MVP=Managed Ventricular Pacing; PM=pacemaker.

patient to repetitive atrioventricular reentrant rhythms, “repetitive nonreentrant VA synchrony,” or “atrioventricular desynchronization arrhythmia,” which manifest as mode switching but which also cause sustained episodes with poor hemodynamics [52]. Thus, based on the available data, it appears that atrial pacing with excessively long atrioventricular delays should be avoided.

Algorithms that minimize ventricular pacing sometimes lead to inadvertent bradycardia or spontaneous premature, beat-related short-long-short RR interval sequences with proarrhythmic potential [53–55]. However, in a study retrospectively analyzing the onset of ventricular tachycardia (VT) in ICD patients, the MVP mode was less frequently associated with the onset of VT compared with the DDD and VVI modes [54]. Atrioventricular decoupling (greater than 40% of atrioventricular intervals exceeding 300 ms) was observed in 14% of the ICD patients in the Marquis ICD MVP study, which might have a negative effect on ventricular filling [56].

In ICD patients with structural heart disease, spontaneous atrioventricular conduction can become prolonged instead of shortening, with increased atrial paced heart rates [33]. This outcome frequently leads to a higher percentage of ventricular paced complexes. In view of the results of the ADEPT trial, which failed to demonstrate the clinical superiority of combined rate modulation and DDD pacing, the need for and aggressiveness of sensor-driven rate responses should be individualized or eliminated [19]. Rate-dependent shortening of atrioventricular delay could have the same effect and should usually be avoided.

Patients with hypertrophic cardiomyopathy represent a small but intricate subset of the ICD population for whom pacing has not been demonstrated to be a consistently effective treatment for outflow tract obstruction. However, according to the 2011 ACCF/AHA Hypertrophic Cardiomyopathy Guideline, dual-chamber ICDs are reasonable for patients with resting LV outflow tract gradients more than 50 mm Hg, and who have indications for ICD implantation to reduce mortality [57]. In these patients, atrioventricular delays should be individually programmed to be short enough to achieve RV preexcitation and decrease LV outflow tract gradient, but not too short, which would impair LV filling; usually in the ranges of 60–150 ms [58,59]. There are few studies of pacing modes in these patients, and they are limited by small numbers and the failure to quantify important cardiac outcomes.

In conclusion, atrioventricular interval programming and choosing between DDDR and MVP or other atrioventricular interval management modes should be performed on an individual basis. The goal is to minimize the percentage of RV pacing and to avoid atrial-based pacing with atrioventricular intervals exceeding 250–300 ms leading to atrioventricular uncoupling. In patients with prolonged PR intervals and impaired LV function, biventricular pacing can be considered.

8. Cardiac Resynchronization Therapy: Consistent Delivery of Ventricular Pacing

CRT in combination with a defibrillator device (CRT-D) improves survival and cardiac function in patients with LV systolic dysfunction, prolonged QRS duration, and mild-to-severe HF [60–62]. The beneficial effect of CRT-D compared with ICD is likely to be derived from biventricular pacing, with a decrease in dyssynchrony and an improvement in cardiac function. The percentage of biventricular pacing capture in the ventricles can be negatively influenced by a number of factors, including atrial tachyarrhythmias, premature ventricular complexes, and programming of the atrioventricular delay, giving way to the intrinsic conduction of the patient and a reduced percentage of biventricular pacing. Some large observational studies have investigated the optimal level of biventricular pacing percentage and found a higher percentage to be associated with more pronounced CRT benefits. An optimal CRT benefit was observed with a biventricular pacing percentage as close to 100% as possible [63–66].

In the analysis of the left bundle branch block population in the MADIT-CRT trial, those patients with less than 90% biventricular pacing had similar rates of HF and death compared with the patients randomized to no CRT. By contrast, biventricular pacing exceeding 90% was associated with a benefit of CRT-D in terms of HF or death when compared with ICD patients and no CRT. Biventricular pacing 97% and greater was associated with a further reduction in HF or death and a significant reduction in death alone. Consistently, every 1% increase in biventricular pacing percentage was associated with a 6% risk reduction in HF or death, a 10% risk reduction in death alone, and an increase in LV reverse remodeling [67]. Therefore, in ICD patients with biventricular pacing, it can be beneficial to adjust the therapy to produce the highest achievable percentage of ventricular

pacing, preferably above 98%, to improve survival and reduce HF hospitalization. Approaches to increasing the percentage of biventricular pacing include programming shorter but hemodynamically appropriate atrioventricular delays and minimizing atrial and ventricular ectopic activity and tachyarrhythmias.

Optimizing the location of ventricular pacing sites and the timing of the pacing pulses can significantly improve cardiac hemodynamics in CRT patients. Echocardiographic optimization of atrioventricular delays in CRT patients can alleviate HF symptoms and increase exercise capacity compared with nominal programming, particularly when approaching nonresponding populations [68]. However, echocardiographic optimization in the PROSPECT study did not support this approach in a randomized trial, and the Frequent Optimization Study Using the QuickOpt Method (FREEDOM) trials failed to provide evidence supporting the benefit of CRT optimization and did not demonstrate superiority of the respective algorithms over nominal or empiric programming [69–71]. There are limited data supporting the use of LV- only stimulation in a small subset of patients who fail to respond to biventricular stimulation [72]. Adaptive CRT (aCRT) is an algorithm that periodically measures intrinsic conduction and dynamically adjusts CRT pacing parameters. The algorithm withholds RV pacing when intrinsic electrical conduction to the RV is normal and provides adjustment of CRT pacing parameters based on electrical conduction. A prospective, multicenter, randomized, double-blind clinical trial demonstrated the safety and efficacy of the aCRT algorithm [73]. This algorithm can increase the longevity of the implantable device and replace a manual device optimization process with an automatic ambulatory algorithm, although echo optimization might still be needed, at least in nonresponders. The Clinical Evaluation on Advanced Resynchronization (CLEAR) study assessed the effects of CRT with automatically optimized atrioventricular and interventricular delays, based on a peak endocardial acceleration (PEA) signal system. PEA-based optimization of CRT in patients with HF significantly increased the proportion of patients who improved with therapy during follow-up, mainly through an improved New York Heart Association (NYHA) class [74].

pacing in preference to single-chamber ventricular pacing to avoid pacemaker syndrome and to improve quality of life.

In ICD patients who have sinus rhythm, mild-to-moderate LV dysfunction, and atrioventricular block where ventricular pacing is expected, it is reasonable to provide CRT in preference to dual-chamber ventricular pacing to improve the combination of HF hospitalization, LV enlargement, and death.	Ila	B-R
In ICD patients who have chronicotropic incompetence, it can be beneficial to program the ICD to provide sensor-augmented rate response, especially if the patient is young and physically active.	Ila	B-NR
In dual-chamber ICD patients with native PR intervals of 230 ms or less, it can be beneficial to program the mode, automatic mode change, and rate response so that the patient's native atrioventricular conduction minimizes ventricular pacing.	Ila	B-R
In biventricular pacing ICD patients, it can be beneficial to adjust the therapy to produce the highest achievable percentage of ventricular pacing, preferably above 98%, to improve survival and reduce HF hospitalization.	Ila	B-NR
In biventricular pacing ICD patients, it can be reasonable to activate the algorithms providing automatic adjustment of atrioventricular delay and/or LV-RV offset to obtain a high percentage of synchronized pacing and reduce the incidence of clinical events.	Ilb	B-R

Bradycardia Mode and Rate Programming Recommendations	Class of Recommendation	Level of Evidence
In ICD patients who also have sinus node disease and guideline-supported indications for a bradycardia pacemaker, it is beneficial to provide dual-chamber pacing to reduce the risk of AF and stroke, to avoid pacemaker syndrome, and to improve quality of life.	I	B-R
In single- or dual-chamber ICD patients without guideline-supported indications for bradycardia pacing, adjusting the pacing parameters is recommended so that ventricular stimulation is minimized to improve survival and reduce HF hospitalization.	I	B-R
In ICD patients who have sinus rhythm, no or only mild LV dysfunction, and atrioventricular block where ventricular pacing is expected, it is reasonable to provide dual-chamber	Ila	B-R

9. Tachycardia Detection Programming

Following significant technological changes in ICDs in recent years, the concept of optimal ICD programming has changed dramatically. From the dawn of this therapy in the early 1980s to the first decade of the 21st century, the rapid detection and treatment of VT and ventricular fibrillation (VF) have been stressed. The argument for rapid detection of VT and VF derived from a number of factors. Initial skepticism regarding the feasibility of sudden death prevention with ICDs, the fact that early ICD patients had all survived one or more cardiac arrests, concern for undersensing and underdetection (of VF in particular), demonstration of an increasing defibrillation threshold with prolonged VF duration, and the increased energy requirement of monophasic defibrilla-

tion all created a culture of programming for rapid tachycardia detection and the shortest possible time to initial therapy [75–77]. The initial generations of ICDs did not record and save electrograms (EGMs), leading to a reduced appreciation for the frequency and impact of inappropriate shocks. With the advent and then dominance of primary prevention indications, avoidable shocks assumed a relatively larger proportion of total therapy [78–83]. Gradually, publications have increased awareness of the frequency and the diverse range of adverse outcomes associated with avoidable ICD therapy, and have demonstrated that avoidable ICD shocks can be reduced by evidence-based programming of the detection rate, detection duration, antitachycardia pacing (ATP), algorithms that discriminate supraventricular tachycardia (SVT) from VT, and specific programming to minimize the sensing of noise [81–92].

10. Duration Criteria for the Detection of Ventricular Arrhythmia

Until recently, default device programming used short- duration “detection” criteria that varied by manufacturer and a tachycardia rate of approximately 2.8 to 5 seconds before either ATP or charging (including detection time plus duration or number of intervals) [82,93]. With increased awareness of the potential harm from inappropriate shocks and the realization from stored pacemaker EGMs that even long episodes of VT can self-terminate, a strategy of prolonged detection settings has been explored. This strategy allows episodes to self-terminate without requiring device intervention and reduces inappropriate therapy for non-malignant arrhythmias. The benefit of programming a prolonged detection duration (30 of 40 beats) was first reported in the Prevention Parameters Evaluation (PREPARE) study on exclusively primary prevention subjects (n=700), and compared outcomes to a historical ICD cohort programmed at “conventional detection delays” with about half programmed to 12 of 16 intervals within the programmed detection zone and half to 18 of 24 intervals [94]. The programming in PREPARE demonstrated a significant reduction in inappropriate shocks for supraventricular arrhythmia and in avoidable shocks for VT. In addition, a composite endpoint was reduced as well: the morbidity index, which consists of shocks, syncope, and untreated sustained VT. Within the limitations of a nonrandomized study, it was concluded that extending detection times reduces shocks without increasing serious adverse sequelae.

In 2009, the Role of Long-Detection Window Programming in Patients with Left Ventricular Dysfunction, Non- Ischemic Etiology in Primary Prevention Treated with a Biventricular ICD (RELEVANT) study confirmed and expanded the results of the PREPARE trial in a cohort of 324 primary prevention CRT-D patients with nonischemic cardiomyopathy [95]. The subjects were treated with simplified VT management, which implies much longer detection for VF episodes (30 of 40) compared with the control group (12 of 16) and a monitor-only window for VT. As in PREPARE, the RELEVANT study group experienced a significantly reduced burden of ICD interventions (81% reduction) without increasing the incidence of syncope. Fewer inappropriate shocks and HF hospitalizations were reported in the RELEVANT study group compared with the control group.

The Multicenter Automatic Defibrillator Implantation Trial: Reduce Inappropriate Therapy (MADIT-RIT), a 3- arm study, compared a conventional programming strategy (a 1-second delay for VF [equivalent to approximately 12 intervals including detection plus delay] and a 2.5-second delay for VT detection [equivalent to approximately 16 intervals including detection plus delay]) (Arm A) to both a high-rate cutoff with a VF zone starting at 200 bpm (Arm B) (discussed in section Rate Criteria for the Detection of

Ventricular Arrhythmia and discussed as referenced in reference [96].) and to a delayed therapy strategy with a 60- second delay for rates between 170 and 199 bpm, a 12- second delay at 200 to 249 bpm, and a 2.5-second delay at 250 bpm (Arm C) [96]. The MADIT-RIT population was exclusively primary prevention and included approximately an equal proportion of nonischemic and ischemic cardiomyopathy patients. All the patients were implanted with either a dual-chamber ICD or a CRT-D programmed to deliver ATP before charging. After a mean 1.4-year follow-up, the prolonged detection group (Arm C) was associated with a reduction in treated VT/VF leading to a 76% reduction in the primary endpoint of the first inappropriate therapy (*P* The Avoid Delivering Therapies for Non-Sustained Arrhythmias in ICD Patients III (ADVANCE III) trial reported that a long detection was associated with a highly significant reduction of overall therapies (appropriate and inappropriate ATP and/ or shocks), inappropriate shocks, and all-cause hospitalizations [97]. Importantly, like PREPARE, RELEVANT, and MADIT-RIT, the extended detection duration used in the ADVANCE III trial (30 of 40) did not negatively impact the rate of syncope events. There was no significant difference in mortality between the optimal and the conventional programming groups. Compared with the MADIT-RIT trial, the ADVANCE III control group had a longer detection duration (primarily in the VF zone), and enrolled a larger cohort of subjects covering all ICD types (single, dual, and CRT with ATP delivered during charging) for both primary and secondary prevention indications. Finally, the Programming Implantable Cardioverter-Defibrillators in Patients With Primary Prevention Indication (PROVIDE) trial randomized 1670 patients to conventional programming (12-beat detection in each of 2 zones) or experimental programming (2 VT and 1 VF zone requiring 25-, 18-, and 12-beat detection, respectively) [98]. PROVIDE observed a significant 36% reduction in the 2-year all-cause shock rate and an improved survival (hazard ratio [HR] 0.7; 95% confidence interval [CI] 0.50–0.98; *P*=.036).

Whereas PREPARE, RELEVANT, MADIT-RIT, and PROVIDE only enrolled primary prevention patients, a subset of the ADVANCE III study evaluated the efficacy and safety of a long-detection approach in secondary prevention patients who have a known higher burden of arrhythmic episodes. In this particular subset of 25% of the enrolled patients, ADVANCE III reported that a long detection duration reduced the overall therapies delivered, primarily due to a significant 36% reduction in appropriate shocks [99]. Syncope episodes related to arrhythmic events and deaths were similar between the 2 groups. Following shortly on the heels of these trials, 2 meta-analyses including the above studies were published in 2014. Tan et al presented the data from the RELEVANT, PREPARE, MADIT-RIT, ADVANCE III, PROVIDE, and EMPIRIC trials [100,101]. A 30% reduction in the risk of death was found in the therapy reduction group when including all 6 studies; however, similar results were observed when separately considering the 4 randomized trials and the 2 observational studies. Data on the appropriateness of shocks were available only for RELEVANT, MADIT-RIT, ADVANCE III, and PROVIDE, and a 50% reduction in inappropriate shock was observed without an increased risk of syncope and appropriate shock. A meta- analysis evaluated the impact of a prolonged arrhythmia detection duration on outcome [102]—thus excluding the EMPIRIC trial (which used 18 of 24 intervals for VF detection), the PREPARE trial (which used a historical control group), and the high-rate therapy arm of the MADIT-RIT. Analyzing the cohort of patients enrolled in RELEVANT, Arm C of MADIT-RIT, ADVANCE III, and PROVIDE, the meta-analysis reported a reduction of overall burden of therapies, driven by the greater than 50% reduction in appropriate and inappropriate ATP and the 50% reduction in inappropriate shocks. A reduction in all-cause mortality was observed without an increase in the risk of syncope. All the reports above clearly stress the necessity to

Table 2
Tachycardia detection evidence.

Study	Participants (N)	Short detection controls	Prolonged detection intervention	Findings
PREPARE	1391 Nonrandomized Primary prevention	12 of 16 (58%) 18 of 24 (42%)	30 of 40	Reduction in inappropriate shocks (SVT), avoidable shocks (VT), and "morbidity index"
RELEVANT	324 Nonrandomized Primary prevention	12 of 16	30 of 40	Reduction in inappropriate shocks (SVT), avoidable shocks (VT), and HF hospitalizations
MADIT-RIT	1500 Randomized Primary prevention	2.5 s (170–199 bpm) 1 s (≥ 200 bpm)	60 s (170–199 bpm) 12 s (200–249 bpm) 2.5 s (≥ 250 bpm)	Reduction in first inappropriate therapy, first appropriate therapy, appropriate ATP, and inappropriate ATP; improved survival
ADVANCE-III	1902 Randomized Primary & secondary prevention	18 of 24	30 of 40	Reduction in overall therapies, inappropriate shocks, and all-cause hospitalizations
PROVIDE	1670 Randomized Primary prevention	12 beats	25 beats (180–214 bpm) 18 beats (214–250 bpm) 12 beats (> 250 bpm)	Reduction in all-cause shock rate; improved survival

consider a long detection window setting as a "default" strategy for ICD programming. Moreover, they underline the importance of choosing to reprogram the ICD rather than using the manufacturers' out-of-the-box settings. A summary of the large comparative datasets of tachycardia detection is presented in Table 2.

11. Limitations of Data on the Duration of Tachycardia Required for Detection

Although the findings on the effect of tachycardia detection duration are based on roughly 7000 patients, there are limitations. Data on secondary prevention patients are limited to 25% of the 1902 patients enrolled in the ADVANCE III trial ($n=477$). Although this proportion is a fair representation of the real-world population receiving an ICD, more data are needed to fully understand the impact of a long-detection strategy in this subgroup of patients. MADIT-RIT and RELEVANT did not include single-chamber ICDs, and MADIT-RIT excluded patients with permanent AF. The PROVIDE and MADIT-RIT trials were designed to assess the time to first therapy and not the overall rate of therapies. MADIT-RIT, ADVANCE III, RELEVANT, and PROVIDE used devices from 3 different manufacturers with detection strategies leading to different detection times, intervals, and definitions. Some manufacturers of ICDs are not represented at all in these trials. Programming in the trial control groups was highly heterogeneous, with time until ATP or charging for VF as varied as about 11–12 intervals (approximately 3.4 seconds at 200 bpm) in MADIT-RIT and PROVIDE and 18 intervals (approximately 5.4 seconds) in ADVANCE III. An approximate translation of the impact of the number of intervals to detection and tachycardia cycle length (CL) are listed in Table 3. A further limitation is the relatively short duration and lack of inclusion of the patients with the most severe illness receiving an ICD. This limitation minimizes the exposure to relatively rare events that might occur in nonclinical trial, "real-world" patients. Lastly, as ICD batteries deplete, the charge time lengthens. The effect of such a delay to shock therapy in addition to prolonged detection times has not been studied.

12. Rate Criteria for the Detection of Ventricular Arrhythmia

Ventricular tachyarrhythmia detection by implantable devices is primarily based on heart rate. Heart rates can be extremely rapid

during ventricular tachyarrhythmias, and it is less likely that such rates are achieved during supraventricular tachyarrhythmias—thus making rate a powerful component of arrhythmia discrimination. However, VT can also present slower rates in the range of those of supraventricular tachyarrhythmias or even of sinus tachycardia. Therefore, any rate cutoff will always imply a tradeoff between maximizing sensitivity for ventricular tachyarrhythmia detection at the expense of inappropriate detection of fast supraventricular tachyarrhythmias and maximizing specificity at the expense of some slow VTs going undetected [103].

Because ICD therapy was initially employed in secondary prevention patients, the cutoff rate was usually tailored to a rate slightly below that of the observed VT. With the development of ICD use in primary prevention, the detection rate came into question because there is no history of sustained tachycardia in these patients. The recognition of a significant rate of inappropriate therapies in primary prevention studies, and their potentially deleterious consequences, prompted the development of studies that tested whether programming faster rate criteria reduced avoidable ICD therapies, particularly shocks. In many of these studies, however, testing also involved programming parameters other than rate, and those have been discussed as described below.

In the MADIT-RIT trial of primary prevention patients, conventional therapy (rate cutoff 170 bpm, $n=514$) was compared with a "high-rate group" in which rate cutoff was 200 bpm ($n=500$) [96]. The primary endpoint of first occurrence of inappropriate therapy was observed in 20% of the conventional group and in 4% of the high-rate group (P follow-up of 1.4 years. ICD shocks occurred in 4% and 2% of patients in the conventional and high-rate groups, respectively. The proportion of patients with appropriate therapies was also significantly different (22% vs 9% in the conventional and high-rate groups, respectively). It is important to note that all-cause mortality in the conventional group (6.6%) was approximately double that of the high-rate group (3.2%, $P=.01$).

In a single-center observational study, 365 primary prevention patients were prospectively studied, with programming including a single shock-only zone over 220 bpm [104]. During a mean follow-up of 42 months, 11% of the patients (7% in the first 2 years) experienced appropriate shocks, and only 6.6% experienced inappropriate shocks. It was notable that in the monitoring zone over 170 bpm, self-limiting VT episodes were detected in 12% of the

Table 3
Approximating the time taken to detect 30 intervals using fixed 8 of 10 interval detection plus adding a time delay, for a range of heart rates.

Arrhythmia characteristic		Interval-based detection	8 of 10 interval detection, then delay	
Beats per minute	Cycle length (ms)	Time to detect 30 intervals (s)	Time to detect 8 intervals (s)	Subsequent delay to approximate a 30-interval detection time
180	333	10.0	2.7	7.0
200	300	9.0	2.4	6.5
220	273	8.2	2.2	6.0
240	250	7.5	2.0	5.5
260	231	6.9	1.8	5.0
280	214	6.4	1.7	4.5
300	200	6.0	1.6	4.5

patients but were symptomatic in only 1.9%. The mortality rate was 17%, with one case of unexplained sudden death.

A recent primary prevention study revealed that there was considerable overlap between the ventricular rates of supraventricular and ventricular arrhythmias, and the majority of inappropriate shocks occurred at rates between 181 and 213 bpm [98]. These data also support the notion that for primary prevention patients it is safe to increase the rate cutoff up to 200 bpm to reduce these potentially avoidable therapies, a practice that was also supported by the results of the MADIT-RIT trial.

In secondary prevention patients, no trial has randomized the detection rate and compared outcomes. However, the ADVANCE III Secondary Prevention substudy confirmed the safety of not programming therapy for rates patient-years [105]. Previously published recommendations suggest a VT zone starting at 10 to 20 bpm slower than the observed tachycardia rate, usually including a 2- or 3-zone arrhythmia detection scheme (as discussed elsewhere) [106]. Clinicians should allow a larger rate differential when starting a patient on an antiarrhythmic drug that might slow the clinical tachycardia rate (e.g., amiodarone).

13. Single- or Multi-Zone Detection

Modern ICDs allow the rate to be classified into single or multiple zones. This classification permits different criteria to be applied for detection (e.g., number of intervals) and for tiered therapy (e.g., different adaptive CLs for slower vs faster VTs and more sequences of ATP for slower and presumably hemodynamically more stable VTs). Additionally, because some manufacturers tie SVT discrimination algorithms to specific VT zones, programming more than one tachycardia zone allows for greater specificity in discriminating VT from SVT (see online Appendix B). Although there are trials in which arms differ in whether a single zone or multiple zones are used, this is typically performed to allow programming of various detection, discrimination, or therapies for comparison. Thus, the number of zones was not the randomization variable being directly compared. Therefore, the concept of single- vs multizone programming as a head-to-head comparison is not well tested. The MADIT-RIT study randomized primary prevention ICD patients into 1 of 3 arms with single-, dual-, or triple-zone programming (the single-zone arm also had a monitoring zone). Although the trial's aim was to compare conventional therapy with high-rate and delayed therapy, the outcome for the single-zone arm (high-rate) was comparable to the triple-zone (delayed) arm and superior to the dual-zone (conventional) arm, with regard to inappropriate shock [96]. This study is consistent with multiple studies in ICD programming in which the use of multiple-zone programming has allowed for flexibility in programming strategies with regard to detection, discrimination, and therapy. Additionally, there are observational data from the ALTITUDE Real World Evaluation of Dual-Zone ICD and CRT-D Programming Compared to Single-Zone Programming (ALTITUDE REDUCES) study that show that dual-zone programming is

associated with fewer shocks than single-zone programming, at least for rates [64]. Therefore, the authors conclude that using more than 1 detection zone can be useful for modern ICD programming. It should be noted that ATP before or during charging was used in the majority of studies described in both the tachycardia detection and therapy sections and thus is recommended for longer detection.

14. Discrimination Between Supraventricular and Ventricular Arrhythmia

The SVT-VT discrimination process classifies a sequence of sensed EGMs that satisfies rate and duration criteria as either SVT (therapy withheld) or VT/VF (therapy given). *Discriminators* are individual algorithm components that provide a partial rhythm classification or a definitive classification for a subset of rhythms. *Discrimination algorithms* combine individual component discriminators to produce a final rhythm classification. Discrimination algorithms vary among manufacturers and between individual ICD models (see online Appendix B). The final rhythm classification can differ depending on the technical details of how each individual discriminator is calculated, the nominal or programmed threshold for each discriminator, the order in which discriminator components are applied, and the logical connections between them (e.g., “and” vs “or”). In some ICDs, rhythms classified as VT/VF undergo a subsequent sensing-verification step to confirm that EGMs represent true cardiac activation.

15. SVT-VT Discriminator Components

Individual discriminators can be considered in relation to the EGMs analyzed as ventricular-only or both atrial and ventricular, by the rhythm that they identify (e.g., AF, sinus tachycardia, VT), or by the type of EGM information analyzed (intervals vs morphology). Note that ventricular rate alone is a mandatory discriminator, as discussed in the section above. We summarize the most commonly used discriminators. More comprehensive discussions are available in the literature [107–111].

16. Rejection of Sinus Tachycardia by Onset

Several interval-based discriminators focus on differences in the onset of sinus tachycardia (gradual and parallel acceleration of atrial and conducted ventricular intervals) compared with VT (typically abrupt, with at least transient atrioventricular dissociation). *Sudden (abrupt) onset* was one of the first single-chamber, interval-based discriminators. It withholds therapy if acceleration across the sinus-VT rate boundary is gradual. Because onset discriminators classify the rhythm only once, and thus cannot correct misclassifications, they are now used infrequently and only with an override feature and/or other discriminators [112–115].

Chamber of onset is a related, interval-based, dual-chamber discriminator that classifies a 1:1 tachycardia as SVT if the atrial rhythm accelerates at the device-defined onset. A related “Sinus Tachycardia^{fi}” discriminator classifies a tachycardia as VT if either the RR or the PR intervals deviate sufficiently from the range of the immediately preceding sinus intervals [116].

16.1. Rejection of AF by Ventricular Interval Regularity

Ventricular interval regularity (*interval stability*) is an explicit single-chamber, interval-based discriminator that classifies the rhythm as AF if the ventricular intervals are sufficiently irregular. Because interval variability in conducted AF decreases at faster rates, stability becomes unreliable in discriminating VT from conducted AF at ventricular rates greater than 170 bpm [112,115]. Interval stability can also fail if drugs (e.g., amiodarone) cause monomorphic VT to become irregular or induce polymorphic VT to slow into the SVT-VT discrimination zone [114,117].

16.2. Diagnosis of VT by Dual-Chamber Components: Atrial vs Ventricular Rate and Atrioventricular Association

In contrast to the single-chamber discrimination algorithms above that diagnose SVT when their criteria are fulfilled, 2 separate, interval-based, dual-chamber discrimination algorithms diagnose VT. First, **atrial rate vs ventricular rate** diagnoses VT if the ventricular rate exceeds the atrial rate [118]. s, **atrioventricular dissociation** identifies isorhythmic VT during sinus tachycardia. Inversely, the atrioventricular **association** discriminator diagnoses SVT in the presence of N:1 (e.g., 2:1, 4:1) atrioventricular association consistent with atrial flutter at a fixed conduction ratio.

16.3. The Ventricular Electrogram Morphology Discriminator

This versatile, single-chamber discriminator is the only algorithm component that does not rely on inter-EGM intervals. It classifies tachycardias as SVT if the morphology (shape) of the ventricular EGM is sufficiently similar to the morphology during a conducted baseline rhythm. It can potentially discriminate any SVT from VT, including SVTs that challenge other discriminators, such as abrupt-onset 1:1 SVTs and irregular VT during AF. Contemporary ICDs (including subcutaneous ICD [S-ICD]) analyze EGMs from the shock electrodes, which record a larger field of view than EGMs from pace-sense electrodes [119]. They operate using a common series of steps and are susceptible to common failure modes [111,120–123]. The first common step is acquisition of a baseline rhythm template by mathematically extracting EGM features and storing them. Both the acquisition of the initial template and the subsequent template updating are automated in most ICDs. Nevertheless, physicians should confirm that the conducted baseline beats match the template both at implant and during follow-up. For CRT patients, the template must be manually collected. If the wavelet signal during template acquisition appears clipped, adjustments specific to the manufacturer might be necessary.

17. SVT-VT Discrimination Algorithms

Discrimination algorithms combine component discriminators to provide a final rhythm classification of VT/VF or SVT. The morphology discriminator frequently forms the primary component of single-chamber algorithms with stability playing a secondary role and sudden onset used sparingly. By contrast, the cornerstone of most dual-chamber algorithms is explicit or implicit comparison of atrial vs ventricular rates. Because the ventricular rate is greater than the atrial rate in more than 80% of VTs, algorithms that

compare atrial and ventricular rates as their first step apply additional SVT discriminators to fewer than 20% of VTs, reducing the risk that they will misclassify VT as SVT [124,125]. Most dual-chamber algorithms further restrict single-chamber discriminators to tachycardias for which they offer the greatest benefit; thus, stability is applied only if AF is confirmed by direct calculation of the atrial rate or the atrial rate is greater than the ventricular rate. Similarly, sudden onset, chamber of onset, or 1:1 atrioventricular association are applied only if the atrial rate equals the ventricular rate. The use of discriminators in redetection varies among manufacturers and has not been systematically studied.

18. Assessing Clinical Benefits and Risks

18.1. What Evidence Supports a Benefit?

1. The annual rate of inappropriate shocks has fallen dramatically from 37%–50% for SVT alone in early studies to 1%–5% for all causes in modern clinical trials [97,118,126–128]. This decrease is likely due to differences in both clinical populations and the programming of multiple ICD parameters, including longer detection time and higher rate cutoffs. Thus, it is difficult to isolate the differential effect of SVT-VT discrimination algorithms using clinical data. These studies have programmed discrimination algorithms to ON, however, so it seems reasonable to use them.
2. Although clinical trials that reported dramatic reductions in shocks for SVT programmed discrimination algorithms consistently, they have been programmed inconsistently in clinical practice, and the rate of inappropriate shocks for SVT has been higher in observational studies of remote-monitoring ICD databases. In the ALTITUDE REDUCES study on 15,991 patients in the Latitude^{fi} database, SVT was the most common cause of shocks when the detection rate was ≤ 180 bpm [129]. For detection rates ≤ 170 bpm, the rate of inappropriate shocks at 1 year was significantly lower with dual-zone programming, which permits SVT-VT discrimination, than single-zone programming, which does not (9.6% vs 4.3%). Similarly, Fischer et al [130] analyzed shocks in 106,513 patients in the CareLink^{fi} database; programming SVT-VT discrimination ON was associated with a 17% reduction in all-cause shocks.
3. Sophisticated simulations indicate that SVT-VT discrimination algorithms have substantial benefit. For example, the SCD-HeFT study on primary prevention patients did not use discriminators. A validated Monte Carlo simulation predicted that use of single- or dual-chamber SVT-VT discriminators alone would have reduced inappropriate shocks for SVT by 75.5% and 78.8%, respectively [131].

18.2. Which Patients are Most Likely to Benefit, and Which are Least Likely to Benefit?

Despite limited direct evidence, it seems clear that patients will benefit most if the rates of their VTs and SVTs overlap. This includes patients with slower monomorphic VT, those at risk for AF with rapid ventricular rates, or those capable of exercising to sinus rates in the VT zone^v [103,132]. In secondary prevention patients with slower VT, older discrimination algorithms reduced shocks for SVT compared with rate-only detection. The benefit is less for primary prevention patients, secondary prevention patients at risk only for VF, and those who cannot sustain rapid atrioventricular conduction. Patients with permanent complete atrioventricular block do not benefit.

18.3. What are the Risks?

The risk of the misclassification of either VT or VF as SVT by the discrimination algorithms can either prevent VT detection or delay the time to therapy (underdetection), as documented in clinically significant situations [112,113,115,125]. When modern algorithms are programmed to recommended parameters, clinically significant underdetection is rare. Large clinical trials on multiple shock-reduction strategies (including SVT-VT discrimination) report no or minimal and statistically insignificant increases in syncope [95,97,126,133]. Most reports do not include the causes of syncope and thus do not permit identification of whether discrimination algorithms contributed to any of the syncopal episodes by prolonging detection. However, in the PREPARE study, no syncopal episode was caused by untreated tachycardia [133]. In general, discriminators that re-evaluate the rhythm classification during ongoing tachycardia reduce the risk of underdetection compared with those that withhold therapy if the rhythm is misclassified by the initial evaluation (e.g., onset, chamber of origin algorithms).

19. Additional Considerations

19.1. SVT Limit

SVT-VT discrimination applies from the VT detection rate to the SVT limit rate, which is programmable independently of the VT/VF therapy zones with some manufacturers (preferable), but which might be linked to one of the zone boundaries in others. The minimum CL for SVT-VT discrimination should be set to prevent clinically significant delays in the detection of hemodynamically unstable VT. PREPARE, EMPIRIC, and MADIT-RIT all support the safety of empirical programming at 200 bpm [96,101,134]. In MADIT-II, approximately 50% of SVT episodes were faster than 170 bpm, and a few were as fast as 250 bpm [82]. In INTRINSIC RV, SVT comprised 19% of episodes, with rates between 200 and 250 bpm [135]. More limited and preliminary data from PainFree SST support programming up to 222–230 bpm [116,136]. We suggest the SVT limit not exceed 230 bpm in adults without a patient-specific indication, based on the low incidence of SVTs in this rate range among ICD patients and the potential—however small—for misclassifying hemodynamically unstable VT.

19.2. Duration-Based “Safety-Net” Features to Override Discriminators

These features deliver VT/VF therapy if a tachycardia satisfies the ventricular rate criterion for a sufficient duration, even if the discrimination algorithm indicates SVT. The premise is that the ventricular rate during transient sinus tachycardia or AF will decrease to below the VT rate boundary before the override duration is exceeded. In one study, an override duration of 3 minutes delivered inappropriate therapy to 10% of SVTs [112]. Because SVT is much more common than VT, programming an override duration of less than 5–10 minutes results primarily or solely in inappropriate SVT therapy [122]. Although more data would be useful, in the absence of a documented benefit, we recommend programming this feature OFF or long (minutes) without a patient-specific or device-specific indication.

19.3. Dual-Chamber vs Single-Chamber Algorithms

Clinical trials and simulated testing of induced arrhythmias that compared single- vs dual-chamber discriminators have reported inconsistent results [10,33,137–139]. Two meta-analyses found no

superiority of dual-chamber ICDs in terms of mortality or inappropriate therapies [11,140]. Any benefit of dual-chamber discrimination is likely restricted to specific patient groups [103,138]. For example, the Dual Chamber and Atrial Tachyarrhythmias Adverse Events (DATAS) trial of predominantly secondary prevention patients with slower VTs reported modest benefit from dual-chamber discrimination, while the recent Reduction and Prevention of Tachyarrhythmias and Shocks Using Reduced Ventricular Pacing with Atrial Algorithms (RAPTURE) trial of primary prevention patients programmed to a fast detection rate (> 182 bpm) and long detection duration (30/40 intervals) did not [103,138,139]. Inappropriate therapy for SVT occurred in only 2% of the patients in each group. Recent data from PainFree SST notes very low rates of inappropriate shocks (3.7% for single chamber; 2.8% for dual and triple chamber after 2 years). The choice of device was not randomized, suggesting that when physicians chose a dual- or triple- chamber device (perhaps due to known atrial arrhythmia or bradycardia), inappropriate shock rates were minimized [136]. The Optimal Anti-Tachycardia Therapy in Implantable Cardioverter-Defibrillator Patients Without Pacing Indications (OPTION) trial randomized 462 patients to single- or dual-chamber programming and noted inappropriate shock rates of 10.3% for single chamber vs 4.3% for dual chamber after 27 months ($P=.015$). Atrial lead-related complications were 1.3%, therapy was delivered from 170 bpm (VT) and 200 bpm (VF), and no difference in ventricular pacing percentage was noted [141]. Dual-chamber algorithms probably reduce the risk of underdetection compared with single- chamber algorithms because more than 80% of VTs with a ventricular rate greater than the atrial rate undergo no further analysis [103,124,125]. However, the rate of clinically significant underdetection with modern programming is so low that this difference is rarely of clinical significance. In most patients, improved SVT-VT discrimination should not be considered an indication for a dual- vs single-chamber ICD. Even if a dual-chamber ICD is implanted, dual-chamber discrimination should be programmed only if the atrial lead becomes chronic or if atrial sensing is unreliable. Accurate sensing of atrial EGMs is essential for dual-chamber SVT-VT discrimination. Atrial lead dislodgments, oversensing of far-field R waves, or undersensing due to low-amplitude atrial signals can cause misclassification of VT/SVT. On implant, it is important to position the atrial lead to minimize far-field R waves.

19.4. Ventricular Oversensing

Excluding recalled leads, ventricular oversensing accounts for less than 10% of inappropriate shocks, but it often results in repetitive shocks and severe symptoms [82,142–144]. Recently introduced features reduce inappropriate therapies from oversensing of physiologic T waves and nonphysiologic signals related to pace-sense lead failures as discussed below.

20. Programming to Reduce T-Wave Oversensing

The problem of T-wave oversensing relates to the basic requirement that ICDs reliably sense VF, which is characterized by RR intervals shorter than the normal QT interval and some EGMs with low amplitudes and slew rates. Approaches to minimizing T-wave oversensing include reprogramming ventricular sensitivity, altering sensing bandwidth, and changing the sensing bipole [109,123,145]. One manufacturer provides an algorithm that withholds therapy after rate and duration criteria for VT/VF are fulfilled if a specific pattern of T-wave oversensing is identified [146]. T-wave oversensing rates vary based on device design; using an appropriate high band-pass filter results in very low rates

of T-wave oversensing [142]. Because T-wave oversensing is unpredictable, features that minimize T-wave oversensing should be enabled proactively at implant, providing they do not cause undersensing in VF [146].

21. Lead-Related Oversensing

Oversensed signals caused by pace-sense lead failure have specific interval patterns and EGM characteristics [145,147,148]. Present algorithms identify three features: (1) intervals can be too short to represent successive ventricular activations; (2) such short intervals are often transient and can be repetitive; and (3) in true bipolar leads, oversensed signals are absent on the shock EGM. Algorithms can provide warning alerts, withhold shocks after spurious detection of VT/VF, or both. All 3 criteria can provide alerts, but only the third is applied to withhold shocks. The present algorithms were developed to identify impending lead failures on recalled leads, notably the Sprint Fidelis. These algorithms might not be appropriate for detecting failures in other leads [144]. There is a high false-positive rate when using these algorithms, and caregivers must carefully review the device data that caused the alert to ensure the lead experienced a true failure [145].

Alerts that combine both oversensing and abrupt changes in impedance trends provide earlier warning of lead failure than a fixed impedance threshold [144,145,149]. Such alerts can be delivered via wireless remote monitoring and/or by notifying the patient via vibration or an audible tone. Caregivers must respond rapidly to alerts to minimize inappropriate shocks [144,149]. Wireless remote monitoring has been reported to reduce response time [150]. The principal disadvantage of lead alerts is false-positive triggers. The principal risk of shock-withholding algorithms is a failure to shock VF, which is extremely rare [151]. In addition to algorithmic approaches, oversensing due to failure of the cable leading to the ring electrode can be prevented by changing the programming of the sensing configuration from true bipolar to integrated bipolar. This approach is appropriate prophylactically or as temporary programming after a ring electrode cable failure; it is not a permanent solution, however, because increased rates of high-voltage cable fractures have been documented after sensing cable fractures [152].

22. The Subcutaneous Defibrillator (S-ICD)

The novel S-ICD follows many of the same principles as intravascular ICDs but is considered here separately for duration criteria, rate criteria, and discrimination algorithms. Candidates for the S-ICD must initially be screened with a modified tri-channel surface electrocardiogram that mimics the sensing vectors of the S-ICD system. This test is designed to assess the R-wave to T-wave ratio for appropriate signal characteristics and relationships. If the screening is not satisfactory for at least 1 of the 3 vectors supine and standing, an S-ICD should not be implanted. On implant, the S-ICD automatically analyzes and selects the optimal sensing vector.

Detection of VT or VF by the S-ICD is programmable using a single or dual zone. In the single-zone configuration, shocks are delivered for detected heart rates above the programmed rate threshold: the “shock zone” [134]. In the dual-zone configuration, arrhythmia discrimination algorithms are active from the lower rate: the “conditional shock zone.” In this latter zone, a unique discrimination algorithm is used to classify rhythms as either shockable or nonshockable. If they are classified as supraventricular arrhythmias or nonarrhythmic oversensing, therapy is withheld.

With dual-zone programming, the shock zone uses rate as the sole method for rhythm analysis. In contrast, the conditional shock zone uses a stepwise discrimination algorithm to distinguish shockable from nonshockable rhythms. The conditional shock zone has a morphology analysis process based on a normal rhythm transthoracic QRS:T wave template. The template uses up to 41 fiducial points to reconstruct morphology for the template as well as the programmed targeted heart rate zones. The comparison of the template to the high-rate rhythm electrocardiogram for discrimination constitutes the static waveform analysis. A good template match designates a sensed beat as supraventricular, thereby preventing a shock. A poor match to the static QRS:T morphology template moves the algorithm to a dynamic waveform analysis that compares single-beat morphologies in groups of 4 beats for uniformity. A consistent dynamic waveform match adjusts the sensing to evaluate QRS width. If a tachycardia has a prolonged QRS width compared with the template width (> 20 ms) and is of sufficient duration, it will lead to a shock.

The system uses an initial 18 of 24 duration criteria (non-programmable) prior to initiating capacitor charging; however, this duration is automatically extended following nonsustained ventricular tachyarrhythmia events. A confirmation algorithm is also used at the end of capacitor charging to ensure persistence of the ventricular arrhythmia prior to shock delivery. Shocks for spontaneous (noninduced) episodes are delivered at a non-programmable 80 J regardless of the therapy zone of origination.

When programmed to include a conditional shock zone, the S-ICD VT detection algorithm has been demonstrated to be more effective than transvenous ICD systems programmed at nominal settings to prevent the detection of induced supraventricular arrhythmias [153]. Furthermore, in the clinical evaluation of the conditional shock zone, the S-ICD system was strongly associated with a reduction in inappropriate shocks from supraventricular arrhythmias and did not result in prolongation of detection times or increased syncope [154].

23. Integrating Tachycardia Detection Data Into Programming Recommendations

When taking data from specific single-manufacturer studies and producing generic guidelines applicable across all ICDs, some compromises and potential pitfalls have been encountered. Nevertheless, it is our intention to convey the general principles of good quality evidence (e.g., extending detection time) to apply to ICD programming in general. Thus, attempts have been made to translate interval-based detection to time-based detection and to provide a range of reasonable heart rate cutoffs that are inclusive of those proven in good-quality trials. We encourage programming ICDs to manufacturer-specific therapies of proven benefit; however, when evidence is lacking, the guidelines provide a framework for programming within the evidence base. See online Appendix B for manufacturer-specific examples of optimal ICD programming.

Tachycardia Detection Programming Recommendations	Class of Recommendation	Level of Evidence
For primary prevention ICD patients, tachyarrhythmia detection duration criteria should be programmed to require the tachycardia to continue for at least 6–12 seconds* or for 30 intervals before	I	A

completing detection, to reduce total therapies.				It is recommended to activate lead-failure alerts to detect potential lead problems.		
<i>*Tachyarrhythmia detection duration is directly related to the tachyarrhythmia rate. Direct evidence to support a delay > 2.5 seconds for rates over 250 bpm is not available, but can be inferred from evidence that 30 detection intervals are safe at that rate.</i>				For secondary prevention ICD patients for whom the clinical VT rate is known, it is reasonable to program the slowest tachycardia therapy zone at least 10 bpm below the documented tachycardia rate but not faster than 200 bpm*, to reduce total therapies.	Ila	C-EO
For primary prevention ICD patients, the slowest tachycardia therapy zone limit should be programmed between 185 and 200 bpm*, to reduce total therapies.	I	A		<i>*Higher minimum rates for detection might be appropriate for young patients or for those in whom SVT-VT discriminators cannot reliably distinguish SVT from VT, provided there is no clinical VT below this rate.</i>		
<i>*Higher minimum rates for detection might be appropriate for young patients or for those in whom SVT-VT discriminators cannot reliably distinguish SVT from VT, provided there is no clinical VT below this rate.</i>				It can be useful to program more than one tachycardia detection zone to allow effective use of tiered therapy and/or SVT-VT discriminators and allow for a shorter delay in time-based detection programming for faster arrhythmias.	Ila	B-R
For secondary prevention ICD patients, tachyarrhythmia detection duration criteria should be programmed to require the tachycardia to continue for at least 6–12 seconds* or for 30 intervals before completing detection, to reduce total therapies.	I	B-R		When a morphology discriminator is activated, it is reasonable to reacquire the morphology template when the morphology match is unsatisfactory, to improve the accuracy of the morphology discriminator.	Ila	C-LD
<i>*Tachyarrhythmia detection duration is directly related to the tachyarrhythmia rate. Direct evidence to support a delay > 2.5 seconds for rates over 250 bpm is not available, but can be inferred from evidence that 30 detection intervals are safe at that rate.</i>				It is reasonable to choose single-chamber ICD therapy in preference to dual-chamber ICD therapy if the sole reason for the atrial lead is SVT discrimination, unless a known SVT exists that may enter the VT treatment zone, to reduce both lead-related complications and the cost of ICD therapy.	Ila	B-NR
Discrimination algorithms to distinguish SVT from VT should be programmed to include rhythms with rates faster than 200 bpm and potentially up to 230 bpm (unless contraindicated*) to reduce inappropriate therapies.	I	B-R		For the S-ICD, it is reasonable to program 2 tachycardia detection zones: 1 zone with tachycardia discrimination algorithms from a rate \leq 200 bpm and a second zone without tachycardia discrimination algorithms from a rate \geq 230 bpm, to reduce avoidable shocks.	Ila	B-NR
<i>*Discrimination algorithms and/or their individual components are contraindicated in patients with complete heart block or if the algorithm/component is known to be unreliable in an individual patient. Dual-chamber discriminators that misclassify VT as SVT if the atrial lead dislodges are discouraged in the perioperative period. Dual-chamber discriminators are contraindicated in patients with known atrial lead dislodgment, atrial undersensing or oversensing of far field R waves, and in those with permanent AF.</i>				Programming a nontherapy zone for tachycardia monitoring might be considered to alert clinicians to untreated arrhythmias.	Iib	B-NR
				It may be reasonable to disable the SVT discriminator timeout function, to reduce inappropriate therapies.	Iib	C-EO
				It may be reasonable to activate lead "noise" algorithms that withhold shocks when detected	Iib	C-EO
	I	B-NR				

VT/VF is not confirmed on a shock or other far-field channel to avoid therapies for non-physiologic signals.

It may be reasonable to activate T- IIB C-LD
wave oversensing algorithms, to
reduce inappropriate therapies.

It may be reasonable to program IIB C-EO
the sensing vector from bipolar
to integrated-bipolar in true-
bipolar leads at risk for failure of
the cable to the ring electrode to
reduce inappropriate therapies.*

**This is not intended as a long-term
solution when a cable fracture
has been identified.*

24. Tachycardia Therapy Programming

Although therapies delivered by the ICD can abort SCD, appropriate and inappropriate ICD shocks have been associated with a considerable increase in the risk of mortality [82,83,155–158]. In the Sudden Cardiac Death in Heart Failure Trial (SCD-HeFT), the risk of mortality was 5-fold higher in patients who received appropriate ICD shocks and 2-fold higher in patients who received inappropriate shocks [83]. Similarly, pooling data from 4 studies of 2135 ICD patients, shocked VT was associated with a 32% increase in the risk of mortality. In that analysis, shocked patients had poorer survival than patients treated with ATP only [155]. ICD shocks are likely a marker of more advanced heart disease and subsequent death, but defibrillation therapies have been associated with troponin release and increased LV dysfunction with the potential of further mortality risk.

The incidence of appropriate and inappropriate ICD shocks depends on the patient's characteristics, including the indication for the device, concomitant medical therapies including antiarrhythmic medications, programming of the ICD, and the duration of follow-up. With regard to ICD programming, faster VT/VF detection rates, longer detection durations, use of a single zone, use of SVT discriminators, and delivery of ATP have been shown to reduce both appropriate and inappropriate shocks and to improve quality of life [91,101,126,129,130,133,159,160]. This programming might improve survival [126]. Indeed, several studies have shown that ATP is effective at terminating slow and fast VT with exceedingly low rates of adverse events like syncope [93,135,161–165]. The initial bias of the ICD community was to reserve ATP therapy for those patients in whom the therapy was demonstrated to be effective, usually during an electrophysiologic study. However, the approach of physician-directed programming based on the knowledge of induced arrhythmias was found to be significantly inferior to the routine strategic (EMPIRIC) programming of ATP. It is not reflective of the arrhythmias experienced outside the electrophysiology laboratory for primary and secondary prevention patients with ischemic and nonischemic substrates [101,166]. Although the ideal number of ATP bursts has not been definitively determined, current data support the use of up to 2 ATP attempts, given additional attempts yield very little additional efficacy [93,135,161–165,167,168]. In one study, up to 5 attempts were found to be safe [168]. The most effective ATP duration is likewise uncertain; however, in the ATP Delivery for Painless ICD Therapy (ADVANCE-D) trial—a prospective RCT of 925 patients—8-pulse ATP was as effective and safe as 15-pulse ATP [169]. The PITAGORA ICD clinical trial randomized 206 patients

with an ICD to 2 ATP strategies: an 88% coupling interval burst vs a 91% coupling interval ramp. The results of the trial showed that over a median follow-up of 36 months and compared with ramp pacing, burst pacing was more effective for terminating fast VT episodes (between CL 240 and 320 ms) [170]. In a prospective study of 602 patients, a strategy of tiered ATP and low-energy shock was efficacious and safe in patients with VT CL greater than 250 ms, with extremely low syncope rates [171]. However, a “real-world” retrospective study on 2000 patients with 5279 shock episodes from the LATITUDE remote monitoring system showed that the success rate of first shock as first therapy was approximately 90%, but the success rate was lower after failed ATP. Therefore, that study recommended programming a higher level of energy after ATP [172]. Finally, a substudy of the Effectiveness and Cost of ICD Follow-Up Schedule with Telecardiology (ECOST) study, which randomly assigned 433 patients to remote monitoring (n=221; active group) vs ambulatory follow-up (n=212; control group), showed that remote monitoring was highly effective in the long-term prevention of inappropriate ICD shocks through early detection and prevention of AF with a rapid ventricular rate, nonsustained VT, or diverted VT episodes [173].

25. Benefits and Risks

The goal of ICD therapy is to prolong life while causing as little morbidity as possible. Although survival is quantifiably objective, morbidity is more subjective and includes both physical and emotional components. Clearly, shocks are usually painful to the patient, whereas ATP is typically not uncomfortable. However, there can be other morbidities related to both therapies, including mild to extreme emotional distress, syncope, palpitations, and proarrhythmia yielding more therapies and occasionally leading to death. Paradoxically, the need for life-saving therapies, including shocks and potentially ATP, might also be associated with increased mortality; however, the causal relationships are unclear. Also, the prevalence of tachycardia amenable to ATP or hemodynamic significance varies with the mechanism of the risk (e.g., long QT vs ischemic cardiomyopathy). In addition, although the risk of having a hemodynamically important or life-threatening arrhythmia can vary from patient group to patient group, the largest proportion of patients in whom ICD therapy is applied has yet to have a previously recorded arrhythmia, and we must therefore strategically choose on the basis of other factors how we will treat the first event and subsequent events.

26. Classification of Therapy

The literature uses definitions of therapies that differ from each other and that impact their results and conclusions. The occurrence rates of these events not only are dependent on their definition but also are highly dependent on the programming of the defibrillation system. Both shock and nonshock therapies can be categorized as being appropriate, inappropriate, and avoidable. Whereas appropriate and inappropriate therapies refer to therapies that were actually delivered, avoidable therapies are theoretical events in the future. These potential future tachycardia therapies, delivered for either appropriately or inappropriately detected events, can frequently be avoided by establishing programming to either prevent the initiation of the arrhythmia or to allow the condition to pass without therapy.

27. Appropriate

A response to a sustained ventricular arrhythmia (VT, VF) or hemodynamically poorly tolerated arrhythmias (e.g., associated with syncope, rate over 200 bpm, or hemodynamically compromising supraventricular arrhythmias).

28. Inappropriate

A response to signals generated by something other than sustained ventricular arrhythmias or hemodynamically poorly tolerated arrhythmias. Possible signals include supraventricular rhythms such as sinus tachycardia, AF, atrial flutter, reentrant SVT, atrial tachycardia, or instances of signal misinterpretation. Signal misinterpretation includes multiple counting of single events (e.g., atrial, T-wave or R-wave), environmental signals such as electromagnetic interference, frequent premature ventricular contractions (PVCs) and nonsustained ventricular arrhythmias, extracardiac physiologic signals (e.g., diaphragmatic or pectoral myopotentials), other implantable electronic devices (e.g., pacemakers, LV assist devices, nerve stimulators), inappropriate lead placement or dislodgment, conductor or insulation failures, header connection instability, and pulse generator failure.

29. Avoidable

Programming of detection and therapy parameters and algorithms so that shock or ATP therapy is withheld from arrhythmias that would be expected to be hemodynamically tolerated. Examples include self-terminating ventricular arrhythmias, ATP-susceptible ventricular arrhythmias, and overdrive suppression responsive rhythms. Many appropriate and most inappropriate therapies are also potentially avoidable.

30. Phantom

These are not true therapies; however, there is the patient's perception that a therapy was delivered. Interrogation of the ICD and/or coincident rhythm monitoring does not identify a tachycardia or therapy.

31. Unintended Consequences of ICD Therapy and ICD Therapy Programming

In the SCD-HeFT and MADIT II trials, inappropriate shocks more than doubled the risk of death. Mortality rates were substantially higher after shocks: 10% within days after the first shock, 25% within 1 year, and 40% by 2 years. The leading cause of death was progressive HF. In an analysis of the MADIT-CRT trial, the patients with appropriate shocks experienced increased mortality when compared with the patients without ICD shocks, after accounting for mechanical remodeling effects; this was not the case for patients who received appropriate ATP only [156]. ICD shocks have also been associated with independent predictors of mortality in the large ALTITUDE registry of 3809 ICD recipients and in a meta-analysis of ICD trials in which ATP was applied [155,157]. Emotional morbidities associated with ICD shocks are well recognized and include anxiety, depression, and posttraumatic stress disorders [174–176]. Phantom shocks can result from fear and/or

anxiety and have a reported incidence of 5% in a European study of ICD recipients over 35 months of follow-up [177]. If possible, and when safe, it is best to avoid both the discomfort and psychological impact of shocks for ventricular arrhythmias, supraventricular arrhythmias, noise events including lead failures, and for self-terminating arrhythmias, as is discussed in the section on tachycardia detection. The 1500-patient MADIT-RIT study demonstrated a mortality reduction by changing both tachycardia detection criteria and tachycardia therapy (shocks and ATP). Therefore, it is difficult to assign the outcome result to ATP, shocks, or both when compared with older, more conventional programming [126]. In addition, in a randomized study of remote follow-up of ICDs, home monitoring showed an incidence of 52% fewer inappropriate shocks, 72% fewer hospitalizations due to inappropriate shocks, 76% fewer capacitor charges, and a significant positive impact on battery longevity [178].

31.1. ATP

Several large clinical trials have established the safety and efficacy of ATP as a first-line therapy to treat even very fast VTs [93,95,101,133]. The use of first-line ATP involving VT at rates between 188 and 250 bpm in the PainFREE Rx II trial resulted in a 71% relative shock reduction [93]. In the PREPARE study, a primary prevention cohort of 700 patients was programmed with 30 of 40 detection intervals with ATP- first for VT between 182 and 250 bpm with SVT discriminators active up to 200 bpm. The results demonstrated a robust absolute risk reduction for shocks at 1 year from 17% to 9% without an increase in arrhythmic syncope when compared with historical controls [133]. Similar findings were noted in the RELEVANT study, which evaluated a cohort of patients with nonischemic heart disease and cardiac resynchronization defibrillators [95]. In the earlier EMPIRIC study, standardized VT detection and ATP therapy parameters demonstrated a reduction in shocks when compared with physician-tailored treatment in a randomized assessment of 900 primary prevention patients [101]. The use of ATP during ICD capacitor charging has been clinically validated as safe and effective [163]. It is important to recognize that inappropriate therapies including inappropriate ATP, delivered primarily in the setting of supraventricular arrhythmias, have been associated with increased mortality in the MADIT-RIT and MADIT-CRT trials [156,179]. However, the overall safety of ATP and its role as a contributor to improved survival are well established, particularly in terms of preventing avoidable ICD shocks.

31.2. Customized vs Strategic Programming

Because primary prevention patients have no prior ventricular arrhythmias, programming individual devices on implant is largely empiric. There are more data for secondary prevention patients, but how the patient will behave in the future is still uncertain. The ability to individualize the antitachycardia programming for patients with both primary and secondary prevention indications was tested in the EMPIRIC trial and found to be an inferior approach to prevent these therapy events [101]. The application of standardized programming and borrowing data from the PainFREE Rx II and PREPARE studies resulted in a comprehensive review of programming and its application across manufacturers.

31.3. Secondary Prevention

For the secondary prevention ICD patient, specific knowledge of the patient’s arrhythmia history facilitates the creation of an effective antitachycardia programming strategy. Using what is known about the ventricular arrhythmia, including any electrocardiograms, available telemetry strips, and EMS recordings, provides insight into the arrhythmia mechanism. In cases of monomorphic VT, discerning the rate (CL) and the hemodynamic impact is useful in making choices, particularly for detection at a minimum; the device must be programmed with active VT detection zones sufficient to cover the clinical arrhythmia. Slower, monomorphic VT that is better tolerated hemodynamically favors a robust approach using ATP termination with at least 2–3 sequences and at least 8 pulses. The use of a second burst of ATP has also been shown to increase effectiveness from 64% to 83% in the fast VT range of 188 to 250 bpm [167]. Although a second burst has clear value, value beyond 2 bursts is limited, except in rare situations [101]. The use of ICDs in patients with implanted LV assist devices allows prolongation of detection times and programming of multiple ATP attempts without significant risk to the patient, and it reduces the opportunity for shock therapies. Adjunct medications and ablation of VT (or SVT) might also be considered for cases in which slow VT occurs or if there is an overlap between the SVT and VT rates, leading to ICD therapies.

Tachycardia Therapy Programming Recommendations	Class of Recommendation	Level of Evidence
It is recommended in all patients with structural heart disease and ATP-capable ICD therapy devices that ATP therapy be active for all ventricular tachyarrhythmia detection zones to include arrhythmias up to 230 bpm, to reduce total shocks except when ATP is documented to be ineffective or proarrhythmic.	I	A
It is recommended in all patients with structural heart disease and ATP-capable ICD therapy devices that ATP therapy be programmed to deliver at least 1 ATP attempt with a minimum of 8 stimuli and a cycle length of 84%–88% of the tachycardia cycle length for ventricular tachyarrhythmias to reduce total shocks, except when ATP is documented to be ineffective or proarrhythmic.	I	A
It is indicated to program burst ATP therapy in preference to ramp ATP therapy, to improve the termination rate of treated ventricular tachyarrhythmias.	I	B-R
It is reasonable to activate shock therapy to be available in all* ventricular tachyarrhythmia therapy zones, to improve the termination rate of ventricular tachyarrhythmias.	Ila	C-EO

**Rarely, to limit patient discomfort and anxiety, hemodynamically stable slow VT can be treated without programming a backup shock.*

It is reasonable to program the initial shock energy to the maximum available energy in the highest rate detection zone to improve the first shock termination of ventricular arrhythmias unless specific defibrillation testing demonstrates efficacy at lower energies.	Ila	C-LD
---	-----	------

32. Intraoperative Testing of Defibrillation Efficacy

The efficacy of the ICD for the primary and secondary prevention of SCD has been well established in several landmark clinical trials [180–185]. Most of these trials have required induction, detection, and termination of VF at the time of implantation as a measure of defibrillation efficacy and as a surrogate of the ICD’s ability to prevent SCD. Testing defibrillation efficacy has been considered an integral part of ICD implantation for many years, and it is performed to establish the appropriate connection of high-voltage electrodes and to test the ability of the ICD to detect and terminate VF with a shock. However, identifying system failures or high defibrillation thresholds is difficult, mainly due to the low prevalence, which also depends on the definition employed, about 5% combined. Significant improvements over the past 2 decades have reduced energy requirements for defibrillation [186–189]. Similarly, current transvenous ICD technology is capable of delivering energies of 35–40 J, raising the question of the value of routine defibrillation testing (DT). Physicians have therefore gravitated to implanting ICDs with minimum or no DT with wide variability in practice, despite a paucity of rigorous data. DT is currently being performed during ICD implant in only about half the procedures [190–195]. Studies evaluating DT are summarized in Table 4.

One of the most important reasons to avoid DT at the time of ICD implantation is that testing might result in complications or even death. The risks of DT include (1) those related to VF itself, which can lead to circulatory arrest and hypoperfusion, (2) risks related to the shocks delivered to terminate VT, and (3) risks related to anesthetic drugs that are required for heavier sedation, which are used to provide patient comfort during testing.

33. Perioperative Mortality

Although improved ICD technology has led to the need for fewer inductions of VF at the time of implantation testing, procedure-related mortality has not been completely eliminated. Using modern ICD technology with transvenous systems and biphasic waveforms, the perioperative mortality rate within 30 days of implantation is reported to be 0.2% to 0.4% [191,196]. Recent data from the National Cardiovascular Data Registry (NCDR) demonstrated an in-hospital mortality of 0.03% following ICD implantation, with death occurring in the laboratory in 0.02% [196]. A Canadian report from 21 implanting centers estimates that 3 of 19,067 deaths (0.016%) are related to DT.

Table 4
Defibrillation testing.

Study (n)	Patients (DT/ no DT)	Results and remarks
CREDIT [193] (361) Prospective multicenter registry	64%/36%	More frequent DT for new implants vs generator replacements (71% vs 32%, $P=.0001$), DT for primary and secondary prevention indications (64% vs 63%, $P=NS$). Reasons for no DT were as follows: unnecessary (44%); persistent atrial fibrillation (37%); no anesthetist (20%); and patient or physician preference (6%). DT was not performed in a third of ICD implants, usually due to a perceived lack of need or relative contraindication. Nonconsecutive patients, single manufacturer.
Ontario DT Registry [225] (2173) Prospective multicenter registry	PP: 65%/45% SP: 67%/43% GR: 24%/76%	Multivariate predictors for DT included new ICD implant (OR 13.9; $P P P=.004$), and LVEF > 20% (OR 1.3; $P=.05$). History of AF (OR 0.58; $P=.0001$) or OAC use (OR: 0.75; $P=.03$) was associated with a lower likelihood of having DT. Complications, including death, were similar: DT 8.7% vs no DT 8.3% ($P=.7$). All consecutive implants at 10 centers in Ontario
NCDR [223] (64,277) Prospective multicenter registry	71%/29%	No DT; older, higher incidence of HF, lower LVEF, atrial arrhythmias, and a primary prevention indication; hospital adverse events; DT 2.56% vs. 3.58% no DT ($P P$ Generator replacement excluded.
Israel DFT Registry [226] (3596) Prospective multicenter registry	17%/83%	Variables associated with ICD testing: implantation for secondary prevention (relative risk [RR] 1.87), prior ventricular arrhythmias (RR 1.81), use of AADs (RR 1.59), and sinus rhythm (RR 2.05). No significant differences in the incidence of mortality, malignant ventricular arrhythmias, or inappropriate ICD discharges were observed between patients who underwent DT compared with those who were not tested. All consecutive implants during 1 year at 22 centers: HOCM: 6.2% DT, 6.3% no DT; ARVC: 0.6% DT, 0.5% no DT; congenital heart disease: 0.8% DT, 2.1% no DT; Long QT: 1.2% DT, 0.26% no DT; Brugada syndrome: 0.3% DT, 0.44% no DT; family history cardiac death: 5.3% DT, 4.7% no DT.
SAFE-ICD [210] 2120 Prospective observational study	836 DT 1284 no DT	Followed up for 24 months. Primary endpoint was composite of severe implant complications, sudden cardiac death, or resuscitation at 2 years. Primary endpoint: Of 34 patients, 12 intraoperative complications (8 in DT; 4 in no DT) and 22 during follow-up (10 in DT; 12 in no DT). Estimated yearly incidence: DT 1.15% (0.73 to 1.83) and no DT 0.68% (0.42 to 1.12); no difference. In 41 Italian centers. The only exclusion criterion was refusal to provide consent. Other ICD indications: 15% DT, 12% no DT.
Healey JS, et al [212] (145) Randomized Multicenter subgroup study	75 DT 70 no DT	All patients in DT arm achieved a successful DT ($\leq 25 J$); 96% without requiring any system modification. No patient experienced perioperative stroke, myocardial infarction, HF, intubation, or unplanned ICU stay. The composite of HF hospitalization or all-cause mortality occurred in 10% of no DT vs. 19% of the DT arm (HR: 0.53; 95% CI 0.21–1.31; $P=.14$). Conclusions: Perioperative complications, failed appropriate shocks, and arrhythmic death are uncommon regardless of DT. There was a nonsignificant increase in the risk of death or HF hospitalization with DT. Excluded: intracardiac thrombus, persistent or permanent AF without appropriate anticoagulation, right-sided implant, or felt ineligible for DT.
SIMPLE [215] 2500 Randomized multicenter trial	1253 DT 1247 no DT	Primary outcome: arrhythmic death or failed appropriate shock occurred in fewer patients (90 [7% per year]) in no DT vs DT (104 [8% per year]; HR 0.86; 95% CI 0.65–1.14; P noninferiority $P=.33$). The second safety composite outcome, including only events most likely to be directly caused by DT, occurred in 3.2% of patients without DT vs 4.5% with DT ($P=.08$). Routine DT at the time of ICD implantation is generally well tolerated but does not improve shock efficacy or reduce arrhythmic death. Single manufacturer, excluded patients on active transplantation list, ICD expected to be right-sided implant. HOCM: 4.2% DT, 3.4% no DT; long QT, Brugada syndrome, or catecholaminergic polymorphic VT: 2.3% DT, 1.9% no DT.
NORDIC ICD [216] 1077 Randomized multicenter trial	540 DT 537 no DT	ICD shocks were programmed to 40 J in all patients. Primary endpoint: first shock efficacy for all true VT and fibrillation episodes during 22.8 months of follow-up. Noninferior with or without DT. First shock efficacy 3.0% in favor of no DT. A total of 112 procedure-related serious adverse events occurred within 30 days in 94 DT patients (17.6%) and 89 events in 74 no-DT patients (13.9%). Excluded were the following: survived an episode of VF due to acute ischemia or potentially reversible causes, listed for heart transplant, life expectancy less than the study duration due to malignant conditions, terminal renal insufficiency, any conditions precluding DT (e.g., left atrial or ventricular thrombus), preexisting or previous ICD or CRT-D, or if the device was intended to be implanted on the right side.

Intraprocedural Testing of Defibrillation Efficacy Recommendations	Class of Recommendation	Level of Evidence
Defibrillation efficacy testing is recommended in patients undergoing a subcutaneous ICD implantation.	I	C-LD
It is reasonable to omit defibrillation efficacy testing in patients undergoing initial left pectoral transvenous ICD implantation procedures where appropriate sensing, pacing, and impedance values are obtained with fluoroscopically well-positioned RV leads.	Ila	B-R
Defibrillation efficacy testing is reasonable in patients undergoing a right pectoral transvenous ICD implantation or ICD pulse generator changes.	Ila	B-NR
Defibrillation efficacy testing at the time of implantation of a transvenous ICD should not be performed on patients with a documented nonchronic cardiac thrombus, atrial fibrillation or atrial flutter without adequate systemic anticoagulation, critical aortic stenosis, unstable CAD, recent stroke or TIA, hemodynamic instability, or other known morbidities associated with poor outcomes.	III (Harm)	C-LD

AAD=antiarrhythmic drug; ARVC=arrhythmogenic right ventricular cardiomyopathy; CAD=coronary heart disease; HOCM=hypertrophic obstructive cardiomyopathy; ICU=intensive care unit; OAC=oral anticoagulant

34. DT-Related Complications

Complications occurring during ICD implantation procedures are infrequent, and many can be directly or indirectly related to DT. Adverse effects related to DT include myocardial injury, depression of contractile function leading to worsening of HF,

persistent hypotension, central nervous system injury, thromboembolic events, or respiratory depression.

Transient central nervous system hypoperfusion and cerebral ischemic changes can be demonstrated during intraoperative electroencephalographic (EEG) monitoring at the time of DT. However, EEG recovery occurs within less than 30 seconds, with a

slightly longer time to the return of middle cerebral blood flow [197–199]. However, the clinical relevance of this transient finding is unclear because DT does not appear to cause cognitive dysfunction 24–48 hours following ICD implantation [200,201]. Although an increase in biochemical markers of myocardial injury can be observed during ICD implantation or after spontaneous clinical shocks, true intraoperative myocardial infarction (MI) is rare, even when extensive DT is performed [202–205]. In 2 recent studies using transvenous ICDs and a more abbreviated testing protocol, there was no significant increase in CK, CK-MB, myoglobin, and NT-proBNP before and after DT, whereas elevated levels of high-sensitive troponin T were observed after DT [206,207]. In the NCDR ICD Registry, the incidence of MI during ICD implantation was reported to be 0.02% [196].

Defibrillator shocks and VF transiently depress contractile function, although fatal pulseless electrical activity is rare at the time of ICD implantation [202,206,208–210]. Refractory VF has been reported to occur during DT, but this is also uncommon, particularly with contemporary devices. One study reported that all tested ICD shocks failed and at least 3 external rescue shocks were required in 0.5% of patients [203]. A Canadian study reported that 27 of 19,067 implants (0.14%) required prolonged resuscitations during DT [211].

Thromboembolic complications can occur during DT in the presence of intracardiac thrombus or when there are less than 3 weeks of therapeutic and uninterrupted anticoagulation in the setting of AF. Stroke or transient ischemic attack (TIA) is reported to occur in 0.026%–0.05% of cases [204,211]. Multiple strategies have been employed, but none were documented to reduce the incidence of thromboembolism, including the avoidance of DT. These include preprocedure transesophageal echocardiography to exclude left atrial appendage thrombus and deferring testing when a thrombus is identified, or using transthoracic echocardiography to detect LV thrombi.

Anesthetic agents can contribute to complications related to a depressant effect on myocardial contractility or can lead to respiratory depression if oversedation occurs. Heavier sedation is typically used in patients undergoing DT. Although patients with underlying chronic obstructive pulmonary disease or sleep apnea might be at increased risk, oversedation and respiratory depression could occur in any patient. Randomized trial data can help to identify which adverse events are directly (or indirectly) related to DT. For example, stroke or TIA might be “directly” related to DT due to dislodgment of intracardiac thrombus during conversion of AF in the absence of therapeutic anticoagulation, and an episode of prolonged hypotension could result in reduced cerebral perfusion. Respiratory depression, respiratory failure requiring intubation, or hypotension might be direct results of DT or might be due to the drugs required to perform testing. Pulseless electrical activity or even death can occur with hemodynamic complications related to induction of VF or multiple external shocks. In contrast, DT can indirectly increase the risk for pneumothorax, perforation, tamponade, lead dislodgment, or infection as more leads are inserted, or the procedure might be prolonged due to the system modifications required to improve defibrillation efficacy; however, all these complications can also occur in the absence of DT. In addition, due to the rates and types of adverse events reported in the literature, it appears that overall complication rates are primarily driven by mechanical complications or infection, most of which are not related to DT.

In a substudy of the Resynchronization for Ambulatory Heart Failure Trial (RAFT), in which 145 patients were randomized to DT compared with no DT at the time of initial ICD implantation, the risk of perioperative complications was extremely low, regardless of DT performance [212]. There was, however, a nonsignificant increase in the risk of death or HF hospitalization in the group that underwent DT. Likewise, no significant difference in implant-

related complications was demonstrated in DT compared with the groups without DT in the Safety of Two Strategies of ICD Management at Implantation (SAFE-ICD) study, a prospective observational study of 2120 patients performed at 41 centers [213]. Similar findings were observed in the prospective randomized Test-No Test Implantable Cardioverter Defibrillator (TNT-ICD) pilot study on 66 patients, in which there was no difference in adverse events between patients who underwent testing compared with those who did not [214].

The Shockless Implant Evaluation (SIMPLE) trial is the largest randomized study assessing the effect of DT on clinical outcomes [215]. This large-scale study randomized 2500 patients to DT or not at the time of ICD implantation; 1253 patients were randomly assigned to DT and 1247 were assigned to no-testing, and were followed for a mean of 3.1 years (SD 1.0). The primary outcome of arrhythmic death or failed appropriate shock was noninferior (90 [7% per year] in the no-testing group compared with patients undergoing DT (104 [8% per year]; HR 0.86; 95% CI 0.65–1.14; *P* noninferiority *P* = .33). The second, pre-specified safety composite outcome, which included only events most likely to be directly caused by testing, occurred in 3.2% of patients with no testing and in 4.5% with DT (*P* = .08). Heart failure needing intravenous treatment with inotropes or diuretics was the most common adverse event (in 20 of 1236 patients [2%] in the no-testing group vs 28 of 1242 patients [2%] in the testing group, *P* = .25). In summary, routine DT at the time of ICD implantation is generally well tolerated without a statistically significantly increased rate of complications, but it also does not improve shock efficacy or reduce arrhythmic death.

Finally, the No Regular Defibrillation Testing In Cardioverter Defibrillator Implantation (NORDIC-ICD) trial, another prospective randomized parallel group multicenter noninferiority trial conducted in 48 centers in Europe, assessed the effects of DT at the time of ICD implantation on first shock efficacy [216]. The primary endpoint was different from the SIMPLE trial and assessed the average first- shock efficacy for all true VT and VF episodes occurring in any patient during follow-up. NORDIC-ICD randomized 540 patients to DT and 537 to no DT at the time of ICD implantation. During a median follow-up of 22.8 months, the first shock efficacy was demonstrated to be noninferior in the patients undergoing ICD implantation without DT, with a difference in first shock efficacy of 3.0% in favor of the no- DT test group (95% CI 3.0%–9.0%; *P* noninferiority procedure-related serious adverse events were reported within 30 days of ICD implantation in 94 patients (17.6%) undergoing DT compared with 74 patients (13.9%) not undergoing DT (*P* = .095). The authors concluded that defibrillation efficacy without DT was noninferior to ICD implantation with DT in left-sided ICD implants. Because no major benefit or harm associated with DT was detected, in patients with a left-sided pectoral implantation it is reasonable to omit routine VF induction and DT during ICD implantation, assuming stable ICD lead position and good sensing and capture function [217–220]. This approach is particularly applicable to patients with ischemic and idiopathic dilated cardiomyopathy, given these entities were well represented in the studied cohort. Patients well represented within the cohort included those with implantation in the left pectoral location, those indicated for primary and secondary prevention of SCD, and patients with ischemic and nonischemic cardiomyopathies. Fewer data are available regarding other cardiomyopathies, such as patients with hypertrophic obstructive cardiomyopathy, congenital channelopathies, patients undergoing generator replacement, and procedures in the right pectoral location. In these instances, and when there is any question of the adequacy of the lead position or function, DT is reasonable. It is worth emphasizing that a nontesting strategy requires an anatomically well- positioned defibrillation lead in the RV with adequate sensing of intrinsic R waves (> 5–7 mV), adequate pacing thresholds, and a thorough verification of proper lead connection.

Other important considerations include the use of alternative RV defibrillation lead sites such as the mid-septum. Pooled data from 2 randomized studies do not indicate a clinically relevant elevation of energy required for defibrillation with mid-septal sites. Positioning of the RV defibrillation lead in other positions such as the RV outflow tract has not been systematically addressed [221].

The SIMPLE trial data were consistent between subgroups, both from patients with single- or dual-coil ICD leads and with or without the use of amiodarone. More recently, the Multicenter Comparison of Shock Efficacy Using Single vs Dual-Coil Lead Systems and Anodal vs Cathodal Polarity Defibrillation in Patients Undergoing Transvenous Cardioverter-Defibrillator Implantation (MODALITY) study was reported [222]. This was a multicenter registry that prospectively followed 469 consecutive patients undergoing DT at the time of implant; 158 (34%) had dual-coil and 311 (66%) had single-coil lead systems configuration, 254 (54%) received anodal shock, and 215 (46%) received cathodal shock. In 35 patients (7.4%), the shock was unsuccessful. No significant differences in the outcome of DT using a single- vs dual-coil lead were observed, but the multivariate analysis showed an increased risk of shock failure using cathodal shock polarity (odds ratio [OR] 2.37; 95% CI 1.12–5.03). These and other registry data support the use of either single- or dual-coil leads, preferably programmed to deliver anodal shocks [211,213,223].

Performing DT has not been determined to be harmful or inappropriate. One reason to perform DT in specific populations is that high defibrillation thresholds have been reported in 2.2% to 12% of subjects undergoing DT. The probabilistic nature of DT with the failure of a single shock 10 J below the maximum ICD output does not necessarily imply long-term ICD failure. Determinations of DT using multiple shock protocols have reported that a safety margin of only 5.2–1.1 J has a 97.3% rate of successful VT/VF conversion [224]; however, the inability to convert VF at maximum output occurs in approximately 1% of procedures during DT. The long-term outcomes of these patients have not been evaluated without modification of the lead system. Further supporters of DT suggest that routine testing is necessary to identify system integrity and sensing failures. R-wave amplitude ≤ 5 –7 mV at implant almost invariably reliably sense VF [190,221]. Failure to sense and some inner insulation failures might only be detected by DT. This situation has not been systematically evaluated.

35. Contraindications to Defibrillation Threshold Testing

A great paucity of systematic data limits the assessment of the literature regarding contraindications to DT. Most implanters tend to avoid DT in patients perceived to be at high risk. Information derived from an NCDR-ICD registry identified advanced age, impaired LVEF, NYHA Class IV HF, atrial fibrillation/flutter, need to withhold warfarin, and several other factors as high-risk situations. Unfortunately, the strength of these associations was weak, given the ORs were under 2 [223]. Other registries have identified patients with broader QRS durations, advanced NYHA class, and CRT as reasons for not performing DT [213]. There are no convincing data to identify high-risk patients, and clinical judgment has likely kept the highest-risk patients, particularly those who were hemodynamically unstable, from being tested in the current literature.

36. S-ICD

Patients receiving a nontransvenous ICD system should routinely undergo DT, given there are no current data regarding the safety and efficacy of not performing DT with this lead configuration and device.

37. Conclusion

In providing focused recommendations for ICD programming and DT of patients implanted with a device we have intentionally left many questions unanswered. There are hundreds of choices for which there are inadequate data to provide evidence or consensus-based recommendations. This document is a long overdue effort to provide analysis and guidance to the clinician as to how to make strategic programming choices in the implementation of ICD therapy. The four continental electrophysiology societies limited the discussion and recommendations to four areas for which there was sufficient consensus and data. In the review process, clearly articulated opinions pointed out that additional recommendations are desirable. However, there is an information gap of insufficient data filled with opinions and logical arguments. Generalizations and inferences were made from the existing data, e.g., taking data from pacemaker trials and applying to this to ICD bradycardia programming, logical arguments bridging the differences between primary and secondary prevention patients for tachycardia detection and therapy, and the use of noninferiority data to make decisions about DT. This document

Table A1
Author Disclosure Table.

Writing Group Member	Institution	Consultant/Advisory Board/Honoraria	Speakers' Bureau	Research Grant	Fellowship Support	Stock Options/ Partner	Board Mbs/Other
Bruce L. Wilkoff, MD, FHRS, CCDS (Chair)	Cleveland Clinic, Cleveland, OH	1: St. Jude Medical, Boston Scientific Corp. 2: Spectranetics Corporation, Medtronic, Inc.	None	None	None	None	Equity interests: 1 CardioMEMS; Royalty Income: 2 Medtronic, Inc. None
Laurent Fauchier, MD, PhD (Co-Chair)	Centre Hospitalier Universitaire Trousseau, Tours, France	1: Bayer HealthCare, LLC; Boehringer Ingelheim; Boston Scientific Corp.; Brital Meyers Squibb; Medtronic, Inc.; Novartis Pharmaceuticals Corp; Sanofi Aventis; Daiichi; Sorin Group	None	None	None	None	None
Martin K. Stiles, MBCHB, PhD (Co-Chair)	Waikato Hospital, Cardiology, Hamilton, New Zealand	1: Medtronic, Inc.; Boston Scientific Corp.; Biotronik	None	None	2: Medtronic, Inc. 3: Biosense Webster, Inc. St. Jude Medical	None	None
					None	None	None

Table A1 (continued)

Writing Group Member	Institution	Consultant/Advisory Board/Honoraria	Speakers' Bureau	Research Grant	Fellowship Support	Stock Options/ Partner	Board Mbs/Other
Carlos A. Morillo, MD, FRCPC, FHRS (Co-Chair)	McMaster University, Hamilton, Canada	1: Sanofi Aventis; Biotronik; 2: Boehringer Ingelheim; Marck Pharmaceuticals	1: Boehringer Ingelheim; Sanofi Aventis; 2: Merck Pharmaceuticals	3: St. Jude Medical; 4: Medtronic, Inc.; Boston Scientific Corp.			
Sana Al-Khatib, MD, MHSc, FHRS, CCDS	Duke University Medical Center, Durham, NC	None	None	None	None	None	None
Jesús Almendral, MD, PhD, FESC	Hospital General Alicante, Alicante, Spain	1: Boston Scientific, St. Jude Medical, Medtronic	None	None	2: St. Jude Medical	None	None
Luis Aguinaga, MD, PhD, FACC, FESC	Centro Privado De Cardiología, Tucuman, Argentina	None	None	None	None	None	None
Ronald D. Berger, MD, PhD, FHRS	Johns Hopkins University, Baltimore, Maryland	2: Boston Scientific Corp;	None	None	3: Medtronic, Inc; 4: St. Jude Medical	None	Royalty Income:3: Zoll Medical Corporation
Alejandro Cuesta, MD, PhD, FESC	Montevideo, Uruguay	None	None	None	None	None	None
James P. Dautbert, MD, FHRS	Duke University Medical Center, Durham, North Carolina	1: Medtronic, Inc.; St. Jude Medical; Boston Scientific Corp.; Sorin Group; CardioFocus, Inc.; Gilead Sciences, Inc.; Biosense Webster, Inc.; Biotronik; Sanofi Aventis	None	5: Boston Scientific Corp.; Biosense Webster, Inc.; Medtronic, Inc.; Gilead Sciences Inc.	3: Medtronic, Inc.; Boston Scientific Corp.; Biotronik; St. Jude Medical; Biosense Webster, Inc.; Bard Electrophysiology	None	None
Sergio Dubner, MD, FACC	Clinica y Maternidad Suizo Argentina and De Los Arcos Sanatorio, Buenos Aires, Argentina	1: Boehringer Ingelheim	None	None	None	None	0; Officer, trustee, director or any other fiduciary role: Cardiology Journal
Kenneth A. Ellenbogen, MD, FHRS	Virginia Commonwealth University Medical Center, Richmond, VA	1: Biotronik; St. Jude Medical; Cameron Health Corp.; American College of Cardiology; American Heart Association; CardioNet, Inc.; Daiichi; Janssen Pharmaceuticals; Biosense Webster, Inc.; AtriCure, Inc.; Pfizer, Inc. 2: Boston Scientific Corp.; Medtronic, Inc.; Sentreheart; Endosense; Elsevier	None	1: National institutes for Health; St. Jude Medical; 2: Biosense Webster, Inc. 3: Medtronic, Inc.; Boston Scientific Corp.	2: Medtronic, Inc. 3: Boston Scientific Corp; Biosense Webster, Inc.	None	None
N.A. Mark Estes III, MD	New England Medical Center, Boston, Massachusetts	1: Medtronic, Inc.; St. Jude Medical; 2: Boston Scientific Corp.	None	4: Boston Scientific Corp.	4: Medtronic, Inc.; Boston Scientific Corp.; St. Jude Medical	None	None
Guilherme Fenelon, MD, PhD	Federal University of S'ao Paulo, S'ao Paulo, Brazil	None	None	2: Biosense Webster, Inc.	None	None	None
Fermin C. Garcia, MD	Hospital of the University of Pennsylvania, Philadelphia, Pennsylvania	None	1: St. Jude Medical; Biotronik; 2: Boston Scientific Corp.	1: Biosense Webster, Inc.		None	None
Maurizio Gasparini, MD	Humanitas Research Hospital, Milan, Italy	None	1: Boston Scientific Corp.; Medtronic, Inc.	None		None	None
David E. Haines, MD, FHRS	William Beaumont Hospital Division of Cardiology, Royal Oak, Michigan	None	None	None		None	1: nContact, Inc.
Jeff S. Healey, MD, MSc, FRCPC, FHRS	Population Health Research Institute, McMaster University, Hamilton, Canada	1: Boehringer Ingelheim; St. Jude Medical; Bayer HealthCare, LLC	None	4: Bristol Meyers Squibb; 5: Boston Scientific Corp.; St. Jude Medical; Boehringer Ingelheim; Medtronic, Inc.; Bristol Meyers Squibb; Pfizer, Inc.		None	None
Jodie L. Hurtwitz, MD	North Texas Heart Center, Dallas, Texas	1: Biosense Webster, Inc.	1: Boehringer Ingelheim; Medtronic, Inc.; St. Jude Medical	None		None	None
Roberto Keegan, MD	Hospital Privado del Sur, Bahia Blanca, Argentina	None	None	None		None	None
Christof Kolb, MD	Deutsches Herzzentrum Munchen, Munich, Germany	1: Biotronik; Boston Scientific; Medtronic; Sorin Group; St.	None	1: Biotronik, Sorin Group; St. Jude Medical		None	None

Table A2 (continued)

Peer Reviewer	Institution	Consultant/Advisory Board/ Honoraria	Speakers' Bureau	Research Grant	Fellowship Support	Stock Options/ Partner	Board Mbs/ Other
Luigi Di Biase, MD, PhD, FACC, FHRS	Montefiore-Einstein Center for Heart & Vascular Care, Bronx, NY	1: Stereotaxis; St. Jude Medical; Boston Scientific Corp.; Medtronic, Inc.; EpiEP, Inc.; Janssen; Pfizer 2: Biosense Webster, Inc.; Biotronik	None	None	None	None	None
Kevin Donahue, MD	University of Massachusetts Medical School, Worcester, MA	None	None	0: Medtronic, Inc.; Boston Scientific Corp.; St. Jude Medical	None	None	1: NIH (salary)
Andrew E. Epstein, MD, FAHA, FACC, FHRS	Philadelphia VA Medical Center, Philadelphia, PA	1: Medtronic, Inc. 2: Boston Scientific Corp.; St. Jude Medical	None	4: Boston Scientific Corp.; Biotronik, Medtronic, Inc.; St. Jude Medical	4: Boston Scientific Corp.; Biotronik; Medtronic; St. Jude Medical	None	None
Michael E. Field, MD, FACC, FHRS	University of Wisconsin School of Medicine and Public Health, Madison, WI	None	None	None	None	None	None
Bulent Gorennek, MD, FACC, FESC	Eskisehir Osmangazi University, Eskisehir, Turkey	None	None	None	None	None	None
Jin-Long Huang, MD, PhD	School of Medicine, National Yang-Ming University, Taichung City, Taiwan	1: Pfizer; Sanofi-Aventis; AstraZeneca; Boehringer Ingelheim	None	None	None	None	None
Julia H. Indik, MD, PhD, FACC, FAHA, FHRS	University of Arizona, Sarver Heart Center, Tucson, AZ	None	None	None	None	None	None
Carsten W. Israel, MD	Chefarzt der Klinik für Innere Medizin-Kardiologie, Diabetologie und Nephrologie des Ev. Krankenhauses Bielefeld Privatdozent der J.W. Goethe-Universität Frankfurt	0: European Society of Cardiology 1: Biotronik; Boston Scientific Corp.; Medtronic, Inc.; Sorin Group; St. Jude Medical	None	1: Biotronik; Boston Scientific Corp.; Medtronic, Inc.; Sorin Group; St. Jude Medical	1: Medtronic, Inc.; Sorin Group; St. Jude Medical	None	None
Mariell L. Jessup, MD, FACC, FAHA, FESC	University of Pennsylvania School of Medicine, Philadelphia, PA	None	None	None	None	None	None
Christophe Leclercq, MD, PhD	CHU Pontchaillou, Rennes, France	1: Biotronik; Medtronic, Inc.; Sorin Group; Boston Scientific Corp.; Bayer; Bristol Meyers Squibb 2: St. Jude Medical	None	None	None	None	None
Robert J. MacFadyen, MD, PhD	Ballarat Health and University of Melbourne, Victoria, Australia	None	None	None	None	None	None
Christopher Madias, MD, FHRS	Rush University Medical Center, Chicago, IL	None	None	None	None	None	None
Manlio F. Marquez, MD, FACC	Sociedad Mexicana de Electrofisiología y Estimulación Cardíaca (SOMEEC), Mexico City, Mexico	None	None	None	None	None	None
Brian Olshansky, MD, FACC, FAHA, FHRS	University of Iowa Carver College of Medicine, Iowa City, IA	1: Daiichi Sankyo; Boston Scientific Corp.; Medtronic, Inc.; Biotronik; BioControl; Amarin; On-X; Boehringer Ingelheim; Lundbeck	None	1: Daiichi Sankyo	None	None	None
Kristen K. Patton, MD	University of Washington, Seattle, WA	None	None	None	None	None	0: American College of Cardiology; American Heart Association
Marwan M. Refaat, MD, mMBA, FACC, FAHA, FHRS, FASE, FESC, FACP, FAAMA	American University of Beirut Faculty of Medicine and Medical Center, Beirut, Lebanon	None	None	None	None	None	None
Cynthia M. Tracy, MD, FACC, FAHA	George Washington University, Washington, DC	None	None	None	None	None	None
Gaurav A. Upadhyay, MD	University of Chicago Medicine and Biological Sciences, Chicago, IL	None	None	4: Medtronic, Inc.	None	None	None
Diego Vanegas, MD, FHRS	Electrophysiology Unit, Hospital Militar Central, Bogot/E, Colombia	None	None	None	None	None	None
Paul J. Wang, MD, FHRS, CCDS	Stanford School of Medicine, Stanford, CA	1: Janssen, Medtronic, Inc.	None	3: Biosense Webster; Johnson & Johnson; St. Jude Medical; Medtronic, Inc.; Boston Scientific Corp.	None	1: VytronUS	None

Number Value: 0=\$0; 1=10,000 to\$25,000 to 50,000 to \$100,000.

is a beginning, necessary because there are now sufficient data to support recommendations that improve the safety, morbidity, and mortality of patients with ICDs.

Appendix. Disclosures

See Tables A1 and A2.

References

- [1] Jacobs AK, Anderson JL, Halperin JL, et al. The evolution and future of ACC/AHA clinical practice guidelines: a 30-year journey: a report of the American College of Cardiology/American Heart Association Task Force on practice guidelines. *Circulation* 2014;130(14):1208–17 Sep 30.
- [2] Anderson JL. Evolution of the ACC/AHA Clinical Practice Guidelines in Perspective. *J Am Coll Cardiol* 2015;65(25):2735–8.
- [3] Nielsen JC, Thomsen PEB, Højberg S, et al. A comparison of single-lead atrial pacing with dual-chamber pacing in sick sinus syndrome. *Eur Heart J* 2011;32(6):686–96.
- [4] Connolly SJ, Kerr CR, Gent M, et al. Effects of Physiologic Pacing versus Ventricular Pacing on the Risk of Stroke and Death Due to Cardiovascular Causes. *N Engl J Med* 2000;342(19):1385–91.
- [5] Lamas GA, Lee KL, Sweeney MO, et al. Ventricular Pacing or Dual-Chamber Pacing for Sinus-Node Dysfunction. *N Engl J Med* 2002;346(24):1854–62.
- [6] Lamas GA, Orav EJ, Stambler BS, et al. Quality of Life and Clinical Outcomes in Elderly Patients Treated with Ventricular Pacing as Compared with Dual-Chamber Pacing. *N Engl J Med* 1998;338(16):1097–104.
- [7] Toff WD, Camm AJ, Skehan JD. Single-Chamber versus Dual-Chamber Pacing for High-Grade Atrioventricular Block. *N Engl J Med* 2005;353(2):145–55.
- [8] Healey JS. Cardiovascular Outcomes With Atrial-Based Pacing Compared With Ventricular Pacing: Meta-Analysis of Randomized Trials, Using Individual Patient Data. *Circulation* 2006;114(1):11–7.
- [9] Casteluovo E, Stein K, Pitt M, Garside R, Payne E. The effectiveness and cost-effectiveness of dual-chamber pacemakers compared with single-chamber pacemakers for bradycardia due to atrioventricular block or sick sinus syndrome: systematic review and economic evaluation. *Health Technol Assess* 2005;9(43):1–246 iii, xi–xiii.
- [10] Theuns DA, Klootwijk AP, Goedhart DM, Jordaens LJ. Prevention of inappropriate therapy in implantable cardioverter-defibrillators: results of a prospective, randomized study of tachyarrhythmia detection algorithms. *J Am Coll Cardiol* 2004;44(12):2362–7.
- [11] Chen BW, Liu Q, Wang X, Dang AM. Are dual-chamber implantable cardioverter-defibrillators really better than single-chamber ones? A systematic review and meta-analysis *J Interv Card Electrophysiol* 2014;39(3):273–80.
- [12] Wilkoff BL, Cook JR, Epstein AE, et al. Dual-chamber pacing or ventricular backup pacing in patients with an implantable defibrillator: the Dual Chamber and VVI Implantable Defibrillator (DAVID). *Trial JAMA* 2002;288(24):3115–23.
- [13] Wilkoff BL, Kudenchuk PJ, Buxton AE, et al. The DAVID (Dual Chamber and VVI Implantable Defibrillator) II trial. *J Am Coll Cardiol* 2009;53(10):872–80 Mar 10.
- [14] Casteluovo E, Stein K, Pitt M, Garside R, Payne E. The effectiveness and cost-effectiveness of dual-chamber pacemakers compared with single-chamber pacemakers for bradycardia due to atrioventricular block or sick sinus syndrome: systematic review and economic evaluation. *Health Technol Assess* 2005;9(43) 11 2005.
- [15] Magri D, Corrà U, Di Lenarda A, et al. Cardiovascular mortality and chronotropic incompetence in systolic heart failure: the importance of a reappraisal of current cut-off criteria. *Eur J Heart Fail* 2013;2017;16(2):201–9.
- [16] Sims DB, Mignatti A, Colombo PC, et al. Rate responsive pacing using cardiac resynchronization therapy in patients with chronotropic incompetence and chronic heart failure. *Europace* 2011;13(10):1459–63.
- [17] Padeletti L, Pieragnoli P, Di Biase L, et al. Is a Dual-Sensor Pacemaker Appropriate in Patients with Sino-Atrial Disease? Results from the DUSISLOG Study Pacing Clin Electrophysiol 2006;29(1):34–40.
- [18] Sulke N, Chambers J, Dritsas A, Sowton E. A randomized double-blind crossover comparison of four rate-responsive pacing modes. *J Am Coll Cardiol* 1991;17(3):696–706.
- [19] Lamas GA, Knight JD, Sweeney MO, et al. Impact of rate-modulated pacing on quality of life and exercise capacity—Evidence from the Advanced Elements of Pacing Randomized Controlled Trial (ADEPT). *Heart Rhythm* 2007;4(9):1125–32.
- [20] Lau CP, Rushby J, Leigh-Jones M, et al. Symptomatology and quality of life in patients with rate-responsive pacemakers: A double-blind, randomized, crossover study. *Clin Cardiol* 1989;12(9):505–12.
- [21] Oto MA, Muderrisoglu H, Ozin MB, et al. Quality of Life in Patients with Rate Responsive Pacemakers: A Randomized, Cross-Over Study. *Pacing Clin Electrophysiol* 1991;14(5):800–6.
- [22] Dell'Orto S, Valli P, Greco EM. Sensors for rate responsive pacing. *Indian Pacing Electrophysiol J* 2004;4(3):137–45.
- [23] Martin DO, Day JD, Lai PY, et al. Atrial support pacing in heart failure: results from the multicenter PEGASUS CRT trial. *J Cardiovasc Electrophysiol* 2012;23(12):1317–25.
- [24] Leung S-K, Lau C-P. DEVELOPMENTS IN SENSOR-DRIVEN PACING. *Cardiol Clin* 2000;18(1):133–55.
- [25] Van Campen LCMC De Cock CC, Visser FC, Visser CA. The Effect of Rate Responsive Pacing in Patients with Angina Pectoris on the Extent of Ischemia on 201-Thallium Exercise Scintigraphy. *Pacing Clin Electrophysiol* 2002;25(4):430–4.
- [26] Brignole M, Auricchio A, Baron-Esquivias G, et al. 2013 ESC Guidelines on cardiac pacing and cardiac resynchronization therapy: The Task Force on cardiac pacing and resynchronization therapy of the European Society of Cardiology (ESC). Developed in collaboration with the European Heart Rhythm Association (EHRA). *Eur Heart J* 2013;34(29):2281–329.
- [27] Geelen P, Brugada J, Andries E, Brugada P. Ventricular Fibrillation and Sudden Death After Radiofrequency Catheter Ablation of the Atrioventricular Junction. *Pacing Clin Electrophysiol* 1997;20(2):343–8.
- [28] Nowinski K, Gadler F, Jensen-Urstad M, Bergfeldt L. Transient proarrhythmic state following atrioventricular junction radiofrequency ablation: pathophysiological mechanisms and recommendations for management. *Am J Med* 2002;113(7):596–602.
- [29] Andersen HR, Nielsen JC, Thomsen PEB, et al. Long-term follow-up of patients from a randomised trial of atrial versus ventricular pacing for sick-sinus syndrome. *The Lancet* 1997;350(9086):1210–6.
- [30] Sweeney MO. Adverse Effect of Ventricular Pacing on Heart Failure and Atrial Fibrillation Among Patients With Normal Baseline QRS Duration in a Clinical Trial of Pacemaker Therapy for Sinus Node Dysfunction. *Circulation* 2003;107(23):2932–7.
- [31] The DTI. Dual-Chamber Pacing or Ventricular Backup Pacing in Patients With an Implantable Defibrillator. *JAMA* 2002;288(24):3115.
- [32] Sharma AD, Rizo-Patron C, Hallstrom AP, et al. Percent right ventricular pacing predicts outcomes in the DAVID trial. *Heart Rhythm* 2005;2(8):830–4.
- [33] Olshansky B, Day JD, Moore S, et al. Is dual-chamber programming inferior to single-chamber programming in an implantable cardioverter-defibrillator? Results of the INTRINSIC RV (Inhibition of Unnecessary RV Pacing With AVSH in ICDs) study *Circulation* 2007;115(1):9–16 Jan 2.
- [34] Steinberg JS, Fischer AVI, Wang P, et al. The Clinical Implications of Cumulative Right Ventricular Pacing in the Multicenter Automatic Defibrillator Trial II. *J Cardiovasc Electrophysiol* 2005;16(4):359–65.
- [35] Smit MD, PFHM Van Dessel, Nieuwland W, et al. Right ventricular pacing and the risk of heart failure in implantable cardioverter-defibrillator patients. *Heart Rhythm* 2006;3(12):1397–403.
- [36] Gillis AM, Purerfellner H, Israel CW, et al. Reducing Unnecessary Right Ventricular Pacing with the Managed Ventricular Pacing Mode in Patients with Sinus Node Disease and AV Block. *Pacing Clin Electrophysiol* 2006;29(7):697–705.
- [37] Kolb C, Schmidt R, Dietl JU, et al. Reduction of Right Ventricular Pacing with Advanced Atrioventricular Search Hysteresis: Results of the PREVENT Study. *Pacing Clin Electrophysiol* 2011;34(8):975–83.
- [38] Akerström F, Arias MA, Pachón M, Puchol A, Jimóñez-López J, Rodríguez-Padial L. The reverse mode switch algorithm: How well does it work? *Heart Rhythm* 2013;10(8):1146–52.
- [39] Pürerfellner H, Brandt J, Israel C, et al. Comparison of Two Strategies to Reduce Ventricular Pacing in Pacemaker Patients. *Pacing Clin Electrophysiol* 2008;31(2):167–76.
- [40] Murakami Y, Tsuboi N, Inden Y, et al. Difference in percentage of ventricular pacing between two algorithms for minimizing ventricular pacing: results of the IDEAL RVP (Identify the Best Algorithm for Reducing Unnecessary Right Ventricular Pacing) study. *Europace* 2009;12(1):96–102.
- [41] Sweeney MO, Bank AJ, Nshah E, et al. Minimizing Ventricular Pacing to Reduce Atrial Fibrillation in Sinus-Node Disease. *N Engl J Med* 2007;357(10):1000–8.
- [42] Sweeney MO, Ellenbogen KA, Tang ASL, et al. Atrial pacing or ventricular backup-only pacing in implantable cardioverter-defibrillator patients. *Heart Rhythm* 2010;7(11):1152–560.
- [43] Steinbach M, Douchet M-P, Bakouboula B, Bronner F, Chauvin M. Outcome of patients aged over 75 years who received a pacemaker to treat sinus node dysfunction. *Arch Cardiovasc Dis* 2011;104(2):89–96.
- [44] Veasey RA, Arya A, Silberbauer J, et al. The relationship between right ventricular pacing and atrial fibrillation burden and disease progression in patients with paroxysmal atrial fibrillation: the long-MinVPACE study. *Europace* 2011;13(6):815–20.
- [45] Rey J-L, Quenum S, Hero M. Effect of a Pacing Mode Preserving Spontaneous AV Conduction on Ventricular Pacing Burden and Atrial Arrhythmias. *Pacing Clin Electrophysiol* 2012;35(5):580–5.
- [46] L. Botto G, Ricci RP, BØnØzet JM, et al. Managed ventricular pacing compared with conventional dual-chamber pacing for elective replacement in chronically paced patients: Results of the Prefer for Elective Replacement Managed Ventricular Pacing randomized study. *Heart Rhythm* 2014;11(6):992–1000.
- [47] Boriani G, Tukkie R, Manolis AS, et al. Atrial antitachycardia pacing and managed ventricular pacing in bradycardia patients with paroxysmal or persistent atrial tachyarrhythmias: the MINERVA randomized multicentre international trial. *Eur Heart J* 2014;35(35):2352–62.

- [48] Chen S, Chen K, Tao Q, et al. Reduction of unnecessary right ventricular pacing by managed ventricular pacing and search AV+ algorithms in pacemaker patients: 12-month follow-up results of a randomized study. *Europace* 2014;16(11):1595–602.
- [49] Gillis AM, Russo AM, Ellenbogen KA, et al. HRS/ACCF Expert Consensus Statement on Pacemaker Device and Mode Selection. *Heart Rhythm* 2012;9(8):1344–65.
- [50] Nielsen JC, Thomsen PEB, Højberg S, et al. Atrial fibrillation in patients with sick sinus syndrome: the association with PQ-interval and percentage of ventricular pacing. *Europace* 2011;14(5):682–9.
- [51] Den Dulk K, Lindemans FW, Brugada P, Smeets JLRM, Wellens HJJ. Pacemaker Syndrome with AAI Rate Variable Pacing: Importance of Atrioventricular Conduction Properties, Medication, and Pacemaker Programmability. *Pacing Clin Electrophysiol* 1998;11(11):1226–33.
- [52] Sweeney MO. Novel Cause of Spurious Mode Switching in Dual-Chamber Pacemakers: Atrioventricular Desynchronization Arrhythmia. *J Cardiovasc Electrophysiol* 2002;13(6):616–9.
- [53] Pascale P, Pruvot E, Graf D. Pacemaker Syndrome During Managed Ventricular Pacing Mode: What is the Mechanism? *J Cardiovasc Electrophysiol* 2009;20(5):574–6.
- [54] Sweeney MO, Ruetz LL, Belk P, Mullen TJ, Johnson JW, Sheldon T. Bradycardia Pacing-Induced Short-Long-Short Sequences at the Onset of Ventricular Tachyarrhythmias. *J Am Coll Cardiol* 2007;50(7):614–22.
- [55] Vavasis C, Slotwiner DJ, Goldner BG, Cheung JW. Frequent Recurrent Polymorphic Ventricular Tachycardia during Sleep Due to Managed Ventricular Pacing. *Pacing Clin Electrophysiol* 2009;33(5):641–4.
- [56] Sweeney MO, Ellenbogen KA, Tang ASL, Johnson J, Belk P, Sheldon T. Severe Atrioventricular Decoupling, Uncoupling, and Ventriculoatrial Coupling During Enhanced Atrial Pacing: Incidence, Mechanisms, and Implications for Minimizing Right Ventricular Pacing in ICD Patients. *J Cardiovasc Electrophysiol* 2008;19(11):1175–80.
- [57] Gersh BJ, Maron BJ, Bonow RO, et al. 2011 ACCF/AHA Guideline for the Diagnosis and Treatment of Hypertrophic Cardiomyopathy: Executive Summary: A Report of the American College of Cardiology Foundation/American Heart Association Task Force on Practice Guidelines. *Circulation* 2011;124(24):2761–96.
- [58] Nishimura RA, Hayes DL, Ilstrup DM, Holmes DR, Jamil Tajik A. Effect of dual-chamber pacing on systolic and diastolic function in patients with hypertrophic cardiomyopathy. *J Am Coll Cardiol* 1996;27(2):421–30.
- [59] Topilski I, Sherez J, Keren G, Copperman I. Long-Term Effects of Dual-Chamber Pacing With Periodic Echocardiographic Evaluation of Optimal Atrioventricular Delay in Patients With Hypertrophic Cardiomyopathy > 50 Years of Age. *Am J Cardiol* 2006;97(12):1769–75.
- [60] Moss AJ, Hall WJ, Cannom DS, et al. Cardiac-Resynchronization Therapy for the Prevention of Heart-Failure Events. *N Engl J Med* 2009;361(14):1329–38.
- [61] Tang ASL, Wells GA, Talajic M, et al. Cardiac-Resynchronization Therapy for Mild-to-Moderate Heart Failure. *N Engl J Med* 2010;363(25):2385–95.
- [62] Chen S, Ling Z, Kiuchi MG, Yin Y, Krucoff MW. The efficacy and safety of cardiac resynchronization therapy combined with implantable cardioverter defibrillator for heart failure: a meta-analysis of 5674 patients. *Europace* 2013;15(7):992–1001.
- [63] Koplan BA, Kaplan AJ, Weiner S, Jones PW, Seth M, Christman SA. Heart Failure Decompensation and All-Cause Mortality in Relation to Percent Biventricular Pacing in Patients With Heart Failure. *J Am Coll Cardiol* 2009;53(4):355–60.
- [64] Hayes DL, Boehmer JP, Day JD, et al. Cardiac resynchronization therapy and the relationship of percent biventricular pacing to symptoms and survival. *Heart Rhythm* 2011;8(9):1469–75.
- [65] Gasparini M, Auricchio A, Regoli F, et al. Four-Year Efficacy of Cardiac Resynchronization Therapy on Exercise Tolerance and Disease Progression. *J Am Coll Cardiol* 2006;48(4):734–43.
- [66] Ousdigian KT, Borek PP, Koehler JL, Heywood JT, Ziegler PD, Wilkoff BL. The Epidemic of Inadequate Biventricular Pacing in Patients With Persistent or Permanent Atrial Fibrillation and Its Association With Mortality. *Circulation: Arrhythmia Electrophysiol* 2014;7(3):370–6.
- [67] Ruwald AC, Kutyla V, Ruwald MH, et al. The association between biventricular pacing and cardiac resynchronization therapy-defibrillator efficacy when compared with implantable cardioverter defibrillator on outcomes and reverse remodelling. *Eur Heart J* 2014;36(7):440–8.
- [68] Mullens W, Grimm RA, Verga T, et al. Insights From a Cardiac Resynchronization Optimization Clinic as Part of a Heart Failure Disease Management Program. *J Am Coll Cardiol* 2009;53(9):765–73.
- [69] Ellenbogen KA, Gold MR, Meyer TE, et al. Primary Results From the SmartDelay Determined AV Optimization: A Comparison to Other AV Delay Methods Used in Cardiac Resynchronization Therapy (SMART-AV) Trial: A Randomized Trial Comparing Empirical, Echocardiography-Guided, and Algorithmic Atrioventricular Delay Programming in Cardiac Resynchronization Therapy. *Circulation* 2010;122(25):2660–8.
- [70] Abraham WT, Gras D, Yu CM, Guzzo L, Gupta MS. Rationale and design of a randomized clinical trial to assess the safety and efficacy of frequent optimization of cardiac resynchronization therapy: The Frequent Optimization Study Using the QuickOpt Method (FREEDOM) trial. *Am Heart J* 2010;159(6):944–8.
- [71] Bax JJ, Gorcsan J. Echocardiography and Noninvasive Imaging in Cardiac Resynchronization Therapy. *J Am Coll Cardiol* 2009;53(21):1933–43.
- [72] Thibault B, Ducharme A, Harel F, et al. Left ventricular versus simultaneous biventricular pacing in patients with heart failure and a QRS complex > / = 120 milliseconds. *Circulation* 2011;124(25):2874–81 Dec 20.
- [73] Martin DO, Lemke B, Bernie D, et al. Investigation of a novel algorithm for synchronized left-ventricular pacing and ambulatory optimization of cardiac resynchronization therapy: Results of the adaptive CRT trial. *Heart Rhythm* 2012;9(11):1807–14.
- [74] Ritter P, Delnoy PPH, Padeletti L, et al. A randomized pilot study of optimization of cardiac resynchronization therapy in sinus rhythm patients using a peak endocardial acceleration sensor vs. standard methods. *Europace* 2012;14(9):1324–33.
- [75] Lown B. Implanted Standby Defibrillators. *Circ* 1972;46:637.
- [76] Mirowski M, Reid PR, Mower MM, et al. Termination of malignant ventricular arrhythmias with an implanted automatic defibrillator in human beings. *N Engl J Med* 1980;303(6):322–4 Aug 7.
- [77] Gradaus R, Bode-Schnurbus L, Weber M, et al. Effect of ventricular fibrillation duration on the defibrillation threshold in humans. *Pacing Clin Electrophysiol* 2002;25(1):14–9.
- [78] Moss AJ, Zareba W, Hall WJ, et al. Prophylactic implantation of a defibrillator in patients with myocardial infarction and reduced ejection fraction. *N Engl J Med* 2002;346(12):877–83 Mar 21.
- [79] Moss AJ, Hall WJ, Cannom DS, et al. Improved Survival with an Implanted Defibrillator in Patients with Coronary Disease at High Risk for Ventricular Arrhythmia. *N Engl J Med* 1996;335(26):1933–40.
- [80] Bardy GH, Lee KL, Mark DB, et al. Amiodarone or an Implantable Cardioverter-Defibrillator for Congestive Heart Failure. *N Engl J Med* 2005;352(3):225–37.
- [81] Klein RC, Raitt MH, Wilkoff BL, et al. Analysis of implantable cardioverter defibrillator therapy in the Antiarrhythmics Versus Implantable Defibrillators (AVID) Trial. *J Cardiovasc Electrophysiol* 2003;14(9):940–8.
- [82] Daubert JP, Zareba W, Cannom DS, et al. Inappropriate implantable cardioverter-defibrillator shocks in MADIT II: frequency, mechanisms, predictors, and survival impact. *J Am Coll Cardiol* 2008;51(14):1357–65 Apr 8.
- [83] Poole JE, Johnson GW, Hellkamp AS, et al. Prognostic importance of defibrillator shocks in patients with heart failure. *N Engl J Med* 2008;359(10):1009–17 Sep 4.
- [84] Grimm W, Flores BT, Marchlinski FE. Shock occurrence and survival in 241 patients with implantable cardioverter-defibrillator therapy. *Circulation* 1993;87(6):1880–8.
- [85] Luderitz B, Jung W, Deister A, Marneros A, Manz M. Patient acceptance of the implantable cardioverter defibrillator in ventricular tachyarrhythmias. *Pacing Clin Electrophysiol* 1993;16(9):1815–21.
- [86] Rosenqvist M, Beyer T, Block M, den Dulk K, Minten J, Lindemans F. Adverse Events With Transvenous Implantable Cardioverter-Defibrillators: A Prospective Multicenter Study. *Circ* 1998;98(7):663–70.
- [87] Gold MR, Peters RW, Johnson JW, Shorofsky SR. Complications Associated with Pectoral Implantation of Cardioverter Defibrillators. *Pacing Clin Electrophysiol* 1997;20(1):208–11.
- [88] Kron J. Clinical significance of device-related complications in clinical trials and implications for future trials: Insights from the Antiarrhythmics Versus Implantable Defibrillators (AVID) trial. *Card Electrophysiol Rev Cardiovasc Med* 2003;7:473–8.
- [89] Schron EB, Exner DV, Yao Q, et al. Quality of life in the antiarrhythmics versus implantable defibrillators trial: impact of therapy and influence of adverse symptoms and defibrillator shocks. *Circulation* 2002;105(5):589–94.
- [90] Vollmann D, Luthje L, Vonhof S, Unterberg C. Inappropriate therapy and fatal proarrhythmia by an implantable cardioverter-defibrillator. *Heart Rhythm* 2005;2(3):307–9.
- [91] Sweeney MO, Wathen MS, Volosin K, et al. Appropriate and Inappropriate Ventricular Therapies, Quality of Life, and Mortality Among Primary and Secondary Prevention Implantable Cardioverter Defibrillator Patients: Results From the Pacing Fast VT Reduces Shock Therapies (PainFREE Rx II) Trial. *Circ* 2005;111(22):2898–905.
- [92] Pinski SL, Fahy GJ. The proarrhythmic potential of implantable cardioverter-defibrillators. *Circ* 1995;92(6):1651–64.
- [93] Wathen MS, DeGroot PJ, Sweeney MO, et al. Prospective randomized multicenter trial of empirical antitachycardia pacing versus shocks for spontaneous rapid ventricular tachycardia in patients with implantable cardioverter-defibrillators: Pacing Fast Ventricular Tachycardia Reduces Shock Therapies (PainFREE Rx II) trial results. *Circulation* 2004;110(17):2591–6 Oct 26.
- [94] Wilkoff BL, Williamson BD, Stern RS, et al. Strategic programming of detection and therapy parameters in implantable cardioverter-defibrillators reduces shocks in primary prevention patients: results from the PREPARE (Primary Prevention Parameters Evaluation) study. *J Am Coll Cardiol* 2008;52(7):541–50 Aug 12.
- [95] A simplified biventricular defibrillator with fixed long detection intervals reduces implantable cardioverter defibrillator (ICD) interventions and heart failure hospitalizations in patients with non-ischaemic cardiomyopathy implanted for primary prevention: the RELEVANT [Role of long dEtection window programming in patients with Left Ventricular dysfunction, Non-ischemic eTiology in primary prevention treated with a biventricular ICD] study. In: Gasparini M, Menozzi C, Proclemer A, et al., editors. *Eur Heart J*, 30; 2009, p. 2758–67.
- [96] Moss AJ, Schuger C, Beck CA, et al. Reduction in Inappropriate Therapy and Mortality through ICD Programming. *New Engl J Med* 2012;367(24):2275–83.

- [97] Gasparini M, Proclemer A, Klersy C, et al. Effect of long-detection interval vs standard-detection interval for implantable cardioverter-defibrillators on antitachycardia pacing and shock delivery: the ADVANCE III randomized clinical trial. *JAMA* 2013;309(18):1903–11 May 8.
- [98] Saeed M, Hanna I, Robotis D, et al. Programming implantable cardioverter-defibrillators in patients with primary prevention indication to prolong time to first shock: results from the PROVIDE study. *J Cardiovasc Electrophysiol* 2014;25(1):52–9.
- [99] Kloppe A, Proclemer A, Arenal A, et al. Efficacy of long detection interval ICD settings in secondary prevention population: Data from the Advance III trial. *Circulation* 2014;130:308–14.
- [100] Tan VH, Wilton SB, Kuriachan V, Sumner GL, Exner DV. Impact of Programming Strategies Aimed at Reducing Nonessential Implantable Cardioverter Defibrillator Therapies on Mortality: A Systematic Review and Meta-Analysis. *Circulation: Arrhythmia Electrophysiol* 2014;7(1):164–70.
- [101] Wilkoff BL, Ousdigian KT, Sterns LD, et al. A comparison of empiric to physician-tailored programming of implantable cardioverter-defibrillators: results from the prospective randomized multicenter EMPERIC trial. *J Am Coll Cardiol* 2006;48(2):330–9 Jul 18.
- [102] Scott PA, Silberbauer J, McDonagh TA, Murgatroyd FD. The Impact of Prolonged ICD Arrhythmia Detection Times On Outcomes: A Meta-analysis. *Heart Rhythm* 2014;11:828–35.
- [103] Bansch D, Steffgen F, Gronefeld G, et al. The 1+1 trial: a prospective trial of a dual- versus a single-chamber implantable defibrillator in patients with slow ventricular tachycardias. *Circulation* 2004;110(9):1022–9.
- [104] Clementy N, Pierre B, Lallemand B, et al. Long-term follow-up on high-rate cut-off programming for implantable cardioverter defibrillators in primary prevention patients with left ventricular systolic dysfunction. *Europace: Eur pacing, Arrhythm, Card Electrophysiol: J Work groups Card pacing, Arrhythm, Card Cell Electrophysiol Eur Soc Cardiol* 2012;14(7):968–74 Jul.
- [105] Kloppe A, Proclemer A, Arenal A, et al. Efficacy of long detection interval implantable cardioverter-defibrillator settings in secondary prevention population: data from the Avoid Delivering Therapies for Nonsustained Arrhythmias in ICD Patients III (ADVANCE III) trial. *Circulation* 2014;130(4):308–14.
- [106] Webber MR, Stiles MK. Recommendations for the programming of implantable cardioverter-defibrillators in New Zealand. *Heart Lung Circ* 2012;21(12):765–77.
- [107] Gard JJ, Friedman PA. Strategies to Reduce ICD Shocks: The Role of Supraventricular Tachycardia-Ventricular Tachycardia Discriminators. *Card Electrophysiol Clin* 2011;3(3):373–87.
- [108] Aliot E, Nitzsche R, Ripart A. Arrhythmia detection by dual-chamber implantable cardioverter defibrillators. A review of current algorithms. *Eur Jol* 2004;6(4):273–86.
- [109] Koneru JN, Swerdlow CD, Wood MA, Ellenbogen KA. Minimizing inappropriate or “unnecessary” implantable cardioverter-defibrillator shocks: appropriate programming. *Circ Arrhythm Electrophysiol* 2011;4(5):778–90.
- [110] Mansour F, Khairy P. Programming ICDs in the Modern Era beyond Out-of-the-Box Settings. *Pacing Clin Electrophysiol* 2011;34(4):506–20.
- [111] Swerdlow C, Gillberg J, Khairy P. Sensing and detection. In: Ellenbogen K, Kay G, Lau C, Wilkoff B, editors. *Clinical Cardiac Pacing, Defibrillation and Resynchronization Therapy*. 4th ed. Philadelphia: W.B. Saunders Company; 2011. p. 56–126.
- [112] Brugada J, Mont L, Figueiredo M, Valentino M, Matas M, Navarro-Lopez F. Enhanced detection criteria in implantable defibrillators. *J Cardiovasc Electrophysiol* 1998;9(3):261–8.
- [113] Weber M, Bocker D, Bansch D, et al. Efficacy and safety of the initial use of stability and onset criteria in implantable cardioverter defibrillators. *J Cardiovasc Electrophysiol* 1999;10(2):145–53.
- [114] Swerdlow CD, Ahern T, Chen PS, et al. Underdetection of ventricular tachycardia by algorithms to enhance specificity in a tiered-therapy cardioverter-defibrillator. *J Am Coll Cardiol* 1994;24(2):416–24.
- [115] Swerdlow CD, Chen PS, Kass RM, Allard JR, Peter CT. Discrimination of ventricular tachycardia from sinus tachycardia and atrial fibrillation in a tiered-therapy cardioverter-defibrillator. *J Am Coll Cardiol* 1994;23(6):1342–55.
- [116] Stadler RW, Gunderson BD, Gillberg JM. An adaptive interval-based algorithm for withholding ICD therapy during sinus tachycardia. *Pacing Clin Electrophysiol* 2003;26(5):1189–201.
- [117] Neuzner J, Pitschner HF, Schlepper M. Programmable VT detection enhancements in implantable cardioverter defibrillator therapy. *Pacing Clin Electrophysiol* 1995;18(3 Pt 2):539–47.
- [118] Dorian P, Philippon F, Thibault B, et al. Randomized controlled study of detection enhancements versus rate-only detection to prevent inappropriate therapy in a dual-chamber implantable cardioverter-defibrillator. *Heart Rhythm* 2004;1(5):540–7.
- [119] Requena-Carrion J, Vaisanen J, Alonso-Atienza F, Garcia-Alberola A, Ramos-Lopez FJ, Rojo-Alvarez JL. Sensitivity and spatial resolution of transvenous leads in implantable cardioverter defibrillator. *IEEE Trans Biomed Eng* 2009;56(12):2773–81.
- [120] Swerdlow CD, Brown ML, Lurie K, et al. Discrimination of ventricular tachycardia from supraventricular tachycardia by a downloaded wavelet-transform morphology algorithm: a paradigm for development of implantable cardioverter defibrillator detection algorithms. *J Cardiovasc Electrophysiol* May 2002;13(5):432–41.
- [121] Theuns DA, Rivero-Ayerza M, Goedhart DM, van der Perk R, Jordaens LJ. Evaluation of morphology discrimination for ventricular tachycardia diagnosis in implantable cardioverter-defibrillators. *Heart Rhythm* Nov 2006;3(11):1332–8.
- [122] Klein GJ, Gillberg JM, Tang A, et al. Improving SVT discrimination in single-chamber ICDs: A new electrogram morphology-based algorithm. *J Cardiovasc Electrophysiol* 2006;17(12):1310–9.
- [123] Swerdlow CD, Friedman PA. Advanced ICD troubleshooting: Part I. *Pacing Clin Electrophysiol* 2005;28(12):1322–46.
- [124] Wilkoff BL, Kuhlkamp V, Volosin K, et al. Critical analysis of dual-chamber implantable cardioverter-defibrillator arrhythmia detection: results and technical considerations. *Circulation* 2001;103(3):381–6.
- [125] Glikson M, Swerdlow CD, Gurevitz OT, et al. Optimal combination of discriminators for differentiating ventricular from supraventricular tachycardia by dual-chamber defibrillators. *J Cardiovasc Electrophysiol* 2005;16(7):732–9.
- [126] Moss AJ, Schuger C, Beck CA, et al. Reduction in inappropriate therapy and mortality through ICD programming. *N Engl J Med* 2012;367(24):2275–83.
- [127] Meijer A, Auricchio A, Kurita T, et al. Inappropriate shock rates in patients with single chamber ICDs using a novel suite of detection algorithms. *Europace*. 152013: ii116-ii117.
- [128] Grimm W, Flores BF, Marchlinski FE. Electrocardiographically documented unnecessary, spontaneous shocks in 241 patients with implantable cardioverter defibrillators. *Pacing Clin Electrophysiol* 1992;15(11):1667–73.
- [129] Gilliam FR, Hayes DL, Boehmer JP, et al. Real world evaluation of dual-zone ICD and CRT-D programming compared to single-zone programming: the ALTITUDE REDUCES study. *J Cardiovasc Electrophysiol* 2011;22(9):1023–9.
- [130] Fischer A, Ousdigian KT, Johnson JW, Gillberg JM, Wilkoff BL. The impact of atrial fibrillation with rapid ventricular rates and device programming on shocks in 106,513 ICD and CRT-D patients. *Heart Rhythm* 2012;9(1):24–31.
- [131] Volosin KJ, Exner DV, Wathen MS, Sherfesele L, Scinicariello AP, Gillberg JM. Combining shock reduction strategies to enhance ICD therapy: a role for computer modeling. *J Cardiovasc Electrophysiol* 2011;22(3):280–9.
- [132] Nanthakumar K, Dorian P, Paquette M, et al. Is inappropriate implantable defibrillator shock therapy predictable? *J Interv Card Electrophysiol* 2003;8(3):215–20.
- [133] Wilkoff BL, Williamson BD, Stern RS, et al. Strategic programming of detection and therapy parameters in implantable cardioverter-defibrillators reduces shocks in primary prevention patients: results from the PREPARE (Primary Prevention Parameters Evaluation) study. *J Am Coll Cardiol* 2008;52(7):541–50.
- [134] Weiss R, Knight BP, Gold MR, et al. Safety and efficacy of a totally subcutaneous implantable-cardioverter defibrillator. *Circulation* 2013;128(9):944–53.
- [135] Sullivan RM, Russo AM, Berg KC, et al. Arrhythmia rate distribution and tachyarrhythmia therapy in an ICD population: results from the INTRINSIC RV trial. *Heart Rhythm* 2012;9(3):351–8.
- [136] Auricchio A, Schloss EJ, Kurita T, et al. Low inappropriate shock rates in patients with single- and dual/triple-chamber implantable cardioverter-defibrillators using a novel suite of detection algorithms: PainFree SST trial primary results. *Heart Rhythm* 2015;12(5):926–36.
- [137] Friedman PA, McClelland RL, Bamlet WR, et al. Dual-Chamber Versus Single-Chamber Detection Enhancements for Implantable Defibrillator Rhythm Diagnosis: The Detect Supraventricular Tachycardia Study. *Circulation* 2006;113(25):2871–9.
- [138] Almendral J, Arribas F, Wolpert C, et al. Dual-chamber defibrillators reduce clinically significant adverse events compared with single-chamber devices: results from the DATAS (Dual chamber and Atrial Tachyarrhythmias Adverse events Study) trial. *Europace: Eur pacing, Arrhythm, Card Electrophysiol: J Work groups Card pacing, Arrhythm, Card Cell Electrophysiol Eur Soc Cardiol* 2008;10(5):528–35.
- [139] Friedman PA, Slusser J, Hodge DO, et al. Prospective Randomized Trial Of Dual Chamber Vs. Single Chamber ICDs To Minimize Shocks In Optimally Programmed Devices. *Heart Rhythm* 2012;9(9):1583.
- [140] Goncalves J, Pereira T. Inappropriate shocks in patients with ICDs: single chamber versus dual chamber. *Arq Bras Cardiol* 2013;101(2):141–8.
- [141] Kolb C, Sturmer M, Sick P, et al. Reduced risk for inappropriate implantable cardioverter-defibrillator shocks with dual-chamber therapy compared with single-chamber therapy: results of the randomized OPTION study. *JACC Heart Fail* 2014;2(6):611–9.
- [142] Powell BD, Asirvatham SJ, Perschbacher DL, et al. Noise, Artifact, and Oversensing Related Inappropriate ICD Shock Evaluation: ALTITUDE NOISE Study. *Pacing Clin Electrophysiol* 2012;35(7):863–9.
- [143] Sears SF, Hauf JD, Kirian K, Hazelton G, Conti JB. Posttraumatic stress and the implantable cardioverter-defibrillator patient: what the electrophysiologist needs to know. *Circ Arrhythm Electrophysiol* 2011;4(2):242–50.
- [144] Swerdlow CD, Gunderson BD, Ousdigian KT, et al. Downloadable algorithm to reduce inappropriate shocks caused by fractures of implantable cardioverter-defibrillator leads. *Circulation* 2008;118(21):2122–9.
- [145] Ellenbogen KA, Gunderson BD, Stromberg KD, Swerdlow CD. Performance of Lead Integrity Alert to assist in the clinical diagnosis of implantable cardioverter defibrillator lead failures: analysis of different implantable cardioverter defibrillator leads. *Circ Arrhythm Electrophysiol* 2013;6(6):1169–77.
- [146] Cao J, Gillberg JM, Swerdlow CD. A fully automatic, implantable cardioverter-defibrillator algorithm to prevent inappropriate detection of ventricular tachycardia or fibrillation due to T-wave oversensing in spontaneous rhythm. *Heart Rhythm* 2012;9(4):522–30.
- [147] Swerdlow CD, Sachanandani H, Gunderson BD, Ousdigian KT, Hjelte M, Ellenbogen KA. Preventing overdiagnosis of implantable cardioverter-

- defibrillator lead fractures using device diagnostics. *J Am Coll Cardiol* 2011;57(23):2330–9 Jun 7.
- [148] Swerdlow CD, Ellenbogen KA. Implantable cardioverter-defibrillator leads: design, diagnostics, and management. *Circulation* 2013;128(18):2062–71 2061–2069.
- [149] Swerdlow CD, Gunderson BD, Ousdigian KT, Abeyratne A, Sachanandani H, Ellenbogen KA. Downloadable software algorithm reduces inappropriate shocks caused by implantable cardioverter-defibrillator lead fractures: a prospective study. *Circulation* 2010;122(15):1449–55.
- [150] Blanck Z, Axtell K, Brodhagen K, et al. Inappropriate shocks in patients with Fidelis(R) lead fractures: impact of remote monitoring and the lead integrity algorithm. *J Cardiovasc Electrophysiol* 2011;22(10):1107–14.
- [151] Gunderson BD, Gillberg JM, Wood MA, Vijayaraman P, Shepard RK, Ellenbogen KA. Development and testing of an algorithm to detect implantable cardioverter-defibrillator lead failure. *Heart Rhythm* 2006;3(2):155–62.
- [152] .*Sprint Fidelis[®] Lead Patient Management Recommendations Update: Models 6949, 6948, 6931, 6930*. 2011; (http://www.medtronic.com/wcm/groups/mdtcom_sg/@mdt/documents/documents/fidelis-phys-ltr-2011-04.pdf). Accessed 4/21/2013, 2013.
- [153] Gold MR, Theuns DA, Knight BP, et al. Head-to-head comparison of arrhythmia discrimination performance of subcutaneous and transvenous ICD arrhythmia detection algorithms: the START study. *J Cardiovasc Electrophysiol* 2012;23(4):359–66.
- [154] Gold MR, Weiss R, Theuns DA, et al. Use of a discrimination algorithm to reduce inappropriate shocks with a subcutaneous implantable cardioverter-defibrillator. *Heart Rhythm* 2014;11(8):1352–8.
- [155] Sweeney MO, Sherfese L, DeGroot PJ, Wathen MS, Wilkoff BL. Differences in effects of electrical therapy type for ventricular arrhythmias on mortality in implantable cardioverter-defibrillator patients. *Heart Rhythm* 2010;7(3):353–60.
- [156] Sood N, Ruwald ACH, Solomon S, et al. Association between myocardial substrate, implantable cardioverter defibrillator shocks and mortality in MADIT- CRT. *Eur Heart J* 2013;35(2):106–15.
- [157] Powell BD, Saxon LA, Boehmer JP, et al. Survival After Shock Therapy in Implantable Cardioverter-Defibrillator and Cardiac Resynchronization Therapy- Defibrillator Recipients According to Rhythm Shocked. *J Am Coll Cardiol* 2013;62(18):1674–9.
- [158] van Rees JB, Borleffs CJW, de Bie MK, et al. Inappropriate Implantable Cardioverter-Defibrillator Shocks. *J Am Coll Cardiol* 2011;57(5):556–62.
- [159] Gasparini M, Proclemer A, Klersy C, et al. Effect of Long-Detection Interval vs Standard-Detection Interval for Implantable Cardioverter-Defibrillators on Antitachycardia Pacing and Shock Delivery. *JAMA* 2013;309(18):1903.
- [160] Gonz/Eliz-Enriquez S, Rodriguez-Entem F, Expósito V, et al. Single-chamber ICD, single-zone therapy in primary and secondary prevention patients: the simpler the better? *J Interv Card Electrophysiol* 2012;35(3):343–9.
- [161] Saeed M, Neason CG, Razavi M, et al. Programming Antitachycardia Pacing for Primary Prevention in Patients With Implantable Cardioverter Defibrillators: Results From the PROVE Trial. *J Cardiovasc Electrophysiol* 2010;21(12):1349–54.
- [162] Sivagangabalan G, Eshoo S, Eipper VE, Thiagalangam A, Kovoov P. Discriminatory Therapy for Very Fast Ventricular Tachycardia in Patients with Implantable Cardioverter Defibrillators. *Pacing Clin Electrophysiol* 2008;31(9):1095–9.
- [163] Schoels W, Steinhaus D, Johnson WB, et al. Optimizing implantable cardioverter-defibrillator treatment of rapid ventricular tachycardia: Antitachycardia pacing therapy during charging. *Heart Rhythm* 2007;4(7):879–85.
- [164] Wathen MS, Sweeney MO, DeGroot PJ, et al. Shock Reduction Using Antitachycardia Pacing for Spontaneous Rapid Ventricular Tachycardia in Patients With Coronary Artery Disease. *Circulation* 2001;104(7):796–801.
- [165] Gasparini M, Anselme F, Clementy J, et al. BiVentricular versus right ventricular antitachycardia pacing to terminate ventricular tachyarrhythmias in patients receiving cardiac resynchronization therapy: The ADVANCE CRT-D Trial. *Am Heart J* 2010;159(6):1116–23 e1112.
- [166] Yee R, Birgersdotter-Green U, Belk P, Jackson T, Christensen J, Wathen MS. The Relationship between Pacing Site and Induction or Termination of Sustained Monomorphic Ventricular Tachycardia by Antitachycardia Pacing. *Pacing Clin Electrophysiol* 2010;33(1):27–32.
- [167] Anguera I, Dallaglio P, Sabaté X, et al. The Benefit of a Second Burst Antitachycardia Sequence for Fast Ventricular Tachycardia in Patients with Implantable Cardioverter Defibrillators. *Pacing Clin Electrophysiol* 2013;37(4):486–94.
- [168] Martins RP, Blangy H, Muresan L, et al. Safety and efficacy of programming a high number of antitachycardia pacing attempts for fast ventricular tachycardia: a prospective study. *Europace* 2012;14(10):1457–64.
- [169] Santini M, Lunati M, Defaye P, et al. Prospective multicenter randomized trial of fast ventricular tachycardia termination by prolonged versus conventional anti-tachyarrhythmia burst pacing in implantable cardioverter-defibrillator patients-Atp DeliVery for pAiNless ICD thErapy (ADVANCE-D) Trial results. *J Interv Card Electrophysiol* 2010;27(2):127–35.
- [170] Gulizia MM, Piraino L, Scherillo M, et al. A Randomized Study to Compare Ramp Versus Burst Antitachycardia Pacing Therapies to Treat Fast Ventricular Tachyarrhythmias in Patients With Implantable Cardioverter Defibrillators: The PITAGORA ICD Trial. *Circulation: Arrhythmia Electrophysiol* 2009;2(2):146–53.
- [171] Sivagangabalan G, Chik W, Zaman S, et al. Antitachycardia Pacing for Very Fast Ventricular Tachycardia and Low-Energy Shock for Ventricular Arrhythmias in Patients With Implantable Defibrillators. *Am J Cardiol* 2013;112(8):1153–7.
- [172] Kanal E, Barkovich AJ, Bell C, et al. ACR guidance document on MR safe practices: 2013. *J Magn Reson Imaging* 2013;37(3):501–30.
- [173] Guðon-Moreau L, Kouakam C, Klug D, et al. Decreased Delivery of Inappropriate Shocks Achieved by Remote Monitoring of ICD: A Substudy of the ECOST Trial. *J Cardiovasc Electrophysiol* 2014;25(7):763–70.
- [174] Camm AJ, Sears SF, Todaro JF, Lewis TS, Sotile W, Conti JB. Examining the psychosocial impact of implantable cardioverter defibrillators: A literature review. *Clin Cardiol* 1999;22(7):481–9.
- [175] Sears SF, Vazquez LD, Matchett M, Pitzalis M. State-of-the-art: anxiety management in patients with implantable cardioverter defibrillators. *Stress Health* 2008;24(3):239–48.
- [176] Magyar-Russell G, Thombs BD, Cai JX, et al. The prevalence of anxiety and depression in adults with implantable cardioverter defibrillators: A systematic review. *J Psychosom Res* 2011;71(4):223–31.
- [177] Kraaier K, Starrenburg AH, Verheggen RM, van der Palen J, Scholten MF. Incidence and predictors of phantom shocks in implantable cardioverter defibrillator recipients. *Neth Heart J* 2012;21(4):191–5.
- [178] Guðon-Moreau L, Lacroix D, Sadoul N, et al. A randomized study of remote follow-up of implantable cardioverter defibrillators: safety and efficacy report of the ECOST trial. *Eur Heart J* 2013;34(8):605–14.
- [179] Ruwald AC, Schuger C, Moss AJ, et al. Mortality Reduction in Relation to Implantable Cardioverter Defibrillator Programming in the Multicenter Automatic Defibrillator Implantation Trial-Reduce Inappropriate Therapy (MADIT-RIT). *Circulation: Arrhythmia Electrophysiol* 2014;7(5):785–92.
- [180] A Comparison of Antiarrhythmic-Drug Therapy with Implantable Defibrillators in Patients Resuscitated from Near-Fatal Ventricular Arrhythmias. *N Engl J Med* 1997;337(22):1576–1584.
- [181] Connolly S. Meta-analysis of the implantable cardioverter defibrillator secondary prevention trials. *Eur Heart J* 2000;21(24):2071–8.
- [182] Moss AJ, Cannom DS, Daubert JP, et al. Multicenter Automatic Defibrillator Implantation Trial II (MADIT II): Design and Clinical Protocol. *Ann Non-invasive Electrocardiol* 1999;4(1):83–91.
- [183] Buxton AE, Lee KL, Fisher JD, Josephson ME, Prystowsky EN, Hailey G. A Randomized Study of the Prevention of Sudden Death in Patients with Coronary Artery Disease. *N Engl J Med* 1999;341(25):1882–90.
- [184] Kadish A, Dyer A, Daubert JP, et al. Prophylactic Defibrillator Implantation in Patients with Nonischemic Dilated Cardiomyopathy. *N Engl J Med* 2004;350(21):2151–8.
- [185] Bardy GH, Lee KL, Mark DB, et al. Amiodarone or an Implantable Cardioverter-Defibrillator for Congestive Heart Failure. *N Engl J Med* 2005;352(3):225–37 January 20.
- [186] Neuzner J, Pitschner HF, Huth C, Schlepper M. Effect of Biphasic Waveform Pulse on Endocardial Defibrillation Efficacy in Humans. *Pacing Clin Electrophysiol* 1994;17(2):207–12.
- [187] Gold MR, Foster AH, Shorofsky SR. Effects of an active pectoral-pulse generator shell on defibrillation efficacy with a transvenous lead system. *Am J Cardiol* 1996;78(5):540–3.
- [188] Bardy GH, Yee R, Jung W. Multicenter experience with a pectoral unipolar implantable cardioverter-defibrillator. *J Am Coll Cardiol* 1996;28(2):400–10.
- [189] Natale A, Sra J, Axtell K, et al. Preliminary experience with a hybrid nonthoracotomy defibrillating system that includes a biphasic device: Comparison with a standard monophasic device using the same lead system. *J Am Coll Cardiol* 1994;24(2):406–12.
- [190] Swerdlow CD. Implantation of Cardioverter Defibrillators Without Induction of Ventricular Fibrillation. *Circulation* 2001;103(17):2159–64.
- [191] Russo AM, Sauer W, Gerstenfeld EP, et al. Defibrillation threshold testing: Is it really necessary at the time of implantable cardioverter-defibrillator insertion? *Heart Rhythm* 2005;2(5):456–61.
- [192] Kolb C, Tzeis S, Zrenner B. Defibrillation Threshold Testing: Tradition or Necessity? *Pacing Clin Electrophysiol* 2009;32(5):570–2.
- [193] Healey JS, Dorian P, Mitchell LB, et al. Canadian Registry of ICD Implant Testing Procedures (CREDIT): Current Practice, Risks, and Costs of Intraoperative Defibrillation Testing. *J Cardiovasc Electrophysiol* 2010;21(2):177–82.
- [194] Vischer AS, Sticherling C, KÜhne MS, Osswald S, Schaer BA. Role of Defibrillation Threshold Testing in the Contemporary Defibrillator Patient Population. *J Cardiovasc Electrophysiol* 2012;24(4):437–41.
- [195] Leong-Sit P, Gula LJ, Diamantouros P, et al. Effect of defibrillation testing on management during implantable cardioverter-defibrillator implantation. *Am Heart J* 2006;152(6):1104–8.
- [196] Kremers MS, Hammill SC, Berul CI, et al. The National ICD Registry Report: Version 2.1 including leads and pediatrics for years 2010 and 2011. *Heart Rhythm* 2013;10(4):e59–65.
- [197] de Vries JW, Bakker PFA, Visser GH, Diephuis JC, van Huffelen AC. Changes in Cerebral Oxygen Uptake and Cerebral Electrical Activity During Defibrillation Threshold Testing. *Anesth Analg* 1998;87(1):16–20.
- [198] Singer I, Edmonds H. Changes in cerebral perfusion during third-generation implantable cardioverter defibrillator testing. *Am Heart J* 1994;127(4):1052–7.
- [199] Vroems EM, Bakker PFA, De Vries JW, Wieneke GH, Van Huffelen AC. The impact of repeated short episodes of circulatory arrest on cerebral function. Reassuring electroencephalographic (EEG) findings during defibrillation threshold testing at defibrillator implantation. *Electroencephalogr Clin Neurophysiol* 1996;98(4):236–42.

- [200] da Silva MP, Rivetti LA, LAST Mathias, Cagno G, Matsui C. Impact of Induced Cardiac Arrest on Cognitive Function after Implantation of a Cardioverter-Defibrillator. *Braz J Anesthesiol* 2009;59(1):37–45.
- [201] Karaoguz R, Altin T, Atbasoglu EC, et al. Defibrillation Testing and Early Neurologic Outcome. *Int Heart J* 2008;49(5):553–63.
- [202] Schluter T, Baum H, Plewan A, Neumeier D. Effects of implantable cardioverter defibrillator implantation and shock application on biochemical markers of myocardial damage. *Clin Chem* 2001;47:459–63.
- [203] Hasdemir CAN, Shah N, Rao AP, et al. Analysis of Troponin I Levels After Spontaneous Implantable Cardioverter Defibrillator Shocks. *J Cardiovasc Electrophysiol* 2002;13(2):144–50.
- [204] Hurst TM, Hinrichs M, Breidenbach C, Katz N, Waldecker B. Detection of myocardial injury during transvenous implantation of automatic cardioverter-defibrillators. *J Am Coll Cardiol* 1999;34(2):402–8.
- [205] Frame R, Brodman R, Furman S, et al. Clinical Evaluation of the Safety of Repetitive Intraoperative Defibrillation Threshold Testing. *Pacing Clin Electrophysiol* 1992;15(6):870–7.
- [206] Toh N, Nishii N, Nakamura K, et al. Cardiac Dysfunction and Prolonged Hemodynamic Deterioration After Implantable Cardioverter-Defibrillator Shock in Patients With Systolic Heart Failure. *Circulation: Arrhythmia Electrophysiol* 2012;5(5):898–905.
- [207] Semmler V, Biermann J, Haller B, et al. ICD Shock, Not Ventricular Fibrillation, Causes Elevation of High Sensitive Troponin T after Defibrillation Threshold Testing—The Prospective, Randomized, Multicentre TropShock-Trial. *Plos One* 2015;10(7):e0131570.
- [208] Mollerus M, Naslund L. Myocardial stunning following defibrillation threshold testing. *J Interv Card Electrophysiol* 2007;19(3):213–6.
- [209] Frame R, Brodman R, Gross JAY, et al. Initial Experience with Transvenous Implantable Cardioverter Defibrillator Lead Systems: Operative Morbidity and Mortality. *Pacing Clin Electrophysiol* 1993;16(1):149–52.
- [210] Cardiac Arrhythmias. *J Am Coll Cardiol* 2008;51(10s1):A1–A34.
- [211] Birnie D, Tung S, Simpson C, et al. Complications associated with defibrillation threshold testing: The Canadian experience. *Heart Rhythm* 2008;5(3):387–90.
- [212] Healey JS, Gula LJ, Birnie DH, et al. A Randomized-Controlled Pilot Study Comparing ICD Implantation with and Without Intraoperative Defibrillation Testing in Patients with Heart Failure and Severe Left Ventricular Dysfunction: A Substudy of the RAFT Trial. *J Cardiovasc Electrophysiol* 2012;23(12):1313–6.
- [213] Brignole M, Occhetta E, Bongiorno MG, et al. Clinical Evaluation of Defibrillation Testing in an Unselected Population of 2,120 Consecutive Patients Undergoing First Implantable Cardioverter-Defibrillator Implant. *J Am Coll Cardiol* 2012;60(11):981–7.
- [214] Russo A, Andriulli J, Ortman M, et al. Outcome following ICD implantation with versus without defibrillation testing: preliminary results of the Prospective Randomized Test-No Test (TNT) Pilot Study. *J Am Coll Cardiol* 2014;63(12):A451.
- [215] Healey JS, Hohnloser SH, Glikson M, et al. Cardioverter defibrillator implantation without induction of ventricular fibrillation: a single-blind, non-inferiority, randomised controlled trial (SIMPLE). *The Lancet* 2015;385(9970):785–91.
- [216] Bansch D, Bonnemeier H, Brandt J, et al. Intra-operative defibrillation testing and clinical shock efficacy in patients with implantable cardioverter-defibrillators: the NORDIC ICD randomized clinical trial. *Eur Heart J* 2015.
- [217] Swerdlow CD, Russo AM, Degroot PJ. The Dilemma of ICD Implant Testing. *Pacing Clin Electrophysiol* 2007;30(5):675–700.
- [218] Smits K, Virag N, Swerdlow CD. Impact of defibrillation testing on predicted ICD shock efficacy: Implications for clinical practice. *Heart Rhythm* 2013;10(5):709–17.
- [219] Healey JS, Brambatti M. Is Defibrillation Testing Necessary for Implantable Transvenous Defibrillators?: Defibrillation Testing Should Not Be Routinely Performed at the Time of Implantable Cardioverter Defibrillator Implantation Circulation: *Arrhythmia Electrophysiol* 2014;7(2):347–51.
- [220] Blatt JA, Poole JE, Johnson GW, et al. No Benefit From Defibrillation Threshold Testing in the SCD-HeFT (Sudden Cardiac Death in Heart Failure Trial). *J Am Coll Cardiol* 2008;52(7):551–6.
- [221] Ellenbogen KA, Wood MA, Stambler BS, Welch WJ, Damiano RJ. Measurement of ventricular electrogram amplitude during intraoperative induction of ventricular tachyarrhythmias. *Am J Cardiol* 1992;70(11):1017–22.
- [222] Baccillieri MS, Gasparini G, Benacchio L, et al. Multicentre comparison of shock efficacy using single-vs. Dual-coil lead systems and Anodal vs. cathodal polarTY defibrillation in patients undergoing transvenous cardioverter-defibrillator implantation. The MODALITY study. *J Interv Card Electrophysiol* 2015;43(1):45–54.
- [223] Russo AM, Wang Y, Al-Khatib SM, Curtis JP, Lampert R. Patient, Physician, and Procedural Factors Influencing the Use of Defibrillation Testing during Initial Implantable Cardioverter Defibrillator Insertion: Findings from the NCDR[®]. *Pacing Clin Electrophysiol* 2013;36(12):1522–31.
- [224] Gold MR. Efficacy and Temporal Stability of Reduced Safety Margins for Ventricular Defibrillation: Primary Results From the Low Energy Safety Study (LESS). *Circulation* 2002;105(17):2043–8.
- [225] Healey JS, Birnie DH, Lee DS, et al. Defibrillation Testing at the Time of ICD Insertion: An Analysis From the Ontario ICD Registry. *J Cardiovasc Electrophysiol* 2010;21(12):1344–8.
- [226] Arnson Y, Suleiman M, Glikson M, et al. Role of defibrillation threshold testing during implantable cardioverter-defibrillator placement: Data from the Israeli ICD Registry. *Heart Rhythm* 2014;11(5):814–21.