

Falsely increased bispectral index values by convective air warming system during kidney transplantation

Se Hun Kim¹, Byeong-Cheol Lee², Yong Han Kim³

ABSTRACT

Bispectral index (BIS) is a reliable parameter for measuring depth of hypnotic level during anesthesia. Convective air warming system is an effective equipment to maintain normothermia during operation. We report falsely elevated BIS value due to convective air warming system while undergoing kidney transplantation.

KEY WORDS: Artifact, Air warmer, Bispectral index, Depth of anesthesia.

doi: <http://dx.doi.org/10.12669/pjms.323.9858>

How to cite this:

Kim SH, Lee BC, Kim YH. Falsely increased bispectral index values by convective air warming system during kidney transplantation. *Pak J Med Sci.* 2016;32(3):793-795. doi: <http://dx.doi.org/10.12669/pjms.323.9858>

This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/3.0>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

INTRODUCTION

Bispectral index (BIS) has been used to measure sedative-hypnotic depth during anesthesia. However, BIS value may be influenced by several agents or movement to lead misinterpretation.¹ Anesthesiologist makes an effort to maintain normal body temperature. Convective air warming system (WarmTouch™, Covidien, USA) is a useful device to prevent hypothermia. WarmTouch™ can produce high value of BIS inaccurately. We report a case of inexact reading with BIS influenced by WarmTouch™ during kidney transplantation.

CASE REPORT

A 38-year-old female patient (height: 153cm, weight: 55kg) was scheduled for kidney transplantation from cadaveric donor. She had been diagnosed with chronic kidney disease about two years ago and on hemodialysis for one year.

Anesthesia was induced with propofol 100mg and atracurium 25mg with equipment of monitoring including BIS, invasive arterial pressure, pulse oximeter and electrocardiogram. After tracheal intubation, the anesthesia was maintained by 6% desflurane, 0.1µg/kg-min remifentanyl and intermittent atracurium. Triple lumen venous catheter was placed in the right internal jugular vein with fluid infusion and central venous pressure monitor. Esophageal probe was inserted to record body temperature. Warming blanket with convective air warming system covered chest, both arms and head to keep body temperature normal (Fig.1). BIS value maintained from 35 to 50 and electromyography (EMG) was almost fixed at 27 during one and half hour of operation. BIS Signal Quality Index (BIS-SQI) and Suppression Ratio (SR) was 100 and 0, respectively. Blood pressure was 131 over 72 mmHg, and heart rate was 69 beats per minute. WarmTouch™ was operated for body heating because esophageal temperature dropped from 36.3 to 36.0°C. BIS and EMG increased to 74 and 37, respectively. BIS-SQI and SR was 100 and

1. Dr. Se Hun Kim, MD.
 2. Dr. Byeong-Cheol Lee, MD.
 3. Prof. Yong Han Kim, MD.
- 1-3: Department of Anesthesiology and Pain Medicine, Haeundae Paik Hospital, Inje University, Busan, South Korea.

Correspondence:

Dr. Yong Han Kim, MD.
Assistant Professor,
Department of Anesthesiology and Pain Medicine,
Haeundae Paik Hospital,
Inje University,
875 Haeundae-ro, Haeundae-gu,
Busan, South Korea.
E-mail: adonis94@naver.com

- * Received for Publication: January 28, 2016
- * Revision Received: January 29, 2016
- * Accepted for Publication: * April 15, 2016

0, unchanged. Hemodynamics was steady state. The anesthesiologist added up desflurane to 8% concentration and inject atracurium 5mg. However, BIS was over 70 and EMG was 37. After turning off WarmTouch™, BIS and EMG returned to 45 and 27. BIS and EMG was raised again to same numerical value after switch on WarmTouch™. Heated and humidified breathing circuit (Heated Circuit Kit®, Ace Medical, South Korea) was applied for normothermia without activation of WarmTouch™. Esophageal temperature was maintained at 36.0 °C to recovery of anesthesia. Total operation time was four hours and the patient did not recall any intraoperative event. The patient was discharged without any symptom or complication on postoperative 22 days.

DISCUSSION

BIS monitor is a helpful device for adequate depth of anesthesia during operation. The value of BIS is superior to clinical signs for hypnotic effect of anesthesia.² However, several anesthetics, devices and clinical conditions could affect BIS value.¹

BIS was increased after ketamine bolus (0.4mg/kg) followed by continuous infusion (1mg/kg.h).³ Halothane anesthesia reflects higher BIS than sevoflurane in breast surgery, therefore halothane inhalation based on only BIS has a possibility of overdose.⁴ BIS of halothane at 1 MAC is 56.1 whereas isoflurane is 33.2 in children.⁵

Some of the electronic devices could interfere BIS parameter. Forced air warmer (Bair Hugger®, Augustine Medical, USA) elevated BIS without hemodynamic change during hepatectomy with isoflurane, sufentanyl, and atracurium.⁶ Falsely increased BIS was also noted during cardiac surgery heated by Bair Hugger® with sevoflurane, fentanyl and rocuronium anesthesia without memory recall.⁷ Endoscopic shaver attributed falsely elevated BIS with sevoflurane, remifentanyl and rocuronium during endoscopic shoulder surgery.⁸ Pacemaker-induced artifact appeared after deep hypothermic circulatory arrest with propofol, remifentanyl and rocuronium in cardiac surgery.⁹ BIS increased falsely in patients with electromagnetic operating device (VTI surgery).

Perioperative normothermia is very important to prevent following complications: delayed recovery, myocardial ischemia, shivering, hypertension, tachycardia, and prolonged coagulation time. Various methods could be adapted for maintenance of temperature including forced air blanket, water mattress, warm intravenous fluid, and heated airway



Fig.1: BIS was surrounded by WarmTouch™ and blanket.

circuit.¹¹ WarmTouch™ is an effective method for preventing complication of hypothermia. The vibration of air warmer could interfere BIS value through high EMG signal. BIS represented higher value than real depth of hypnotic level. The ventilation outlet of WarmTouch™ was set close to BIS as it seems in Fig.1.

In conclusion, BIS can be interpreted incorrectly by convective air warming system, therefore, deserves a special attention.

ACKNOWLEDGEMENT

This study was presented in 41st Romanian Congress of Anaesthesia and Intensive Care as a poster, 13th -17th May 2015. The authors declare that there is no conflict of interests.

Grant Support & Financial Disclosures: None.

REFERENCES

1. Dahaba AA. Different conditions that could result in the bispectral index indicating an incorrect hypnotic state. *Anesth Analg.* 2005;101:765-773. DOI:10.1213/01.ane.0000167269.62966.af
2. Smajic JI, Praso M, Hodzic M, Hodzic S, Srabovic-Okanovic A, Smajic N, Djonlagic Z. Assessment of depth of anesthesia: PRST score versus bispectral index. *Med Arh.* 2011;65(4):216-220. doi: 10.5455/medarh.2011.65.216-220
3. Vereecke HE, Struys MM, Mortier EP. A comparison of bispectral index and ARX-derived auditory evoked potential index in measuring the clinical interaction between ketamine and propofol anaesthesia. *Anaesthesia.* 2003;58(10):957-961. DOI: 10.1046/j.1365-2044.2003.03403.x
4. Edwards JJ, Soto RG, Thrush DM, Bedford RF. Bispectral index scale is higher for halothane than sevoflurane during intraoperative anesthesia. *Anesthesiology.* 2003;99(6):1453-1455.
5. Davidson AJ, Czarnecki C. The Bispectral Index in children: comparing isoflurane and halothane. *Br J Anaesth.* 2004;92(1):14-17. doi: 10.1093/bja/aeh011
6. Guignard B, Chauvin M. Bispectral index increases and decreases are not always signs of inadequate anesthesia. *Anesthesiology.* 2000;92(3):903.
7. Hemmerling TM, Fortier JD. Falsely increased bispectral index values in a series of patients undergoing cardiac surgery using forced-air-warming therapy of the head. *Anesth Analg.* 2002;95(2):322-323. DOI: 10.1213/00000539-200208000-00014
8. Hemmerling TM, Migneault B. Falsely increased bispectral index during endoscopic shoulder surgery attributed to interferences with the endoscopic shaver device. *Anesth Analg.* 2002;95(6):1678-1679. DOI: 10.1097/00000539-200212000-00038
9. Bang JO, Son HJ, Lee EH, Hahm KD, Choi IC. Falsely increased bispectral index score during deep hypothermic circulatory arrest in cardiac surgery. *Korean J Anesthesiol.* 2012;63(4):372-373. doi: 10.4097/kjae.2012.63.4.372.
10. Sessler DI. Temperature regulation and monitoring, *Miller's Anesthesia* (Seventh edition), Miller RD, Churchill Livingstone Elsevier, Philadelphia, USA. 2010;pp 1533-1556.

Authors' Contribution:

SHK editing of manuscript.

BCL and YHK clinical data acquisition.

YHK did manuscript writing, review and final approval.