

Electrocardiogram Signals Denoising Using Improved Variational Mode Decomposition

Abstract

Background: Electrocardiogram (ECG) plays a vital role in the analysis of heart activity. It can be used to analyze the different heart diseases and mental stress assessment also. Various noises, such as baseline wandering, muscle artifacts and power line interface disturbs the information within the ECG signal. To acquire correct information from ECG signal, these noises should be removed. **Methods:** In the proposed work, the improved variational mode decomposition (IVMD) method for the removal of noise in ECG signals is used. In the proposed method, the weighted signal amplitude integrated over the timeframe of the ECG signal varies the window size during decomposition. Raw ECG data are extracted from 10 subjects and ECG data are also taken from the MIT BIH database for the proposed method. **Results:** The performance comparison of traditional variational mode decomposition (VMD) and the proposed technique is also calculated using mean square error, percentage root mean square difference, signal to noise ratio and correlation coefficient. The extracted highest signal to noise ratio (SNR) value of acquired ECG signals using traditional VMD is 42db whereas highest value of signal to noise ratio (SNR) using improved VMD (IVMD) is 83db. **Conclusion:** The proposed IVMD technique represented better performance than traditional VMD for denoising of ECG signals.

Keywords: Denoising, electrocardiogram, variational mode decomposition

Submitted: 29-Feb-2020

Revised: 22-Apr-2020

Accepted: 24-May-2020

Published: 24-May-2021

Introduction

Electrocardiogram (ECG) patterns reflect the electrical activity produced by the heart. It accumulates a ton of information for human heart well-being.^[1] ECG signal is commonly weak and it is mostly affected by noise. Noise is an undesirable component. Noise does not store any heart-related information.^[2] Noise is the basic reason for the error in analysis measurement. Different types of noises are baseline wander noise (0.15–3 Hz) produced because of breath, electromyography (EMG) or muscle artifacts created because of muscle constriction, power line interface (PLI) (50–60 Hz) produced because of power supply, and electrosurgical noise (100 kHz to 1 MHz) produced because of others clinical-related machines. Electrode contact noise created because of inadequate closeness among electrodes and skin.^[3] However, the most widely recognized noises are PLI, baseline wandering, and motion artifacts.^[4] Thus, for accurate and

reliable analysis, these noises should be removed from the corrupted signal. The techniques to denoise the ECG signal can be separated into three classifications: (1) frequency domain, (2) spatiotemporal technique, and (3) statistical method. Distinctive decompositions, for example, empirical mode decomposition (EMD), variational mode decomposition (VMD), and ensemble EMD (EEMD), are used in frequency-domain method to denoise the ECG signal. Different neural networks such as principal component analysis (PCA) and independent component analysis are used in the statistical method.

Various filters such as median filter, adaptive filter, and low-pass filter are also used under the spatiotemporal method.^[5] Error in reference signal reduces the effectiveness of the adaptive filter. Edges of the ECG signals are not protected in the Wavelet method. The neural network system and principal component analysis (PCA) do not represents good performance for noise cancellation in a single channel ECG Signal.^[6]

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: WKHLRPMedknow_reprints@wolterskluwer.com

**Vikas Malhotra,
Mandeep Kaur
Sandhu**

*Department of Electronic and
Communication Engineering,
University School of
Engineering and Technology,
Rayat Bahra University, Mohali,
Punjab*

Address for correspondence:

*Er. Vikas Malhotra,
18-B, Jagdish Colony, Rajpura
Road, Patiala, Punjab, India.
E-mail: vikas_malhotra10@
yahoo.com*

Access this article online

Website: www.jmssjournal.net

DOI: 10.4103/jmss.JMSS_17_20

Quick Response Code:



How to cite this article: Malhotra V, Sandhu MK. Electrocardiogram signals denoising using improved variational mode decomposition. *J Med Sign Sens* 2021;11:100-7.

Because of the variety in the ECG signal range, high-pass filtering can distort the ECG waveform.^[7] It is hard to remove EMG noise using various filters, such as infinite impulse response and finite impulse response filter, because of overlap ECG recurrence use, i.e., 20–200 Hz.^[8]

Numerous analysts used various novel techniques for ECG denoising. Yadav *et al.*^[9] proposed a nonlocal wavelet transform (NLWT) technique to denoise the ECG signal distorted by additive white Gaussian noise (AWGN). This NLWT technique outperformed the nonlocal mean (NLM) and hybrid EMD methods, but complexity is increased because of adaptive thresholding.

Cuomo *et al.*^[10] used a recursive filtering technique (Gaussian filter based) to remove the noises, such as electrical PLI and baseline drift noise in real-time ECG signals. The result represented that the time and memory utilization are less when contrasted with other filtration techniques, such as bandpass filter, low-pass filter, Kalman filter, double-stage moving average filter, and single- and double-stage median filter. Tobon and Falk^[11] proposed an adaptive spectrotemporal technique using a bandpass filter to remove noises such as baseline wandering and muscle artifacts from synthetic ECG signal as well as from long haul recorded ECG signal, but the spectrotemporal method takes twice the computational time when contrasted with standard techniques (EMD and wavelet). Hesar and Mohebbi^[12] used marginalized particle extended Kalman filter (MP-EKF) to denoise the MIT-BIH typical sinus rhythm ECG signals. The ECG signals are altered with artificial white Gaussian noise, muscle artifacts, pink noise, and brown noise at various input signal-to-noise ratios (SNRs) in MP-EKF method. Underestimated MP-EKF is much slower in speed than in extended kalman filter (EKF)/Kalman smoother denoising algorithm (EKS). The backtracking issue in EKS provided smooth but not the best result. Lahmiri^[13] compared two hybrid denoising systems, for example, EMD-DWT and VMD-DWT. Gurjit^[14] evaluated the performances of various windows (Hamming, Rectangular, Welch, Kaiser, and Hanning) for denoising ECG signals for various noises (power noise, muscle noise, and EMG noise).

El B'charri *et al.*^[15] used a dual-tree wavelet transform with an appropriate modified threshold function for denoising the synthetic ECG. The synthetic ECG signal is combined with various noises (colored, white, baseline wander, electromyogram, and motion artifacts), but some distortion because of flicker noise is present in the denoised signal. Oliveira *et al.*^[16] proposed a novel discrete wavelet transform-based technique without a thresholding method for the removal of PLI noise in the ECG signal. This novel discrete wavelet transform technique presented better results over the notch filter, but the performance of a novel discrete wavelet transform technique relies on adequate sampling frequency and decomposition level.

EMD is generally used for nonstationary and nonlinear signals. It is usually implemented by most of the

researchers. The performance of EMD-based method is superior to the wavelet technique.^[17] However, the EMD technique experienced the issue of mode mixing and loss of evolutionary trend information in the final residue. Even though the problem of mode mixing has been resolved in EEMD, missing out of evolutionary trend information in final residue is still an issue in the EEMD method. In addition, CEEMDAN with wavelet technique is used to evacuate the mode mixing issue of EMD technique. Even though the CEEMDAN with wavelet method represented better results than EMD, the impotence of wavelet technique to the removal of baseline wandering noise in the real-time ECG signal is a disadvantage.^[18] EMD-based method is useful, but the drawback of this method is that it is incapable to estimate the accurate morphological QRS complexes.^[19]

Prabhakararao and Manikandan^[20] suggested using VMD technique for ECG denoises, instead of other denoising techniques such as EMD, DFT, and DWT methods, as VMD has better capability to reduce noises. The VMD technique is appropriate for the ECG signal examination. However, existing VMD includes smooth windowing, for example, the static window of little and fixed size that works well on the shorter chunks of signals, over which the signals are almost stationary. Due to usage of small window size in existing VMD algorithm, the processing time of algorithm increases. This limitation can be minimized using the dynamic window of variable size. In the proposed work, the window size is being made as a function of weighted signal amplitude integrated over the timeframe.

Variational mode decomposition

Dragomiretskiy and Zosso^[21] proposed the VMD method. VMD is a nonrecursive signal processing technique. The variational issue is to discover the mode functions (modes) of a signal with the end goal that the sum of the bandwidth of every mode is least, and the requirement is that the sum of the bandwidth of every mode is equivalent to the original signal.^[21]

For the variational problem, VMD breaks down the signal into the discrete number of inherent modes. Every mode has explicit sparsity properties in the frequency domain, and it is compressed around a center frequency.

The bandwidth of every mode is dictated by the following criteria.

1. The single-frequency spectrum of every mode is controlled by the Hilbert transform of every mode $i(k)$. Symmetrical pair for every Intrinsic mode function (IMF) is a stage moved by 90° through Hilbert transform. Every IMF set and its symmetrical pair can be utilized to assess the momentary variation in magnitude and frequency of the IMF as for time. An analytical function is shaped after Hilbert change as demonstrated as follows:

$$\left(\delta(t) + \frac{j}{\pi t} \right) * i_k(t)$$

Where $i_k(t)$ is intrinsic mode function, $*$ is the convolution, $\delta(t)$ is the unit impulse function (Dirac distribution), and j is an imaginary part.

The low-frequency components represented by mode i with a higher value of k .

2. The frequency spectrum of each intrinsic mode function (mode) is shifted to baseband (center frequency) by multiplying the analytical function:

$$\left[\left(\delta(t) + \frac{j}{\pi t} \right) * i_k(t) \right] e^{-jw_k t}$$

Where w_k is the center frequency.

The bandwidth of each intrinsic mode function $i_k(t)$ is determined by the integral of the square of the time derivative of this frequency translated signal.

$$\min_{i_k, w_k} \left\{ \sum_k \left\| \partial t \left[\left(\delta(t) + \frac{j}{\pi t} \right) * i_k(t) \right] e^{-jw_k t} \right\|_2^2 \right\} \quad (1)$$

Subjected to: $\sum_k i_k(t) = x(t)$

Where ∂t is partial derivative and $x(t)$ is input signal to be decomposed.

Now to solve the optimizing problem (i.e., a minimum point) of Eq. 1, Dragomiretskiy and Zosso proposed the Lagrangian function and penalty term for Eq. 1 as given below:

$$\eta(\{i_k\}, \{w_k\}, \lambda) = \alpha \sum_{k=1}^K \left\| dt \left[\left(\delta(t) + \frac{j}{\pi t} \right) * i_k(t) \right] e^{-jw_k t} \right\|_2^2 + \left\| x(t) - \sum_{k=1}^K i_k(t) \right\|_2^2 + \langle \lambda(t), x(t) - \sum_{k=1}^K i_k(t) \rangle \quad (2)$$

Where λ = Lagrangian multiplier and α = penalty parameter, and by increasing it, the bandwidth of IMF is decreased.

To obtain the corresponding updated equation of i_k and w_k , an alternate direction method of multipliers (ADMM) is used. The updated equation of each mode in the frequency domain is obtained as given below:

$$\hat{i}_k^{n+1}(w) = \frac{\hat{x}(w) - \sum_{e=1, e \neq k}^k \hat{i}_e(w) + \frac{\hat{\lambda}(w)}{2}}{1 + 2\alpha(w - w_k)^2} \quad (3)$$

Where $\hat{\cdot}$ are frequency domain variables and $n + 1$ is the number of iteration.

This Eq. 3 is the result of the current residue considered as the Wiener filtering with the signal prior $\frac{1}{(w - w_k)^2}$

Similarly, the center frequency is transferred into the frequency domain and the update equation is given below:

$$w_k^{n+1} = \frac{\int_0^\infty w |\hat{i}_k(w)|^2 dw}{\int_0^\infty |\hat{i}_k(w)|^2 dw} \quad (4)$$

Dual ascent for all $w \geq 0$, Lagrangian multiplier (λ) is updated as:

$$\hat{\lambda}^{n+1}(w) = \hat{\lambda}^n(w) + \tau \left(\hat{x}(w) - \sum_k \hat{i}_k^{n+1}(w) \right) \quad (5)$$

Where τ is time-step of the dual ascent.

Repeat the up-gradation of Eq. 3, 4, 5 until

$$\sum_k \left\| \hat{i}_k^{n+1} - \hat{i}_k^n \right\|_2^2 / \left\| \hat{i}_k^n \right\|_2^2 < \varepsilon \quad (6)$$

Where ε is the tolerance of the convergence criterion, and in the VMD algorithm, it required to set manually.

Now, adaptive decomposition of the signal band can be used due to frequency domain characteristics.

Before performing the VMD algorithm, the Gaussian window is simply applied to the input signal $x(t)$. By simple addition after decomposition, the individual modes can be sewed together without error amplification, instead of window division. Because window division affects, the reconstruction fidelity closes to the window borders. The flowchart of traditional variational mode decomposition is shown in Figure 1.

Improved Variational Mode Decomposition

In improved VMD (IVMD), the estimation of K is assessed by estimation of r with subject to the threshold ceiling estimation of 0.05, where r is the correlation coefficient (CC) of Y and X . The CC can be represented as:

$$r = \text{cov}(Y, X) / \text{std}(X) \text{std}(Y) \quad (7)$$

Where $\text{cov}(Y, X)$ is the covariance between X and Y , X is the combined modal segments, and Y is the original input signal.

The idea of the dynamic window involved an element of weighted signal amplitude integrated over a period of time, and it is implemented using inverse short-time Fourier transform (STFT) and STFT rather than traditional discrete Fourier transform. The output of the STFT function is complex STFT coefficients, a time vector, and a frequency vector. STFT-self does not improve the frequency or time resolution. The dynamic window is implemented by the Hamming window and Blackman–Harris window. Hamming window cancels the nearest side lobes for the better results than other windowing techniques. The Blackman–Harris window has good side lobe compression. Blackman–Harris breaks our signal in segments that are




Figure 1: Flowchart of traditional VMD [21]

processed independently and added to the other processed segments using the WOLA algorithm. This technique helps avoid the anomalies on boundaries encountered in the VMD algorithm.

Improved algorithm

- Generate the analysis window and the synthesis window:
 - $awin = \text{blackmanharris}(wlen, \text{"periodic"})$;
 - $swin = \text{hamming}(wlen, \text{"periodic"})$;
 Where $wlen$ is the window length.
- Generate the zeros vector:
 - Equal to the length of the signal ($xlen$) to save the successive results obtained from each processed window segment
 - Equal to the length of the signal ($xlen$) to save the successive modes (number of modes had been calculated earlier) results obtained from each processed window segment.
- Calculate the number of window frames to be processed. $L = 1 + \text{fix}((xlen - wlen) / \text{hop})$ where hop is the shifting parameter to move to the next window.
 - Note: these windows overlap for the "hop" number of points in the signal.
- For each window:
 - Generate the window part by multiplying each $wlen$ long part of the signal by an analysis window
Signal $(1 + l * \text{hop} : wlen + 1 * \text{hop}) * awin$
 - Process it with the traditional VMD method
 - Save the processed signal window by adding

it to the zero vectors after multiplying it with the synthesis window "swin" as it has been truncated along sides in VMD processing and we want smooth WOLA addition of the overlapping processed signal segments

- Do the same for all the modes of the VMD processed windowed signal.
- Now revert the amplitude change induced by a window function ($W0$) by:
 - $W0 = \text{sum}(awin * swin)$
 - Processed signal = Processed signal/ $W0$
 - Processed signal modes = Processed signal modes/ $W0$.

Experimental Work

Noisy ECG data acquired from 10 subjects (between age 30 and 35) and ECG data from the MIT-BIH database are also used for the proposed algorithm. The signal's duration is taken as 10 s and sampled at a sampling frequency of 720 Hz, with a resolution of 10 bits per sample. The 118 and 119 signals of noise stress test data taken from the MIT-BIH database are already altered with three different types of noises, i.e., baseline wander, muscle (EMG) artifact, and electrode motion artifact, whereas other MIT-BIH database signals (100, 101, and 102) are altered with electrode motion (em) noise using `nstgen()` script with different noise levels. The proposed IVMD is used for decomposition for all these ECG signals. The noises are removed by eliminating the lowest and the highest IMF functions after decomposition using both VMD as well as IVMD technique. MATLAB R2019a platform is used for all data processing.

Results and Discussion

The four parameters, i.e., mean square error (MSE), percentage root mean square difference (PRD), SNR, and CC, are used to investigate the performance of proposed IVMD method. MSE, percentage root mean square, and CC at the different SNR of the ECG signals are taken from the MIT-BIH database. SNR parameter is extracted from acquired ECG signals of 10 subjects.

If $x(n)$ is original signal, $x_n(n)$ is noisy signal, and $y(n)$ is denoised ECG signal, N is the length of the ECG signal, the parameters for performance evaluation can be expressed as follows:

Mean square error

To estimate the original signal, the filtering technique's accuracy has been traced by the MSE. The energy of the error signal in the noise removal process is generally defined by the MSE. For better estimation of the original signal and better protection of signal details, the MSE value is generally lower.

$$MSE = \frac{1}{N} \sum_{n=0}^{N-1} [x[n] - y[n]]^2 \quad (8)$$

Percentage root mean square difference

It is generally used for the detection of the efficiency of the noise removal technique during signal extraction with the protection of the necessary medical information within the signal. The superior preservation of necessary physiological information in the denoised ECG signal shows a lower value of PRD.

$$PRD = \sqrt{\frac{\sum_{n=0}^{N-1} (x[n] - y[n])^2}{\sum_{n=0}^{N-1} x^2[n]}} \times 100 \quad (9)$$

Signal-to-noise ratio

The energy of the signal concerning related noise is called the SNR. This SNR parameter value should be higher for a better denoising method. It can be defined as:

$$SNR = 10 \log \left(\frac{\sum_{n=0}^{N-1} [x_n[n]]^2}{\sum_{n=0}^{N-1} [x_n[n] - y[n]]^2} \right) \quad (10)$$

Correlation coefficient

It represents the statistical relationship between the original signal and the denoised signal. A higher value of the CC of a method represents the better reinstate property of the original signal by that method.

$$CC = \frac{\sum_{n=0}^{N-1} (x[n] - \bar{x})(y[n] - \bar{y})}{\sqrt{\sum_{n=0}^{N-1} (x[n] - \bar{x})^2 \sum_{n=0}^{N-1} (y[n] - \bar{y})^2}} \quad (11)$$

The graphical representation of the CC of arrhythmia signal (118) taken from MIT-BIH database using VMD and IVMD method at different SNR is shown in Figure 2. Similarly, the graphical representations of the MSE and PRD of arrhythmia signal (118) taken from MIT-BIH database using VMD and IVMD method at different SNR are shown in Figures 3 and 4, respectively. The graphical representation of the SNR of acquired signals using VMD and IVMD method at different SNR is shown in Figure 5. The result of noisy acquired ECG signal [Figure 6] after denoising using VMD and IVMD is represented in Figures 7 and 8, respectively. It can be seen that IVMD represented better results as compared to VMD decomposition.

The minimum extracted values of PRD and MSE and the highest extracted value of CC of proposed IVMD method for MIT-BIH ECG data signals are 71.63, 0.191, and 0.952, respectively, as shown in Table 2. The minimum extracted values of PRD and MSE and the highest extracted value of the SNR of traditional VMD technique for MIT-BIH ECG data signals are 83.87, 0.229, and 0.852, respectively, as also shown in Table 2. In addition, for acquired ECG

data, proposed VMD represented better result than traditional VMD technique as the highest extracted value of the SNR of traditional VMD is 42 dB and the highest extracted value of the SNR of improved variational mode decay is 83 dB as shown in Table 1. As compared to existing methods such as EMD with adapting switching filter method,^[19] periodic nonlinear mean filter method,^[8] and EMD with NLM method,^[6] the proposed method represented better results with the lowest value in terms of MSE. The proposed method still involves the limitation of the Fourier spectrum, i.e., in the Fourier spectrum, different components cannot be separated. However, the proposed method can be used for ECG analysis in the future to find various diseases, arrhythmia with enhanced performance.

Table 1: Signal-to-noise ratio after variational mode decomposition and improved variational mode decomposition processing of acquired noisy real electrocardiogram data

ECG signals	SNR (dB)	
	IVMD	Traditional VMD
N1	76	42
N2	77	35
N3	70	40
N4	83	36
N5	60	37
N6	68	32
N7	65	40
N8	70	27
N9	65	41
N10	66	35

ECG – Electrocardiogram; SNR – Signal-to-noise ratio; VMD – Variational mode decomposition; IVMD – Improved VMD




Figure 2: Graphical representation of the correlation coefficient using variational mode decomposition and improved variational mode decomposition process at different signal-to-noise ratio of signal 118.




Figure 3: Graphical representation of the mean square error using variational mode decomposition and improved variational mode decomposition process at different signal-to-noise ratio of signal 118




Figure 4: Graphic representation of the percentage root mean square difference using variational mode decomposition and improved variational mode decomposition process at different signal-to-noise ratio of the signal 118




Figure 5: Graphic representation of the signal to noise ratio of the acquired electrocardiogram signals after using traditional variational mode decomposition and improved variational mode decomposition process




Figure 6: An acquired noisy raw electrocardiogram signal




Figure 7: Denoised electrocardiogram signal after using the traditional variational mode decomposition




Figure 8: Denoised electrocardiogram signal after using the improved variational mode decomposition

Conclusion

This IVMD represented better results than traditional VMD. It is increasingly proficient to utilize this method for noise removal in ECG data.

Financial support and sponsorship

None.

Conflicts of interest

There are no conflicts of interest.

Table 2: Different performance parameters for MIT-BIH noisy electrocardiogram data

ECG data	SNR (dB)	IVMD			Traditional VMD		
		MSE ($\times 10^4$)	PRD	CC	MSE ($\times 10^4$)	PRD	CC
100	0	0.764	127.69	0.404	0.979	139.1	0.109
	6	0.425	95.23	0.633	0.461	99.21	0.413
	12	0.358	87.41	0.826	0.416	94.28	0.647
	18	0.342	85.45	0.923	0.408	93.38	0.713
	24	0.339	85.09	0.946	0.407	93.21	0.726
101	0	0.929	128.89	0.417	0.915	127.9	0.217
	6	0.485	93.16	0.658	0.509	95.43	0.528
	12	0.375	81.99	0.812	0.456	90.30	0.671
	18	0.332	77.08	0.888	0.443	89.03	0.725
	24	0.322	75.89	0.923	0.440	88.71	0.741
102	0	0.406	111.63	0.476	0.475	120.8	0.381
	6	0.246	86.97	0.714	0.260	89.38	0.714
	12	0.208	80.00	0.846	0.236	85.13	0.807
	18	0.195	77.50	0.911	0.230	84.11	0.842
	24	0.191	76.62	0.933	0.229	83.87	0.852
118	0	4.641	111.67	0.415	5.264	118.9	0.127
	6	3.287	93.98	0.624	3.527	97.35	0.553
	12	2.960	89.18	0.793	3.393	95.48	0.684
	18	3.043	90.42	0.875	3.362	95.04	0.724
	24	3.068	90.79	0.886	3.354	94.93	0.734
119	0	3.355	96.42	0.566	3.987	105.1	0.433
	6	2.273	79.36	0.787	2.979	90.85	0.665
	12	2.026	74.92	0.899	2.873	89.22	0.734
	18	1.852	71.63	0.942	2.849	88.84	0.754
	24	1.869	71.96	0.952	2.844	88.77	0.758

ECG – Electrocardiogram; SNR – Signal to noise ratio; VMD – Variational mode decomposition; MSE – Mean square error; PRD – Percentage root mean square difference; CC – Correlation coefficient; IVMD- Improved VMD

References

- Sahoo S, Biswas P, Das T, Sabut S. De-noising of ECG signal and detection using Hilbert transform and adaptive thresholding. *Process Technol* 2016;25:68-75.
- Wing Z, Wan F, Wong CM, Zhang L. Adaptive Fourier decomposition based ECG de-noising. *Comput Biol Med* 2016;77:195-206.
- Berkaya SK, Uysal AK, Uysal ES, Gunal ES, Ergin S. A survey on ECG analysis. *Biomed Sig Process Control* 2018;43:216-35.
- Tayel MB, Eltrass AS, Ammar AI. A new multistage combined kernel filtering approach for ECG noise removal. *Electron Cardiol J* 2017;51:265-75.
- Tian X, Li Y, Zhou H, Li X, Chen L, Zhang X. Electrocardiogram signal de-noising using extreme-point symmetric mode decomposition and non-local mean. *Sensors* 2016;16:10.
- Kumar S, Panigrahy D, Sahu PK. De-noising of electrocardiogram signal by using empirical mode decomposition with non-local mean technique. *Biocybern Biomed Eng* 2018;38:297-312.
- Gupta P, Sharma KK, Joshi SD. Baseline wander removal of electrocardiogram signals using multivariate empirical mode decomposition. *IEEE Health Technol Lett* 2015;2:164-6.
- Lee YD, Hwang D. Periodicity-based non-local means denoising method for electrocardiography in low SNR non-white noisy conditions. *Biomed Sig Process Control* 2018;39:284-93.
- Yadav SK, Sinha R, Bora PK. Electrocardiogram signal de-noising using non-local wavelet transform domain filtering. *IET Sig Process* 2015;9:88-96.
- Cuomo S, Pietro GD, Farina R, Galletti A, Sannino G. A revised scheme for real-time ECG signal de-noising based on recursive filtering. *Biomed Sig Process Control* 2016;27:134-44.
- Tobon DP, Falk TH. Adaptive spectro-temporal filtering for electrocardiogram signal enhancement. *IEEE J Biomed Health Inform* 2018;22:2.
- Hesar HD, Mohebbi M. ECG de-noising using marginalized particle extended kalman filter with automatic particle weighting strategy. *IEEE J Biomed Health Inform* 2017;21:3.
- Lahmiri S. Comparative study of ECG signal de-noising by wavelet thresholding in empirical and variational mode decomposition domains. *Health Technol Lett* 2014;1:104-9.
- Gurjit PK. Noise removal in ECG signal using windowing technique and its optimization. *Adv Biotechnol Microbiol* 2017;6:1.
- El B'charri O, Latif R, Elmansouri K, Abenaou A, Jenkal W. ECG signal performance de-noising assessment based on threshold tuning of dual-tree wavelet transform. *Biomed Eng Online* 2017;16:26.
- Oliveira BR, Duarte MA, Abreu CC. A wavelet-based method for power line interface removal in ECG signals. *Res Biomed Eng* 2018;34:73-86.
- Suchetha M, Kumaravel N, Jagannath M, Jaganathan SK. A comparative analysis of EMD based filtering methods for 50Hz noise cancellation in ECG signal. *Inform Med Unloc* 2017;8:54-9.
- Xu Y, Luo M, Li T, Song G. ECG signal de-noising and baseline wander correlation based on CEEMDAN and wavelet threshold. *Sensors* 2017;17:12.

19. Rakshit M, Das S. An efficient ECG de-noising methodology using empirical mode decomposition and adaptive switching mean filter. *Biomed Sig Procees Control* 2018;40:140-8.
20. Prabhakararao E, Manikandan MS. On the use of Variational Mode Decomposition for Removal of Baseline Wander in ECG Signals. *National Conference on Communications (NCC)*; 2016. p. 1-6.
21. Dragomiretskiy K, Zosso D. Variational mode decomposition. *IEEE Trans Sig Proce* 2014;62:531-44.
22. Liu S, Tang G, Wang X, He Y. Time-frequency analysis based on improved variational mode decomposition and teager energy operator for rotor system fault diagnosis. *Math Probl Eng* 2016;11:1-19.

BIOGRAPHIES



Vikas Malhotra received his B.E degree in Electronics and Communication (ECE) in the year 2005 from Shaheed Udham Singh College of Engineering & Technology, Punjab Technical University, Punjab, M.E. degree in Electronics and Communication (ECE) in the year 2014. He is currently a research scholar with the University School of Engineering and Technology (USET), Department of Electronics and Communication (ECE), Rayat Bahra University (RBU), Mohali, Punjab. His research interest includes biomedical signal processing, Embedded system.

Email: vikas_malhotra10@yahoo.com



Mandeep Kaur Sandhu Mandeep Kaur Sandhu received her Bachelor's degree in Electronics and Communication Engineering from Punjabi University and the Masters degree from Punjabi University in 2005 and 2007 respectively. She has done her Ph.D. in the Department of Electrical and Electronics Engineering, Maharishi Markandeshwar University (MMU), Mullana, India in the year 2018. Presently, she is working as a Assistant Professor in Electronics & Communication Engineering department, University School of Engineering and Technology (USET), Rayat-Bahra University (RBU), Mohali. Her research interest includes Optical Networks and Wireless Networks, Neural networks.

Email: mandeep.sandhu@rayatbahrauniversity.edu.in