

POSTER PRESENTATION

Open Access

3D cardiac Chemical Shift Imaging of [1-13C] hyperpolarized acetate and pyruvate in pigs

Luca Menichetti^{1,2}, Francesca Frijia^{2*}, Alessandra Flori³, Vincenzo Lionetti³, Matteo Liserani⁴, Giulio Giovannetti¹, Giacomo Bianchi², Simone L Romano³, Vincenzo Positano^{2,1}, Jan Henrik Ardenkjaer-Larsen^{5,6}, Rolf F Schulte⁷, Fabio A Recchia², Luigi Landini⁸, Maria Filomena Santarelli^{1,2}, Massimo Lombardi²

From 16th Annual SCMR Scientific Sessions
San Francisco, CA, USA. 31 January - 3 February 2013

Background

¹³C Dynamic Nuclear Polarization (DNP) with rapid dissolution together with Magnetic Resonance Chemical Shift Imaging (CSI) have been used for non-invasive real-time metabolic assessment in cardiac experimental models on a clinical 3T scanner. Here, we report an in vivo comparison of hyperpolarized [1-¹³C] pyruvate and [1-¹³C] acetate perfusion and metabolism: a method based on a 3D Spiral CSI sequence is presented for obtaining spatially and spectrally-resolved information on whole heart cardiac metabolism.

Methods

In this work hyperpolarized [1-¹³C] pyruvate and [1-¹³C] acetate were injected in vivo to obtain spatially and spectrally resolved information of basal metabolism on whole heart in middle size animal models. Five healthy male farm pigs (38±2 kg) were studied in basal condition and subjected to imaging experiments performed on a 3T GE Signa HDx scanner using a ¹³C-quadrature birdcage coil (Rapid Biomedical). An HyperSense DNP polarizer (Oxford Inst.) was employed for the studies: a procedure for the hyperpolarization and dissolution of a large dose of TRIS-[1-¹³C]acetate water/glycerol mixture was set up while the preparation of a large dose of [1-¹³C] pyruvic acid was performed as recently published by this group. An anatomical region of interest covering the whole heart was first acquired with a proton reference scan and the metabolic information was then obtained using 3D IDEAL spiral CSI on the same region. Image re-slicing along cardiac short axis (SA) views and image fusion of ¹³C

metabolite maps and anatomical 1H reference images were performed by PMOD software.

Results

A graph of the γ -variate and mono-exponential fitting of hyperpolarized [1-¹³C] acetate myocardial spectroscopic signals is reported in Figure 1 while a representative map in SA orientation through the heart is shown in Figure 2: [1-¹³C] acetate is extracted inside the heart and clearly detected in the heart-chambers and myocardial wall. Representative maps of spatial distribution of [1-¹³C] bicarbonate, [1-¹³C] lactate and [1-¹³C] pyruvate in SA orientation through the heart are also produced using hyperpolarized [1-¹³C] pyruvate.

Conclusions

A comparison between acetate and pyruvate ¹³C-mapping has been realised as far as we know for the first time in pigs with this experimental approach. This ongoing study demonstrates the feasibility of whole-heart ¹³C-cardiac metabolic imaging in pigs for detecting and mapping cardiac metabolism in basal condition with hyperpolarized [1-¹³C]acetate in comparison with [1-¹³C] pyruvate.

This study is the first step towards the optimization of the [1-¹³C] acetate concentration and the acquisition sequence parameters to ensure suitable MR signals in myocardial tissue and to study its metabolic fate.

Funding

Self funding.

²Fondazione G. Monasterio CNR-Regione Toscana, Pisa, Italy
Full list of author information is available at the end of the article




Figure 1 ^{13}C dynamic spectra were acquired using a slice selective pulse-and-acquire sequence (bandwidth 5000 Hz, 2048 pts, 10° FA). A long-axis slice of 20 mm was selected during excitation. Spectra were acquired from the beginning of the injection of the hyperpolarized [1- ^{13}C] acetate, every 2 s, for 120 s. Diagrammatic representation of the γ -variate and mono-exponential fitting of cardiac spectroscopic signal to obtain rate constants (N=4).




Figure 2 Representative maps in SA view of the heart showing the in vivo spatial distribution of hyperpolarized TRIS-[1- ^{13}C] acetate in pigs; spectroscopic data were normalized to the maximum value of signal amplitude.

Author details

¹Institute of Clinical Physiology, National Research Council, Pisa, Italy. ²Fondazione G. Monasterio CNR-Regione Toscana, Pisa, Italy. ³Scuola Superiore Sant'Anna, Pisa, Italy. ⁴Faculty of Physics, University of Pisa, Pisa, Italy. ⁵GE Healthcare, Hillerod, Denmark. ⁶Department of Electrical Engineering, Technical University of Denmark, Kongens Lyngby, Denmark. ⁷GE Global Research, Munich, Germany. ⁸Department of Information Engineering, University of Pisa, Pisa, Italy.

doi:10.1186/1532-429X-15-S1-P10

Cite this article as: Menichetti *et al.*: 3D cardiac Chemical Shift Imaging of [1- ^{13}C] hyperpolarized acetate and pyruvate in pigs. *Journal of Cardiovascular Magnetic Resonance* 2013 **15**(Suppl 1):P10.