



POSTER PRESENTATION

Open Access

Real time cardiac MRI and its clinical usefulness in arrhythmias and wall motion abnormalities

Christina Unterberg-Buchwald^{1,2*}, Martin Fasshauer², Jan M Sohns², Wieland Staab², Andreas Schuster¹, Dirk Voit³, Johannes T Kowallick¹, Michael Steinmetz⁴, Jens Frahm³, Joachim Lotz²

From 17th Annual SCMR Scientific Sessions
New Orleans, LA, USA. 16-19 January 2014

Background

Analysis of cardiac function in patients with arrhythmias is very limited or nearly impossible in ECG synchronized cine acquisitions with balanced SSFP. A real-time method at 1.5 T with SSFP contrast was used to show that this method is superior for image quality and analysis of ventricular function in a subset of patients with atrial fibrillation and AV-Block.

Methods

Radial gradient-echo sequences with fully balanced SSFP gradients and at least 15-fold undersampling (real-time SSFP, RT) and conventional ECG-synchronized cine SSFP CMR (Cine) was used on a 1.5 T scanner system. Patients who had permanent arrhythmias (most often atrial fibrillation, AA; n = 8) or wall motion abnormalities (WMA; n = 3) were scanned in the standard views

and compared to patients in sinus rhythm (SR; n = 21) without wall motion abnormalities. Image reconstruction of RT was performed offline by regularized non-linear inversion. Quality (IQ) of scan was detected using an image quality score (ranging from 0 = no diagnostic quality to 1 = reduced diagnostic quality, 2 = many artifacts, 3 = some artifacts, and 4 = optimal diagnostic quality). Cardiac function was analyzed using a semiautomatic contour detection (Q mass, Medis, NL) applied to 5 consecutive beats in RT and Cine averaging data of 10-12 beats. All analysis was done by two blinded observers (3 and 8 years experience of CMR evaluation).

Results

IQ was comparable for Cine and RT SSFP in patients with WMA or patients in sinus rhythm without WMA given in Table 1. However, IQ was significantly better

Table 1

Views for IQ and quantitative parameters	Patient groups	Number of views	RT	Cine
Short axis base (IQ)	AA	16	$3.94 \pm 0.25^*$	2.75 ± 0.68
	WMA	6	4.00 ± 0.00	4.00 ± 0.00
	SR	42	3.62 ± 0.54	3.95 ± 0.22
Short axis mid (IQ)	AA	16	$3.81 \pm 0.40^*$	2.69 ± 0.60
	WMA	6	3.67 ± 0.5	4.00 ± 0.00
	SR	42	3.48 ± 0.67	3.93 ± 0.26
Short axis apex (IQ)	AA	16	$3.81 \pm 0.40^*$	2.19 ± 0.91
	WMA	6	3.00 ± 1.26	3.50 ± 0.55
	SR	42	3.07 ± 0.97	3.70 ± 0.55
Enddiastolic volume [ml]	AA	16	55.15 ± 10.17	36.80 ± 20.13
	WMA	3	52.39 ± 3.91	43.10 ± 6.73
	SR	16	47.50 ± 12.70	51.32 ± 18.4
Endsystolic volume [ml]	AA	16	32.30 ± 13.11	21.60 ± 16.50
	WMA	3	30.50 ± 9.80	24.10 ± 6.86
	SR	16	25.10 ± 13.0	22.40 ± 10.70

*Cardiology, University Clinic Goettingen, Goettingen, Germany
Full list of author information is available at the end of the article

© 2014 Unterberg-Buchwald et al.; licensee BioMed Central Ltd. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<http://creativecommons.org/licenses/by/2.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated.

Table 1 (Continued)

Beat volume [ml]	AA	16	22.83 ± 6.30	15.20 ± 6.20
	WMA	3	21.90 ± 8.50	19.00 ± 1.27
	SR	16	22.37 ± 5.30	29.84 ± 16.7
Ejection fraction [%]	AA	16	43.50 ± 15.10	46.30 ± 16.10
	WMA	3	41.70 ± 16.70	45.00 ± 6.90
	SR	16	50.00 ± 12.20	57.90 ± 12.50

* p < 0.001 RT vs Cine

for RT compared to Cine in all three short axis in patients with arrhythmias. Functional parameters showed no significant differences, however there was a trend to lower values for enddiastolic, endsystolic and beat volumes for Cine (10-12 beats) compared to RT (5 consecutive beats) in RT.

Conclusions

RT cardiac MRI is a robust method with high image quality that has the potential to allow functional analysis of sequences in patients with arrhythmias that are often difficult or impossible to analyze by cine SSFP.

Funding

No funding.

Authors' details

¹Cardiology, University Clinic Goettingen, Goettingen, Germany. ²Institute of Diagnostic and Interventional Radiology, University Clinic Goettingen, Goettingen, Germany. ³Biomedizinische NMR ForschungsGmbH, Max-Planck-Institut für biophysikalische Chemie, Goettingen, Germany. ⁴Pediatric Cardiology and Intensive Care, University Clinic Goettingen, Goettingen, Germany.

Published: 16 January 2014

doi:10.1186/1532-429X-16-S1-P34

Cite this article as: Unterberg-Buchwald et al.: Real time cardiac MRI and its clinical usefulness in arrhythmias and wall motion abnormalities. *Journal of Cardiovascular Magnetic Resonance* 2014 16(Suppl 1):P34.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit

