

Original paper

Does dual-energy abdominal computed tomography increase the radiation dose to patients: a prospective observational study

Dilip John^{B,C,E,F}, Athira R.^B, Swathigha Selvaraj^{C,F}, Rupa Renganathan^{A,E}, Kannan Gunasekaran^{B,C}, Venkatesh Kasi Arunachalam^{A,D,E}

Department of Radiology, Kovai Medical Center and Hospital, Coimbatore, India

Abstract

Purpose: The aim of our study was to compare single-energy (SECT) and dual-energy (DECT) abdominal computed tomography (CT) examinations in matched patient cohorts regarding the differences in effective radiation dose (ERD) and image quality performed in a third-generation dual-source computed tomography (DSCT) scanner.

Material and methods: Our study included 100 patients, who were divided randomly into 2 groups. The patients included in Group A were scanned by SECT, and Group B members were scanned by DECT. Volume CT dose index (CTDI_{vol}), dose length product (DLP), and ERD for venous phase acquisition were recorded in each patient and were normalised for 40 cm. Analyses were performed by using statistical software (SPSS version 20.0 for windows), and Bonferroni correction for multiple comparisons was applied for *p*-values and confidence intervals.

Results: Average ERD based on DLP values normalised for 40 cm acquisition were obtained for both Group A and Group B. The mean ERD for Group A was 11.89 mSv, and for group B it was 6.87 mSv. There was a significant difference in these values between Group A and Group B as shown by a *p*-value of < 0.001. On subjective and objective analysis, there was no statistically significant difference in image quality between the 2 groups.

Conclusions: The protocols in third-generation DSCT using dual-energy mode resulted in significant reductions in the effective radiation dose (by approximately 58%) compared to SECT in routine abdominal examination in matched cohorts. Therefore, the quantitative imaging potential of DECT can be utilised in needed patients with decreased radiation dose in third-generation DSCT.

Key words: radiation dose, third generation, DSCT, DECT, SECT, image quality.

Introduction

Dual-energy computed tomography (DECT) has been increasingly used in abdominal studies because it helps in material decomposition and also facilitates hepatic iron load differentiation, gallstones, and renal calculi. Abdominal spectral DECT has been used to analyse iodine uptake for characterisation of hepatic, renal, and adrenal lesions, assessment of bone mineral density of the lumbar spine, and to decrease beam-hardening artefacts. The post-processing of DECT data sets has also shown potential to improve image contrast and increase iodine signal intra-

vascularly in suboptimal contrast conditions. Such techniques can also be used to reduce the volume of contrast in contrast-enhanced CT.

In the current era, because of its increased applications, DECT is routinely performed. However, the common misconception among clinicians is that DECT increases the radiation dose because the patient is scanned by dual energy levels [1]. We need convincing evidence to suggest there is no radiation dose penalty associated with DECT, so that all DECT advanced applications can be used in routine clinical practice. Previous studies have shown that second-generation DECT can be performed

Correspondence address:

Venkatesh Kasi Arunachalam, Department of Radiology, Kovai Medical Center and Hospital, Coimbatore, India, e-mail: drkasivenkatesh@yahoo.co.in

Authors' contribution:

A Study design · B Data collection · C Statistical analysis · D Data interpretation · E Manuscript preparation · F Literature search · G Funds collection

with no or only minimal dose penalty compared to single-energy CT (SECT) acquisition [2,3]. With the recent third-generation DSCT system, which has substantially higher X-ray tube current reserves, additional dual-energy tube voltage combinations, and a thicker tin filter for both tubes, an increase in spectral separation and further reduction in the effective radiation dose to the patient is possible.

The purpose of this study was to calculate and compare the effective radiation dose of SECT and DECT of abdomen examinations performed with third-generation dual-source CT (DSCT).

Material and methods

The study was approved by the local ethics committee and written informed consent was obtained from each participant. This was an observational study and was done during the period October 2019 to March 2020. Patients with age > 18 years, who underwent contrast-enhanced CT abdomen examinations were included in the study. Patients with absolute contraindications for iodinated contrast and scans with movement artefacts were excluded from the study.

A total of 100 patients were included and divided into 2 groups of 50 each (Group A and Group B). Groups were matched by gender (25 males and 25 females in each group) and body mass index (BMI) to allow for a direct comparison of study groups. The patients included in Group A were scanned by SECT, and Group B members were scanned by DECT.

Computed tomography technique (image acquisition and post processing)

Scans of groups A and B were performed on third-generation DSCT (Somatom Force, Siemens Healthcare Sector, Forchheim, Germany). All the patients in Group A and Group B underwent scans during the venous phase from the level of the dome of the diaphragm to the lower end of the ischial tuberosity in SECT and DECT, respectively. These patients underwent plain and arterial phases whenever there was a need according to the clinical scenario and the above-mentioned phases in both groups were performed with SECT in our institution.

Dose-optimised SECT was performed using online dose modulation (CARE Dose 4D) and automatic voltage control (Care kV, Siemens). The scan was performed with a detector collimation of 192×0.6 mm in craniocaudal direction and a pitch of 1 (Figures 1-3). The DECT scan was performed by using 2 different tubes voltages (100 kV and tin filtered 150 kV [Sn 150 kV]) and online dose modulation (CARE Dose 4D, Siemens). The scan was acquired with a detector collimation of 128×0.6 mm in craniocaudal direction and a pitch of 1 (Figures 4-6). The acquisition parameters are detailed in Table 1.

Iohexol 350 mg/ml was used as a contrast agent, and a dose of 1 ml/kg was injected at a flow rate of 4 ml/s followed by a 30 ml saline flush using an automated dual-syringe power injector through an 18 gauge IV access placed in the right median cubital vein. The timing of the scan was performed using a bolus tracking software application. The region of interest (ROI) was placed in the abdominal aorta, when HU of 200 was reached, and the scan was triggered with a delay of 4 seconds for early arterial phase, 15 seconds after bolus tracking for late arterial/early portal venous phase and 50-60 seconds delay for the venous phase.

The images in both SECT and DECT were reconstructed using a standard soft tissue reconstruction kernel (Br36) using modern iterative reconstruction technique, namely the Advanced Modeled Iterative Reconstruction (ADMIRE) algorithm, with slice thickness of 0.6 mm. Images with 3 mm thickness in axial, sagittal, and coronal planes were also generated. A single image series was created for each DECT examination using standard linear blending from the spectral datasets (weighting factor of 0.5) which corresponds to a 120 kV scan. After completion of image acquisition and reconstruction, images were transferred to a dedicated workstation (syngo.via, Siemens) for further analysis.

Radiation dose analysis

Patient protocols for each patient in the 2 groups were analysed, and the tube current, volume CT dose index ($CTDI_{vol}$), dose length product (DLP), and acquisition length in cm for venous phase acquisition were recorded.

To account for variations in scan acquisition length between patients, DLP was normalised for a typical abdominal acquisition covering 40 cm. The ERD estimate was calculated for each patient by multiplying the DLP with an abdomen-specific conversion coefficient κ of $0.017 \text{ mSv}/[\text{mGy} \cdot \text{cm}]$ [4,5]. Previous studies have demonstrated that this conversion factor can be used for both SECT and DECT acquisition [6].

Subjective image analysis

Subjective image analysis was done by a radiologist with more than 10 years' experience in abdominal cross-sectional imaging using the 5-point scale for anatomical details and artefacts (Table 2) in a 3 mm axial slice image dataset. This scoring criteria was developed on the basis of previous studies done by Li *et al.* [7] and De Cecco *et al.* [8].

Objective image quality evaluation

The image quality was objectively evaluated using the contrast-to-noise ratio (CNR) and figure-of-merit (FOM) values. Initially CT attenuation values were measured by placing a circular ROI over the liver (ROI size, 50-100 mm²), spleen (ROI size, 50-100 mm²), portal vein (ROI size,



Figure 1. Axial abdominal contrast enhanced venous phase computed tomography image, acquired in single-energy mode, of a 34-year-old female with a body mass index of 22.15 kg/m²

30-60 mm²), erector spinae muscle (ROI size, 50-100 mm²), and anterior abdominal wall subcutaneous fat (ROI size, 30-100 mm²) on 3 mm axial images of all 100 data sets. Areas with focal heterogeneity of contrast enhancement, focal calcification, and adjacent anatomical structures were excluded to facilitate homogenous measurements. The measurements were repeated 3 times to avoid data inaccuracies, and the mean values were calculated. The image noise was defined as the standard deviation of the fat.

The contrast-to-noise ratios (CNR) of the liver, spleen, and portal vein were measured by dividing the absolute difference of the attenuations within the ROI and the



Figure 2. Coronal and sagittal reformatted abdominal contrast enhanced venous phase computed tomography images acquired in single-energy mode of the same patient

Scan	kV	mAs / ref.	CTDIvol* mGy	DLP mGycm	TI s	cSL mm
Patient Position F-SP						
Topogram	1	120 18 mA	0.04 L	2.1	3.1	0.6
Plain Scan	2	120 91 /147	6.14 L	315.6	0.5	0.6
PreMonitoring Contrast	3	100 23	0.79 L	0.8	0.5	10.0
Monitoring	4	100 23	5.53 L	5.5	0.5	10.0
Arterial Phase	11	120 202 /300	13.57 L	733.7	0.5	0.6
Venous Phase	12	120 188 /300	12.62 L	682.6	0.5	0.6
Medium	Type	Iodine Conc. mg/ml	Volume ml	Flow ml/s	CM Ratio	
Saline			30	3.6		
Contrast		0	70	3.9	100%	
Saline			30	3.7		

Figure 3. Dose chart of the above-mentioned patient done in single-energy computed tomography



Figure 4. Axial abdominal contrast-enhanced venous phase computed tomography images acquired in dual-energy mode of a 60-year-old female with a body mass index of 22.3 kg/m²

erector spinae muscle by the image noise. To account for differences in tube voltage settings of SECT and DECT, figure-of-merit (FOM) values were calculated [3,9]. FOM values were calculated as the ratio of CNR² to effective radiation dose. These values were used for CNR assessment independent of ERD.

Statistical analysis

The results were collected in a Microsoft Excel file. Analyses were performed by using statistical software (SPSS version 20.0 for windows). The independent sample t test was performed. A *p*-value < 0.05 indicated a statistically significant difference. Bonferroni correction for multiple comparisons was applied for *p*-values and confidence intervals. Comparisons among the 2 groups were analysed




Figure 5. Coronal and sagittal reformatted abdominal contrast enhanced venous phase computed tomography images of the same patient

Scan	kV	mAs / ref.	CTDIvol* mGy	DLP mGycm	TI s	cSL mm
Patient Position F-SP						
Topogram	1	120	18 mA	0.04 L	2.1	3.2
Plain Scan	2	120	78 /147	5.25 L	266.3	0.5
PreMonitoring Contrast	3	100	23	0.79 L	0.8	0.5
Monitoring	4	100	23	7.12 L	7.1	0.5
Arterial Phase	13	120	171 /300	11.48 L	622.5	0.5
Portal Phase	14	120	155 /300	10.37 L	359.0	0.5
DE_Abdomen	15A	100	90 /180			
	15B	Sn150	51 / 90	5.48 L	285.7	0.5
Delayed Phase	16	120	126 /250	8.46 L	284.3	0.5
Medium	Type		Iodine Conc. mg/ml	Volume ml	Flow ml/s	CM Ratio
Saline				30	3.6	
Contrast			0	70	3.9	100%
Saline				30	3.7	

Figure 6. Dose chart of the above-mentioned patient done in dual-energy computed tomography

Table 1. Acquisition parameters of SECT (Group A) and DECT (Group B)

Parameter	Group A	Group B
Acquisition mode	SECT	DECT
Tube voltage (kV)	120	90/Sn 150
Tube current (ref. mAs)	300	180/90
Pitch	1	1
Rotation time (s)	0.5	0.5
Collimation	192 × 0.6	128 × 0.6
Section thickness (mm)	0.6 mm axial, 3 mm in all three planes	0.6 mm axial, 3 mm in all three planes
Iterative reconstruction algorithm	ADMIRE, strength 3	ADMIRE, strength 3
Increment (mm)	1	1
Kernel	Br36	Br36
Linear blending in dual-energy mode	–	50%/50% (90/150 Kv)
Tin filter	–	Selective photon shield II

ADMIRE – advanced modelled iterative reconstruction

Table 2. Subjective image analysis criteria based on artefacts and anatomical details

Grade	Artefacts	Anatomical details and lesions
1	Severe artefacts	Organs – not assessable, lesions – not visible
2	Artefacts affecting diagnostic information	Anatomical details of organs – not clearly visible, lesions – not clear
3	Obvious artefacts but acceptable	Anatomical details of organs and lesions – seen to an acceptable extent
4	Artefacts were seen in other organs not interfering with the diagnosis	Organs and lesions – seen clearly
5	Minimum or no artefacts	Organs and lesions – seen clearly

with repeated measures analysis of variance if data were normally distributed according to the Shapiro-Wilk test. The Kruskal-Wallis 2-way analysis of variance was used if data were not normally distributed. If the effective dose was not normally distributed, the mean difference between ERD in Group A and Group B was tested through non-parametric Mann-Whitney *U* test.

Results

The mean age of sample Group A was 47.92 years and of Group B was 54.72 years. The age ranged between 19 and 78 years. The majority of the patients were between the age of 41 and 60 years. The mean BMI in Group A was 22.47 kg/m² and in Group B it was 22.47 kg/m² with a range of between 15.6 and 29.3 kg/m². Because the patient cohorts were matched by gender and BMI, no significant differences regarding these parameters ($p > 0.99$) were observed, although the standard deviations varied (Table 3).

Average effective radiation dose based on DLP values normalised for 40 cm acquisition was obtained for both Group A and Group B (Table 4).

The mean CTDI volume for Group A was 13.41 and for Group B it was 7.94. The mean DLP for Group A was 704.53 and for Group B it was 404.55. The mean ERD for Group A was 11.89 mSv and for group B it was 6.87 mSv. There was a significant difference in these val-

ues between Group A and Group B as shown by a p -value of < 0.001 . The average effective radiation dose based for DLP values normalised for 40 cm acquisition was lowest in group B, with a significant difference compared to group A ($p < 0.001$).

Subjective score

The subjective score was analysed using a 5-point scale ranging from grade 1 to 5, where 1 was the worst and 5 was the best possible score. The mean subjective score obtained for the 2 groups (A and B) were 4.38 ± 0.52 and 4.4 ± 0.53 (mean \pm SD), respectively. The p -value for individual differences of subjective score between Group A and Group B is 0.32, indicating that the difference in image quality was not statistically significant between the 2 groups (Table 5).

Objective image quality evaluation

There was no statistically significant difference in image noise between the 2 groups. The highest CNR and FOM values for liver, spleen, and portal vein were noted in the dual-energy group. However, the differences in CNR and FOM values of liver, spleen, and portal vein between both groups were not statistically significant. A detailed list of all objective image quality parameters is shown in Table 6.

Table 3. Statistical analysis for age, height, weight, and BMI for Groups A and B

Parameter	Number of samples	Mean	Standard deviation	Standard error mean	p-value (A vs. B)
Age					
Group A	50	47.92	16.070	2.273	0.019
Group B	50	54.72	12.022	1.700	
Height (cm)					
Group A	50	160.10	10.363	1.466	0.027
Group B	50	164.16	7.506	1.062	
Weight (kg)					
Group A	50	57.770	11.4706	1.6222	0.2
Group B	50	60.540	9.9662	1.4094	
BMI (kg/m ²)					
Group A	50	22.4760	3.27128	.46263	0.993
Group B	50	22.4700	3.27721	.46347	

Table 4. Statistical analysis for CTDI_{vol}, DLP, and effective dose for Groups A and B

Parameter	Number of samples	Mean	Standard deviation	Standard error mean	p-value (A vs. B)
CTDI _{vol}					
Group A	50	13.4152	4.16509	0.58903	0.000
Group B	50	7.9440	2.39462	0.33865	
DLP					
Group A	50	704.536	243.7485	34.4712	0.000
Group B	50	404.550	128.7641	18.2100	
Effective dose					
Group A	50	11.892506	4.1329150	0.5844824	0.000
Group B	50	6.877350	2.1889903	0.3095700	

Table 5. Mean subjective score and SD in Group A and Group B

Subjective score	Mean	SD
Group A	4.38	0.52
Group B	4.4	0.53

Paired t-test; p-value – 0.322223; non-significant

Table 6. Objective image quality parameters

Parameter	Group A	Group B	p-value	Significant (S)/Non-significant (NS)
Image noise	8.45 ± 2.50	8.40 ± 2.06	0.945	NS
Contrast-to-noise ratio (CNR)				
Liver	5.19 ± 3.06 (1.8-11.1)	4.88 ± 2.43 (1.8-12)	0.73	NS
Spleen	6.33 ± 3.49 (2.5-12.3)	5.78 ± 2.63 (2.6-13.2)	0.553	NS
Portal vein	11.72 ± 6.17 (3.2-24.2)	10.43 ± 4.62 (3.1-24.4)	0.462	NS
Figure of merit CNR				
Liver	4.21 ± 6.57 (0.3-15.9)	5.070 ± 7.00 (0.4-30.3)	0.69	NS
Spleen	5.28 ± 6.67 (0.5-24.9)	6.76 ± 8.44 (0.6-36.7)	0.543	NS
Portal vein	18.27 ± 22.94 (1.2-65.3)	22.38 ± 28.16 (2.9-125.6)	0.616	NS

Data are given as mean ± standard deviation and range in parenthesis.

Discussion

With the current technological advancements, DECT has gained worldwide acceptance. In abdominal imaging, DECT examinations are mainly considered for oncological staging and follow-up, evaluation of vascular pathology, preoperative workups, and evaluation of gastrointestinal bleeding. DECT also has potential in radiation dose reduction.

The previously used first- and second-generation DSCT had certain disadvantages, such as poor spatial resolution, motion-related artefacts, and increased noise and radiation dose [10-13]. These problems were addressed in the third-generation DSCT. Although there are different software and hardware tools among the third-generation DSCT scanners for radiation dose reduction, there is much apprehension regarding the same among the clinicians [1]. We performed this study to compare the radiation dose in single- and dual-energy CT of abdomen examinations in matched patient cohort.

Evaluation of radiation dose

The results of our study show that dual-energy CT can be performed without radiation dose penalty compared to SECT with a third-generation DSCT scanner. When we evaluated a single venous phase in a matched patient cohort in third-generation DSCT using dual-energy mode, a dose reduction of 57.8% was achieved compared to single-energy CT.

The most efficient way to reduce radiation dose in CT is by adapting the scan parameters to the patient's anatomy. Centring the patient correctly, using the right protocols, and adjusting the X-ray tube output to the patient's size and shape help to minimize radiation exposure. Third-generation DSCT has control mechanisms that automatically adjust the radiation dose to the patient's anatomy. This technique is called CARE Dose4D, which automatically adapts the tube current to the size and shape of the patient and achieves optimal tube current modulation [14,15]. In this the tube current is first varied on the basis of a topogram, by comparing the actual patient to a "standard-sized" patient. The tube current as expected is increased for larger patients and reduced for smaller patients. The differences in attenuation in distinct body regions are also taken into account. Also, the real-time angular dose modulation measures the actual attenuation in the patient during the scan and adjusts the tube current accordingly – not only for different body regions, but also for different angles during rotation. The improved X-ray tube current capabilities of third-generation DSCT allow usage of a lower tube voltage (i.e. 100 kV) in abdominal imaging.

The next most important factor useful in reducing radiation dose in third-generation DSCT is tin filters. In third-generation DSCT, each X-ray source has a tin filter

that filters out unnecessary photons for powerful low-dose scanning. It helps to achieve an optimized spectrum for dose efficiency and helps to deliver powerful low-dose scanning at the level of conventional X-ray examinations. Also, the third-generation DSCT systems are equipped with digital detectors (Stellar infinity detectors), which are more sensitive to electron influx. These detectors can measure smaller signals over a wider dynamic range, which directly enhances CT image quality especially for applications with extremely low signal levels. Such extremely low signal levels play an important role when scanning large patients and in low-dose scans, as well as in the low-kV datasets of dual-energy examinations. All the above-mentioned factors result in a significant reduction in radiation dose in third-generation dual-source CT. Our results for radiation doses with third-generation dual-energy CT (6.87 ± 4.37 mSv) averaged for 40-cm acquisitions are well below recently reported results associated with contrast-enhanced routine abdomen-pelvis CT (median of 16 mSv) [14] and are in line with the suggestions (10 mSv) from a patient education website [15].

Some of the prior studies that have investigated the radiation dose delivered by DECT compared to SECT have shown varying results. De Cecco *et al.* assessed second-generation DSCT protocols and compared SECT and DECT [4] and reported a small but significant increase in the radiation dose of the order 1 mSv with dual-energy CT. No significant differences in radiation dose between abdominal second-generation SECT and DECT was found in the study by Primak *et al.* [16]. A significant decrease in radiation dose with DECT compared to SECT in patients who underwent abdominal second-generation SECT and DECT for hepatocellular carcinoma screening was observed in a study by Purysko *et al.* [17]. In their study they observed a 37% reduction in overall dose with DECT compared to SECT. In our study we observed that third-generation DECT showed a significant decrease in radiation dose (57.8%) compared to SECT. Thus, dual-energy CT can be implemented in routine clinical use without negatively impacting image quality while lowering the radiation dose to the patient.

In our study, we observed the lowest radiation doses with DECT in third-generation DSCT. The mean effective radiation dose in second-generation dual-source CT in the study by Purysko *et al.* was 12.97 mSv [17]. However, in their study the results were not normalised for 40 cm acquisition. In a study by De Cecco *et al.*, the mean radiation dose in second-generation dual-source CT normalised for 40 cm was 10.2 mSv [4]. In our study the mean effective radiation dose was 6.88 mSv. Our results were similar to those in the study by Wichmann *et al.* [3]. They concluded that with both second and third DSCT generations, abdominal DECT can be routinely performed without radiation dose penalty compared to SECT, while third-generation DSCT shows improved dose efficiency. David *et al.* in their study observed that




Figure 7. Axial sections of abdomen in single-energy computed tomography (A) and dual-energy computed tomography (B) with grade 5 image quality

the mean dose-length product (DLP) and effective dose of DECT were significantly lower than the DLP and effective dose of SECT ($p < 0.05$) [18]. The mean radiation dose in their study was 6.1 mSv. A reliable explanation of this reduced radiation dose could be the improved tube current capabilities in third-generation DSCT.

The other factor that plays an important role in dose reduction was modern iterative reconstruction algorithms. In our CT scan the latest generation of Siemens iterative reconstruction, called the ADMIRE algorithm, was used for the reconstruction of both SECT and DECT images. It was applied at a strength of 2 (ADMIRE 2) out of a maximum strength of 5. This level 2 strength represents a low to medium level of noise suppression due to iterative reconstruction. ADMIRE reduces image noise on CT images and maintains the spatial resolution and contrast-to-noise ratio. Some of the studies in the previous literature have mentioned that ADMIRE plays an important role in reducing radiation dose by compensating image noise with its post processing capabilities [19,20].

Further reduction in radiation dose is possible in DECT scans in multiphase examination, by utilising virtual non-contrast images and avoiding true unenhanced phase/plain scan. In a study by De Cecco *et al.*, it was observed that by avoiding true unenhanced acquisition the radiation dose can be reduced by 32.9% [8]. Similarly, in a study by Chen *et al.*, a decrease in radiation dose of 57% was achieved by omitting true unenhanced images in a single-phase DECT scan [21].

Evaluation of subjective criteria

The basic aim of the protocols used in this study was to generate good quality images with minimal radiation exposure. In our study we used the 5-point scoring system for assessing the subjective criteria. These scoring criteria were developed on the basis of the previous studies done by De Cecco *et al.* [8] and Li *et al.* [7]. All the images assessed had a score of 3 or more, indicating that they were suitable for interpretation (Figure 7). The p -value for in-

dividual differences of subjective score between Group A and Group B was 0.32, indicating that the difference in image quality was not statistically significant between the 2 groups. This result is similar to the study done by Wichmann *et al.* [3], which showed no significant difference in image quality between single- and dual-energy scans. In a study by Tawfik *et al.*, no significant difference in subjective image quality was observed between SECT and DECT protocols [22].

Evaluation of objective image quality

Objective image quality assessment revealed the highest values for CNR and FOM for all the assessed organs in the DECT group. In our study, the differences in the CNR and FOM values of liver, spleen, and portal vein between the SECT and DECT groups were not statistically significant. The images obtained using a third-generation dual-source scanner in SECT and DECT modes were comparable in our study. The result was similar to that from the study done by Lenga *et al.* [23], which showed no significant difference in objective image quality in images obtained using SECT and DECT modes in a third-generation dual-source scanner. In another study done by Lenga *et al.* [24], in the evaluation of chest CT, there were no statistically significant differences in CNR values of descending thoracic aorta, pulmonary trunk, and thyroid gland as well as the FOM value of the pulmonary trunk between SECT and DECT groups of a third-generation scanner.

Hence, DECT can be considered in all scenarios because it can increase both lesion detection and characterization and has the capacity to decrease artifacts and possibly contrast media dose, all while maintaining or even decreasing the radiation dose to the patient. The spectral imaging potential of DECT can be utilised routinely in abdominal DSCT.

Our study had some limitations. This study was done on the dual-source dual-energy technique. Hence, the effective radiation dose values obtained cannot be generalised to other techniques of dual-energy CT like rapid

kilovolt switching, multilayer detector technique, etc., which might show different results. The study was a single institutional study with limited subjects. Further large-scale studies and meta-analysis are needed to confirm our results.

Conclusions

In third-generation dual-source dual-energy CT there is a significant reduction in the effective radiation dose to the patient compared to single-energy CT in matched cohorts. In our study, the routine abdominal examination

protocols in third-generation DSCT using dual-energy mode resulted in an approximately 58% dose reduction compared to single-energy CT with similar image quality. Therefore, the quantitative imaging potential of DECT can be routinely utilised with decreased radiation dose in third-generation dual-source dual-energy CT in abdominal imaging.

Conflict of interest

The authors report no conflict of interest.

References

- Sodickson AD, Keraliya A, Czakowski B, et al. Dual energy CT in clinical routine: how it works and how it adds value. *Emerg Radiol* 2021; 28: 103-117.
- Uhrig M, Simons D, Kachelrieß M, et al. Advanced abdominal imaging with dual energy CT is feasible without increasing radiation dose. *Cancer Imaging* 2016; 16: 15.
- Wichmann JL, Hardie AD, Schoepf UJ, et al. Single- and dual-energy CT of the abdomen: comparison of radiation dose and image quality of 2nd and 3rd generation dual-source CT. *Eur Radiol* 2017; 27: 642-650.
- De Cecco CN, Darnell A, Macias N. Second-generation dual-energy computed tomography of the abdomen: radiation dose comparison with 64- and 128-row single-energy acquisition. *J Comput Assist Tomogr* 2013; 37: 543-546.
- The 2007 recommendations of the International Commission on Radiological Protection. ICRP publication 103. *Ann ICRP* 2007; 37: 1-332.
- Christner JA, Kofler JM, McCollough CH. Estimating effective dose for CT using dose-length product compared with using organ doses: consequences of adopting International Commission on Radiological Protection publication 103 or dual-energy scanning. *Am J Roentgenol* 2010; 194: 881-889.
- Li Y, Li Y, Jackson A, et al. Comparison of virtual unenhanced CT images of the abdomen under different iodine flow rates. *Abdom Radiol* 2017; 42: 312-321.
- De Cecco CN, Muscogiuria G, Schoepf UJ, et al. Virtual unenhanced imaging of the liver with third-generation dual-source dual-energy CT and advanced modelled iterative reconstruction. *Eur J Radiol* 2016; 85: 1257-1264.
- Schindera ST, Nelson RC, Mukundan S Jr, et al. Hypervascular liver tumors: low tube voltage, high tube current multi-detector row CT for enhanced detection – phantom study. *Radiology* 2008; 246: 125-132.
- Chiro GD, Brooks RA, Kessler RM, et al. Tissue signatures with dual-energy computed tomography. *Radiology* 1979; 131: 521-523.
- Millner MR, McDavid WD, Waggner RG, et al. Extraction of information from CT scans at different energies. *Med Phys* 1979; 6: 70-71.
- Kalender WA, Perman WH, Vetter JR, et al. Evaluation of a prototype dual-energy computed tomographic apparatus. I. Phantom studies. *Med Phys* 1986; 13: 334-339.
- Kelcz F, Joseph PM, Hilal SK. Noise considerations in dual energy CT scanning. *Med Phys* 1979; 6: 418-425.
- Smith-Bindman R, Lipson J, Marcus R, et al. Radiation dose associated with common computed tomography examinations and the associated lifetime attributable risk of cancer. *Arch Intern Med* 2009; 169: 2078-2086.
- Patient safety: radiation dose in X-ray and CT exams. Available at: http://www.radiologyinfo.org/en/safety/?pg=sfty_xray#part3 (Accessed: 28.04.2015).
- Primak AN, Giraldo JC, Eusemann CD. Dual-source dual-energy CT with additional tin filtration: dose and image quality evaluation in phantoms and in vivo. *AJR Am J Roentgenol* 2010; 195: 1164-1174.
- Puryško AS, Primak AN, Baker ME. Comparison of radiation dose and image quality from single-energy and dual-energy CT examinations in the same patients screened for hepatocellular carcinoma. *Clin Radiol* 2014; 69: e538-544.
- Schmidt D, Söderberg M, Nilsson M, et al. Evaluation of image quality and radiation dose of abdominal dual-energy CT. *Acta Radiol* 2018; 59: 845-852.
- Gordic S, Morsbach F, Schmidt B, et al. Ultralow-dose chest computed tomography for pulmonary nodule detection: first performance evaluation of single energy scanning with spectral shaping. *Invest Radiol* 2014; 49: 465-473.
- Schmid AI, Uder M, Lell MM. Reaching for better image quality and lower radiation dose in head and neck CT: advanced modelled and sinogram-affirmed iterative reconstruction in combination with tube voltage adaptation. *Dentomaxillofac Radiol* 2017; 46: 20160131.
- Chen CY, Hsu JS, Jaw TS, et al. Split-bolus portal venous phase dual-energy CT urography: protocol design, image quality, and dose reduction. *AJR Am J Roentgenol* 2015; 205: W492-W501.
- Tawfik AM, Kerl JM, Razek AA, et al. Image quality and radiation dose of dual-energy CT of the head and neck compared with a standard 120-kVp acquisition. *Am J Neuroradiol* 2011; 32: 1994-1999.
- Lenga L, Trapp F, Albrecht MH, et al. Single- and dual-energy CT pulmonary angiography using second- and third-generation dual-source CT systems: comparison of radiation dose and image quality. *Eur Radiol* 2019; 29: 4603-4612.
- Lenga L, Leithner D, Peterke JL, et al. Comparison of radiation dose and image quality of contrast-enhanced dual-source CT of the chest: single-versus dual-energy and second-versus third-generation technology. *Am J Roentgenol* 2019; 212: 741-747.